

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3584</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/10/2025 09:36</u>
Lab File ID:	<u>VX048527.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.405	0.480		18.52	20
Chloromethane	0.537	0.565	0.1	5.21	20
Vinyl Chloride	0.608	0.621		2.14	20
Bromomethane	0.384	0.424		10.42	20
Chloroethane	0.398	0.393		-1.26	20
Trichlorofluoromethane	0.963	0.965		0.21	20
1,1,2-Trichlorotrifluoroethane	0.557	0.576		3.41	20
1,1-Dichloroethene	0.582	0.562		-3.44	20
Acetone	0.330	0.343		3.94	20
Carbon Disulfide	1.638	1.516		-7.45	20
Methyl tert-butyl Ether	2.055	2.069		0.68	20
Methyl Acetate	0.906	0.850		-6.18	20
Methylene Chloride	0.678	0.639		-5.75	20
trans-1,2-Dichloroethene	0.617	0.596		-3.4	20
1,1-Dichloroethane	1.142	1.122	0.1	-1.75	20
Cyclohexane	0.936	0.927		-0.96	20
2-Butanone	0.488	0.511		4.71	20
Carbon Tetrachloride	0.577	0.527		-8.67	20
cis-1,2-Dichloroethene	0.754	0.725		-3.85	20
Bromochloromethane	0.561	0.484		-13.73	20
Chloroform	1.179	1.127		-4.41	20
1,1,1-Trichloroethane	1.029	0.979		-4.86	20
Methylcyclohexane	0.569	0.601		5.62	20
Benzene	1.571	1.482		-5.66	20
1,2-Dichloroethane	0.538	0.510		-5.2	20
Trichloroethene	0.375	0.381		1.6	20
1,2-Dichloropropane	0.361	0.377		4.43	20
Bromodichloromethane	0.526	0.561		6.65	20
4-Methyl-2-Pentanone	0.571	0.642		12.43	20
Toluene	0.827	0.921		11.37	20
t-1,3-Dichloropropene	0.541	0.584		7.95	20
cis-1,3-Dichloropropene	0.561	0.625		11.41	20
1,1,2-Trichloroethane	0.330	0.369		11.82	20
2-Hexanone	0.425	0.474		11.53	20
Dibromochloromethane	0.406	0.451		11.08	20
1,2-Dibromoethane	0.358	0.397		10.89	20
Tetrachloroethene	0.363	0.364		0.28	20
Chlorobenzene	1.187	1.191	0.3	0.34	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3584</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/10/2025 09:36</u>
Lab File ID:	<u>VX048527.D</u>	Init. Calib. Date(s):	<u>11/07/2025 11/07/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:35 16:29</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.993	2.013		1	20
m/p-Xylenes	0.751	0.779		3.73	20
o-Xylene	0.734	0.763		3.95	20
Styrene	1.232	1.286		4.38	20
Bromoform	0.367	0.373	0.1	1.63	20
Isopropylbenzene	3.629	3.726		2.67	20
1,1,2,2-Tetrachloroethane	1.280	1.279	0.3	-0.08	20
1,3-Dichlorobenzene	1.756	1.740		-0.91	20
1,4-Dichlorobenzene	1.814	1.782		-1.76	20
1,2-Dichlorobenzene	1.686	1.710		1.42	20
1,2-Dibromo-3-Chloropropane	0.297	0.312		5.05	20
1,2,4-Trichlorobenzene	1.120	1.177		5.09	20
1,2,3-Trichlorobenzene	1.042	1.088		4.41	20
1,2-Dichloroethane-d4	0.697	0.616		-11.62	20
Dibromofluoromethane	0.378	0.321		-15.08	20
Toluene-d8	1.180	1.100		-6.78	20
4-Bromofluorobenzene	0.440	0.405		-7.95	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.