

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048524.D
 Acq On : 07 Nov 2025 16:29
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Nov 08 04:43:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	180041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	307413	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	271611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	137081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	123374	49.178	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.360%		
35) Dibromofluoromethane	5.373	113	107888	46.370	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.740%		
50) Toluene-d8	8.635	98	369557	50.927	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.860%		
62) 4-Bromofluorobenzene	11.061	95	136244	50.381	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.760%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	76351	52.335	ug/l	100
3) Chloromethane	1.301	50	106667	55.126	ug/l	100
4) Vinyl Chloride	1.380	62	107960	49.279	ug/l	100
5) Bromomethane	1.612	94	76656	55.356	ug/l	100
6) Chloroethane	1.691	64	68875	48.005	ug/l	100
7) Trichlorofluoromethane	1.892	101	159505	46.021	ug/l	100
8) Diethyl Ether	2.136	74	70765	50.864	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	92738	46.218	ug/l	100
10) Methyl Iodide	2.453	142	148356	49.960	ug/l	100
11) Tert butyl alcohol	2.941	59	124670	262.256	ug/l	100
12) 1,1-Dichloroethene	2.325	96	98074	46.794	ug/l	100
13) Acrolein	2.240	56	18158	266.459	ug/l	100
14) Allyl chloride	2.666	41	188637	48.341	ug/l	100
15) Acrylonitrile	3.056	53	363119	261.664	ug/l	100
16) Acetone	2.374	43	296864	249.987	ug/l	100
17) Carbon Disulfide	2.514	76	268053	45.450	ug/l	100
18) Methyl Acetate	2.697	43	156429	47.934	ug/l	100
19) Methyl tert-butyl Ether	3.105	73	381141	51.497	ug/l	100
20) Methylene Chloride	2.788	84	117536	48.142	ug/l	100
21) trans-1,2-Dichloroethene	3.093	96	107632	48.453	ug/l	100
22) Diisopropyl ether	3.751	45	371667	49.602	ug/l	100
23) Vinyl Acetate	3.715	43	1644477	250.599	ug/l	100
24) 1,1-Dichloroethane	3.605	63	198326	48.238	ug/l	100
25) 2-Butanone	4.538	43	455146	259.087	ug/l	100
26) 2,2-Dichloropropane	4.465	77	159000	45.363	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	130445	48.042	ug/l	100
28) Bromochloromethane	4.885	49	97819	48.431	ug/l	100
29) Tetrahydrofuran	4.989	42	313988	250.943	ug/l	100
30) Chloroform	5.087	83	201702	47.521	ug/l	100
31) Cyclohexane	5.458	56	156130	46.347	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	175119	47.275	ug/l	100
36) 1,1-Dichloropropene	5.678	75	131207	42.114	ug/l	100
37) Ethyl Acetate	4.696	43	205009	46.746	ug/l	100
38) Carbon Tetrachloride	5.666	117	149370	42.117	ug/l	100
39) Methylcyclohexane	7.367	83	156904	44.852	ug/l	100
40) Benzene	6.025	78	432976	44.833	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
 Data File : VX048524.D
 Acq On : 07 Nov 2025 16:29
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Nov 08 04:43:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 04:36:23 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/10/2025
 Supervised By :Semsettin Yesilyurt 11/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	90811	47.085	ug/l	100
42) 1,2-Dichloroethane	6.074	62	147877	44.690	ug/l	100
43) Isopropyl Acetate	6.324	43	285622	46.331	ug/l	100
44) Trichloroethene	7.111	130	109058	47.244	ug/l	100
45) 1,2-Dichloropropane	7.415	63	110204	49.629	ug/l	100
46) Dibromomethane	7.568	93	82377	50.863	ug/l	100
47) Bromodichloromethane	7.806	83	167028	51.695	ug/l	100
48) Methyl methacrylate	7.678	41	142167	51.877	ug/l	100
49) 1,4-Dioxane	7.647	88	45365	1086.384	ug/l	100
51) 4-Methyl-2-Pentanone	8.555	43	952270	271.365	ug/l	100
52) Toluene	8.702	92	269581	53.020	ug/l	100
53) t-1,3-Dichloropropene	8.964	75	175854	52.820	ug/l	100
54) cis-1,3-Dichloropropene	8.354	75	186249	54.027	ug/l	100
55) 1,1,2-Trichloroethane	9.135	97	110910	54.597	ug/l	100
56) Ethyl methacrylate	9.098	69	190086	56.242	ug/l	100
57) 1,3-Dichloropropane	9.293	76	190119	53.768	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	544454	302.318	ug/l	100
59) 2-Hexanone	9.415	43	713506	273.339	ug/l	100
60) Dibromochloromethane	9.507	129	134082	53.758	ug/l	100
61) 1,2-Dibromoethane	9.592	107	117046	53.197	ug/l	100
64) Tetrachloroethene	9.257	164	88926	45.076	ug/l	100
65) Chlorobenzene	10.061	112	299156	46.404	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.147	131	106481	49.295	ug/l	100
67) Ethyl Benzene	10.177	91	512428	47.335	ug/l	100
68) m/p-Xylenes	10.287	106	391085	95.921	ug/l	100
69) o-Xylene	10.622	106	190935	47.864	ug/l	100
70) Styrene	10.634	104	328665	49.116	ug/l	100
71) Bromoform	10.781	173	94463	47.401	ug/l	100
73) Isopropylbenzene	10.945	105	475097	47.748	ug/l	100
74) N-amyl acetate	10.823	43	245758	49.190	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.195	83	168512	48.023	ug/l	100
76) 1,2,3-Trichloropropane	11.220	75	138398m	47.352	ug/l	
77) Bromobenzene	11.177	156	124641	48.510	ug/l	100
78) n-propylbenzene	11.287	91	551985	47.178	ug/l	100
79) 2-Chlorotoluene	11.348	91	334083	47.255	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	386446	47.641	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.000	75	59424	51.468	ug/l	100
82) 4-Chlorotoluene	11.433	91	392809	47.366	ug/l	100
83) tert-Butylbenzene	11.695	119	403731	48.121	ug/l	100
84) 1,2,4-Trimethylbenzene	11.732	105	397157	48.526	ug/l	100
85) sec-Butylbenzene	11.872	105	479253	46.218	ug/l	100
86) p-Isopropyltoluene	11.988	119	409592	46.983	ug/l	100
87) 1,3-Dichlorobenzene	11.951	146	225309	46.787	ug/l	100
88) 1,4-Dichlorobenzene	12.024	146	230056	47.350	ug/l	100
89) n-Butylbenzene	12.311	91	372536	45.552	ug/l	100
90) Hexachloroethane	12.518	117	75599	45.515	ug/l	100
91) 1,2-Dichlorobenzene	12.317	146	222592	48.150	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.921	75	40699	49.910	ug/l	100
93) 1,2,4-Trichlorobenzene	13.567	180	146070	47.549	ug/l	100
94) Hexachlorobutadiene	13.707	225	55716	43.386	ug/l	100
95) Naphthalene	13.756	128	507140	51.347	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	139324	48.759	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110725\
Data File : VX048524.D
Acq On : 07 Nov 2025 16:29
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2025
Supervised By :Semsettin Yesilyurt 11/10/2025

Quant Time: Nov 08 04:43:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 04:36:23 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

