

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110725

Review By	John Carlone	Review On	11/10/2025 8:28:24 AM
Supervise By	Amit Patel	Supervise On	11/10/2025 9:54:56 AM
SubDirectory	VX110725	HP Acquire Method	HP Processing Method

STD. NAME	STD REF.#
Tune/Reschk	VP136104
Initial Calibration Stds	VP136097,VP136098,VP135099,VP136100,VP136101,VP135102
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136103
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048515.D	07 Nov 2025 13:06		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX048516.D	07 Nov 2025 13:35		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX048517.D	07 Nov 2025 13:56		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX048518.D	07 Nov 2025 14:16		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX048519.D	07 Nov 2025 14:37	spiked low	JC/MD	Not Ok
6	VSTDIC100	VSTDIC100	VX048520.D	07 Nov 2025 14:58		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX048521.D	07 Nov 2025 15:18		JC/MD	Ok,M
8	VIBLK	VIBLK	VX048522.D	07 Nov 2025 15:39		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX048523.D	07 Nov 2025 16:08	spike error	JC/MD	Not Ok
10	VSTDICCC050	VSTDICCC050	VX048524.D	07 Nov 2025 16:29		JC/MD	Ok,M
11	VSTDICV050	ICVVX110725	VX048525.D	07 Nov 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX111025

Review By	John Carlone	Review On	11/11/2025 8:25:59 AM
Supervise By	Semsettin Yesilyurt	Supervise On	11/11/2025 12:40:16 PM
SubDirectory	VX111025	HP Acquire Method	HP Processing Method 82X110725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136105
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136106,VP136107

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048526.D	10 Nov 2025 08:34		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048527.D	10 Nov 2025 09:36	pH#Lot#V12668	JC/MD	Ok,M
3	VX1110WBL01	VX1110WBL01	VX048528.D	10 Nov 2025 10:04		JC/MD	Ok
4	VX1110MBL01	VX1110MBL01	VX048529.D	10 Nov 2025 11:28		JC/MD	Ok
5	VX1110WBS01	VX1110WBS01	VX048530.D	10 Nov 2025 11:49		JC/MD	Ok,M
6	VX1110WBSD01	VX1110WBSD01	VX048531.D	10 Nov 2025 12:16		JC/MD	Ok,M
7	Q3573-01	STORAGE-BLANK-SO	VX048532.D	10 Nov 2025 12:37	vial A pH<2	JC/MD	Ok
8	Q3573-02	STORAGE-BLANK-WA	VX048533.D	10 Nov 2025 12:58	vial A pH<2	JC/MD	Ok
9	Q3573-03	STORAGE-BLANK-WA	VX048534.D	10 Nov 2025 13:19	vial A pH<2	JC/MD	Ok
10	Q3573-04	STORAGE-BLANK-SAI	VX048535.D	10 Nov 2025 13:43	vial A pH<2	JC/MD	Ok
11	Q3569-03	TB-2	VX048536.D	10 Nov 2025 14:04	vial A pH<2 TB	JC/MD	Ok
12	Q3584-04	EB-2	VX048537.D	10 Nov 2025 14:25	vial A pH<2 EB;Hit of Com.#16	JC/MD	ReRun
13	Q3584-05	FB-2	VX048538.D	10 Nov 2025 14:45	vial A pH<2 FB	JC/MD	Ok
14	Q3584-06	TB-3	VX048539.D	10 Nov 2025 15:06	vial A pH<2 TB	JC/MD	Ok
15	Q3584-01	SW-1	VX048540.D	10 Nov 2025 15:27	vial A pH<2	JC/MD	Ok
16	Q3584-02	SW-2	VX048541.D	10 Nov 2025 15:48	vial A pH<2	JC/MD	Ok
17	Q3584-03	SW-3	VX048542.D	10 Nov 2025 16:08	vial A pH<2	JC/MD	Ok
18	Q3569-01	MW-2	VX048543.D	10 Nov 2025 16:29	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX111025

Review By	John Carlone	Review On	11/11/2025 8:25:59 AM
Supervise By	Semsettin Yesilyurt	Supervise On	11/11/2025 12:40:16 PM
SubDirectory	VX111025	HP Acquire Method	HP Processing Method 82X110725W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136105		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136106,VP136107		

19	Q3569-02	MW-12	VX048544.D	10 Nov 2025 16:49	vial A pH<2 Need 5X	JC/MD	Dilution
20	Q3584-07	SEEP-1	VX048545.D	10 Nov 2025 17:10	Need Straight Run	JC/MD	Not Ok
21	IBLK	IBLK	VX048546.D	10 Nov 2025 17:30		JC/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VX048547.D	10 Nov 2025 17:51		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX111125

Review By	John Carlone	Review On	11/12/2025 8:11:22 AM
Supervise By	Maresh Dadoda	Supervise On	11/12/2025 11:43:18 AM
SubDirectory	VX111125	HP Acquire Method	HP Processing Method 82X110725W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136142		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136143,VP136144		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048548.D	11 Nov 2025 08:09		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048549.D	11 Nov 2025 09:11	pH#Lot#V12668	JC/MD	Ok,M
3	VX1111MBL01	VX1111MBL01	VX048550.D	11 Nov 2025 09:41		JC/MD	Ok
4	VX1111WBL01	VX1111WBL01	VX048551.D	11 Nov 2025 10:02		JC/MD	Ok
5	VX1111WBS01	VX1111WBS01	VX048552.D	11 Nov 2025 10:35	BS failed low for com.#40	JC/MD	Ok,M
6	VX1111WBSD01	VX1111WBSD01	VX048553.D	11 Nov 2025 11:03		JC/MD	Ok,M
7	IBLK	IBLK	VX048554.D	11 Nov 2025 11:23		JC/MD	Ok
8	Q3584-04	EB-2	VX048555.D	11 Nov 2025 11:43	vial B pH<2	JC/MD	Ok
9	Q3584-07	SEEP-1	VX048556.D	11 Nov 2025 12:04	vial A pH<2	JC/MD	Ok
10	Q3569-02DL	MW-12DL	VX048557.D	11 Nov 2025 12:24	vial B pH<2	JC/MD	Ok
11	Q3574-01	304641-01 Liquid	VX048558.D	11 Nov 2025 12:45	Surrogate fail;Run @ Lower Dilution	JC/MD	Not Ok
12	Q3574-01	304641-01 Liquid	VX048559.D	11 Nov 2025 13:25	Surrogate fail;Need straight run	JC/MD	Not Ok
13	IBLK	IBLK	VX048560.D	11 Nov 2025 13:45		JC/MD	Ok
14	VSTDCCC050	VSTDCCC050EC	VX048561.D	11 Nov 2025 15:06		JC/MD	Ok,M

M : Manual Integration