

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG NO.: Q3584
 Lab File ID: VX048527.D Date Analyzed: 11/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:36
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127604	12.00€				
UPPER LIMIT	255208	12.50€				
LOWER LIMIT	63802	11.50€				
EPA SAMPLE NO.						
SW-1	185006	12.01				
SW-2	195935	12.01				
SW-3	204501	12.01				
FB-2	142954	12.01				
TB-3	201653	12.01				
VX1110WBL01	160725	12.01				
VX1110WBS01	112690	12.01				
VX1110WBSD01	139918	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.