

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG NO.: Q3584
 Lab File ID: VX048549.D Date Analyzed: 11/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:11
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	129793	12.00€				
UPPER LIMIT	259586	12.50€				
LOWER LIMIT	64896.5	11.50€				
EPA SAMPLE NO.						
EB-2	202689	12.01				
SEEP-1	225901	12.01				
VX1111WBL01	195630	12.01				
VX1111WBS01	141984	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.