

DATA PACKAGE

SUB - DATA

PROJECT NAME : EDISON LANDFILL

REMINGTON & VERNICK ENGINEERS

2059 Springdale Road

Cherry Hill, NJ - 08003

Phone No: 856-795-9595

ORDER ID : Q3585

ATTENTION : Kyle Carlson



Cover Page

Order ID : Q3585

Project ID : Edison Landfill

Client : Remington & Vernick Engineers

Lab Sample Number

Q3585-04
Q3585-05
Q3585-06
Q3585-07
Q3585-08
Q3585-09

Client Sample Number

SW-1
SW-2
SW-3
EB-2
FB-2
SEEP-1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 11/20/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

December 03, 2025

Yazmeen Gomez
ATG - NEWARK LAB
284 Sheffield Street
Mountainside, NJ 07092
TEL:
FAX:
RE: Q3585

Order No.: 25110874

Dear Yazmeen Gomez:

Alliance Technical Group - Akron received 6 sample(s) on 11/12/2025 for the analyses presented in the following report.

There were no problems with the analytical events associated with this report unless noted in the Case Narrative.

Quality control data is within laboratory defined or method specified acceptance limits except where noted.

If you have any questions regarding these tests results, please feel free to call the laboratory.

Sincerely,

A handwritten signature in black ink, appearing to read "Jennifer Woolf". The signature is fluid and cursive, with the first name "Jennifer" written in a larger, more prominent script than the last name "Woolf".

Jennifer Woolf

Project Manager

3310 Win St.
Cuyahoga Falls, Ohio 44223

Arkansas 88-0735, California 2943, Colorado, Connecticut PH-0828, Florida NELAC E87688, Idaho OH00923, Illinois 200061, Indiana C-OH-13, ISO/IEC 17025:2017 119125 L22-544, Kansas E-10347, Kentucky (Underground Storage Tank) 3, Kentucky 90146, Maryland 339, Michigan 9988, Minnesota 1780279, Nevada OH009232020-1, New Hampshire 2996, New Jersey OH006, New York 11777, North Carolina 39705 and 631, North Dakota R-201, Ohio DW, Ohio VAP CL0052, Oklahoma 2019-155, Oregon OH200001, Pennsylvania 68-01335, Rhode Island LA000317, South Carolina 92016001, Texas T104704466-19-16, Utah OH009232020-12, Virginia VELAP 10381, West Virginia 9957C



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Case Narrative

WO#: 25110874
Date: 12/3/2025

CLIENT: ATG - NEWARK LAB

Project: Q3585

WorkOrder Narrative:

25110874: This report in its entirety consists of the following documents: Cover Letter, Case Narrative, Analytical Results, QC Summary Report, Applicable Accreditation Information, Chain-of-Custody, Cooler Receipt Form, and other applicable forms as necessary. All documents contain the Alliance Technical Group Work Order Number assigned to this report.

Alliance Technical Group holds the accreditations/certifications listed at the bottom of the cover letter that may or may not pertain to this report. Please refer to the "Accreditation Program Analytes Report" for accredited analytes list.

The information contained in this analytical report is the sole property of Alliance Technical Group and that of the customer. It cannot be reproduced in any form without the consent of Alliance Technical Group or the customer for which this report was issued. The results contained in this report are only representative of the samples received. Conditions can vary at different times and at different sampling conditions. Alliance Technical Group is not responsible for use or interpretation of the data included herein.

All results for solid samples are reported on an "as received" or "wet weight" basis unless indicated as "dry weight" using the "-dry" designation on the reporting units.

This report is believed to meet all of the requirements of the accrediting agency, where applicable. Any comments or problems with the analytical events associated with this report are noted below.

WorkOrder Comments:

SED-1, SED-2, SED-3 reported on 25110871

Analytical Sequence Sample Notes:

25110874-008A SVOC-EPA1633_NPW: Sample exhibited low recoveries for multiple IDAs. Potential high bias in associated analytes. Confirmed via extract dilution analysis.

Original

These commonly used Qualifiers and Acronyms may or may not be present in this report.

Qualifiers

U	The compound was analyzed for but was not detected above the MDL.
J	The reported value is greater than the Method Detection Limit but less than the Reporting Limit.
H	The hold time for sample preparation and/or analysis was exceeded. Not Clean Water Act compliant.
E	The result exceeded the linear range of the calibration or is estimated due to interference.
MC	The result is below the Minimum Compound Limit.
*	The result exceeds the Regulatory Limit or Maximum Contamination Limit.
m	Manual integration was used to determine the area response.
D	Manual integration in which peak was deleted.
N	The result is presumptive based on a Mass Spectral library search assuming a 1:1 response.
P	The second column confirmation exceeded 25% difference.
C	The result has been confirmed by GC/MS.
X	The result was not confirmed when GC/MS Analysis was performed.
B	The analyte was detected in the Method Blank at a concentration greater than the RL.
MB+	The analyte was detected in the Method Blank at a concentration greater than the MDL.
G	The ICB or CCB contained reportable amounts of analyte.
QC-/+	The CCV recovery failed low (-) or high (+).
R/QDR	The RPD was outside of accepted recovery limits.
QL-/+	The LCS or LCSD recovery failed low (-) or high (+).
QLR	The LCS/LCSD RPD was outside of accepted recovery limits.
QM-/+	The MS or MSD recovery failed low (-) or high (+).
QMR	The MS/MSD RPD was outside of accepted recovery limits.
QV-/+	The ICV recovery failed low (-) or high (+).
S	The spike result was outside of accepted recovery limits.
W	Samples were received outside temperature limits (0° – 6° C). Not Clean Water Act compliant.
Z	A deviation from the method was performed; Please refer to the Case Narrative for additional information.

Acronyms

ND	Not Detected
QC	Quality Control
MB	Method Blank
LCS	Laboratory Control Sample
LCS	Laboratory Control Sample Duplicate
QCS	Quality Control Sample
DUP	Duplicate
MS	Matrix Spike
MSD	Matrix Spike Duplicate
RPD	Relative Percent Different
ICV	Initial Calibration Verification
ICB	Initial Calibration Blank
CCV	Continuing Calibration Verification
CCB	Continuing Calibration Blank
RLC	Reporting Limit Check
DF	Dilution Factor

RL	Reporting Limit
MDL	Method Detection Limit
LOD	Level of Detection
LOQ	Level of Quantitation
PQL	Practical Quantitation Limit
CRQL	Contract Required Quantitation Limit
PL	Permit Limit
RegLvl	Regulatory Limit
MCL	Maximum Contamination Limit
MinCL	Minimum Compound Limit
RA	Reanalysis
RE	Reextraction
TIC	Tentatively Identified Compound
RT	Retention Time
CF	Calibration Factor
RF	Response Factor

This list of Qualifiers and Acronyms reflects the most commonly utilized Qualifiers and Acronyms for reporting. Please refer to the Analytical Notes in the Case Narrative for any Qualifiers or Acronyms that do not appear in this list or for additional information regarding the use of these Qualifiers on reported data.



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Workorder Sample Summary

WO#: 25110874
03-Dec-25

CLIENT: ATG - NEWARK LAB
Project: Q3585

Lab SampleID	Client Sample ID	Tag No	Date Collected	Date Received	Matrix
25110874-004	SW-1		11/6/2025 2:15:00 PM	11/12/2025 10:40:00 AM	Non-Potable Water
25110874-005	SW-2		11/7/2025 8:30:00 AM	11/12/2025 10:40:00 AM	Non-Potable Water
25110874-006	SW-3		11/7/2025 8:15:00 AM	11/12/2025 10:40:00 AM	Non-Potable Water
25110874-007	EB-2		11/6/2025 3:25:00 PM	11/12/2025 10:40:00 AM	Non-Potable Water
25110874-008	SEEP-1		11/7/2025 10:40:00 AM	11/12/2025 10:40:00 AM	Non-Potable Water
25110874-009	FB-2			11/12/2025 10:40:00 AM	DI Water



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

DATES REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

Sample ID	Client Sample ID	Collection Date	Matrix	Test Name	Leachate Date	Prep Date	Analysis Date
25110874-004A	SW-1	11/6/2025 2:15:00 PM	Non-Potable Water	PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 9:12:00 AM
				TSS for 1633			11/21/2025 3:00:00 PM
25110874-005A	SW-2	11/7/2025 8:30:00 AM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 9:27:00 AM
				TSS for 1633			11/21/2025 3:00:00 PM
25110874-006A	SW-3	11/7/2025 8:15:00 AM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 9:41:00 AM
				TSS for 1633			11/21/2025 3:00:00 PM
25110874-007A	EB-2	11/6/2025 3:25:00 PM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 9:56:00 AM
				TSS for 1633			11/14/2025 4:05:00 PM
25110874-008A	SEEP-1	11/7/2025 10:40:00 AM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 8:45:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/17/2025 8:32:00 AM	11/21/2025 10:11:00 AM
				TSS for 1633			11/14/2025 4:05:00 PM
25110874-009A	FB-2		DI Water	PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/25/2025 7:48:00 AM	12/2/2025 5:33:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/25/2025 7:48:00 AM	12/1/2025 11:17:00 PM
				TSS for 1633			11/21/2025 3:00:00 PM

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/6/2025 2:15:00 PM

Project: Q3585

Lab ID: 25110874-004

Matrix: NON-POTABLE WATER

Client Sample ID: SW-1

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
11CI-PF3OUdS	ND	0.204	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
3:3 FTCA	ND	0.431	9.35	U	ng/L	1	11/21/2025 9:12:00 AM
4:2 FTS	ND	1.60	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
5:3 FTCA	ND	1.62	46.8	U	ng/L	1	11/21/2025 9:12:00 AM
6:2 FTS	ND	2.45	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
7:3 FTCA	ND	2.86	46.8	U	ng/L	1	11/21/2025 9:12:00 AM
8:2 FTS	ND	2.88	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
9CL-PF3ONS	ND	0.281	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
ADONA	ND	0.319	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
HFPO-DA	1.11	0.390	7.48	J	ng/L	1	11/21/2025 9:12:00 AM
NEtFOSA	ND	0.342	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
NEtFOSAA	ND	0.950	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
NEtFOSE	ND	3.29	18.7	U	ng/L	1	11/21/2025 9:12:00 AM
NFDHA	ND	0.272	3.74	U	ng/L	1	11/21/2025 9:12:00 AM
NMeFOSA	ND	0.573	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
NMeFOSAA	ND	0.833	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
NMeFOSE	ND	3.53	18.7	U	ng/L	1	11/21/2025 9:12:00 AM
PFBA	ND	0.768	7.48	U	ng/L	1	11/21/2025 9:12:00 AM
PFBS	3.11	0.456	1.87		ng/L	1	11/21/2025 9:12:00 AM
PFDA	ND	0.597	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFDoA	ND	0.537	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFDoS	ND	0.374	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFDS	ND	0.385	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFEESA	ND	0.226	3.74	U	ng/L	1	11/21/2025 9:12:00 AM
PFHpA	3.18	0.627	1.87		ng/L	1	11/21/2025 9:12:00 AM
PFHpS	ND	0.597	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFHxA	6.14	0.379	1.87		ng/L	1	11/21/2025 9:12:00 AM
PFHxS	2.14	1.14	1.87		ng/L	1	11/21/2025 9:12:00 AM
PFMBA	ND	0.259	3.74	U	ng/L	1	11/21/2025 9:12:00 AM
PFMPA	ND	0.124	3.74	U	ng/L	1	11/21/2025 9:12:00 AM
PFNA	1.11	0.354	1.87	J	ng/L	1	11/21/2025 9:12:00 AM
PFNS	ND	0.761	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFOA	8.97	0.500	1.87		ng/L	1	11/21/2025 9:12:00 AM
PFOS	5.42	1.23	1.87		ng/L	1	11/21/2025 9:12:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/6/2025 2:15:00 PM

Project: Q3585

Lab ID: 25110874-004

Matrix: NON-POTABLE WATER

Client Sample ID: SW-1

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
PFOSA	0.627	0.276	1.87	J	ng/L	1	11/21/2025 9:12:00 AM
PFPeA	5.70	0.545	3.74		ng/L	1	11/21/2025 9:12:00 AM
PFPeS	ND	0.626	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFTeDA	ND	0.405	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFTrDA	ND	0.311	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
PFUnA	ND	0.669	1.87	U	ng/L	1	11/21/2025 9:12:00 AM
Surr: 13C2-PFDA	96.3	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 13C2-PFHxA	89.8	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 13C3-PFBA	55.2	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 13C4-PFOA	100	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 13C4-PFOS	100	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 13C5-PFNA	97.4	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM
Surr: 18O2-PFHxS	110	50-200	50-200		%Rec	1	11/21/2025 9:12:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 8:30:00 AM

Project: Q3585

Lab ID: 25110874-005

Matrix: NON-POTABLE WATER

Client Sample ID: SW-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
11CI-PF3OUdS	ND	0.208	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
3:3 FTCA	ND	0.438	9.52	U	ng/L	1	11/21/2025 9:27:00 AM
4:2 FTS	ND	1.63	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
5:3 FTCA	ND	1.64	47.6	U	ng/L	1	11/21/2025 9:27:00 AM
6:2 FTS	ND	2.50	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
7:3 FTCA	ND	2.91	47.6	U	ng/L	1	11/21/2025 9:27:00 AM
8:2 FTS	ND	2.93	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
9CL-PF3ONS	ND	0.286	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
ADONA	ND	0.325	7.62	U	ng/L	1	11/21/2025 9:27:00 AM
HFPO-DA	2.19	0.397	7.62	J	ng/L	1	11/21/2025 9:27:00 AM
NEtFOSA	ND	0.348	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
NEtFOSAA	ND	0.967	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
NEtFOSE	ND	3.35	19.0	U	ng/L	1	11/21/2025 9:27:00 AM
NFDHA	ND	0.277	3.81	U	ng/L	1	11/21/2025 9:27:00 AM
NMeFOSA	ND	0.583	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
NMeFOSAA	ND	0.849	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
NMeFOSE	ND	3.59	19.0	U	ng/L	1	11/21/2025 9:27:00 AM
PFBA	3.51	0.782	7.62	J	ng/L	1	11/21/2025 9:27:00 AM
PFBS	1.81	0.464	1.90	J	ng/L	1	11/21/2025 9:27:00 AM
PFDA	ND	0.608	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFDoA	ND	0.547	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFDoS	ND	0.381	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFDS	ND	0.392	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFEESA	ND	0.230	3.81	U	ng/L	1	11/21/2025 9:27:00 AM
PFHpA	2.27	0.638	1.90		ng/L	1	11/21/2025 9:27:00 AM
PFHpS	ND	0.608	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFHxA	4.53	0.386	1.90		ng/L	1	11/21/2025 9:27:00 AM
PFHxS	1.54	1.16	1.90	J	ng/L	1	11/21/2025 9:27:00 AM
PFMBA	ND	0.264	3.81	U	ng/L	1	11/21/2025 9:27:00 AM
PFMPA	ND	0.127	3.81	U	ng/L	1	11/21/2025 9:27:00 AM
PFNA	0.857	0.360	1.90	J	ng/L	1	11/21/2025 9:27:00 AM
PFNS	ND	0.775	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFOA	6.26	0.509	1.90		ng/L	1	11/21/2025 9:27:00 AM
PFOS	4.07	1.25	1.90		ng/L	1	11/21/2025 9:27:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 8:30:00 AM

Project: Q3585

Lab ID: 25110874-005

Matrix: NON-POTABLE WATER

Client Sample ID: SW-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
PFOSA	0.533	0.281	1.90	J	ng/L	1	11/21/2025 9:27:00 AM
PFPeA	3.79	0.555	3.81	J	ng/L	1	11/21/2025 9:27:00 AM
PFPeS	ND	0.637	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFTeDA	ND	0.412	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFTrDA	ND	0.316	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
PFUnA	ND	0.682	1.90	U	ng/L	1	11/21/2025 9:27:00 AM
Surr: 13C2-PFDA	81.9	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 13C2-PFHxA	81.4	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 13C3-PFBA	52.2	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 13C4-PFOA	82.9	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 13C4-PFOS	81.9	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 13C5-PFNA	75.9	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM
Surr: 18O2-PFHxS	85.6	50-200	50-200		%Rec	1	11/21/2025 9:27:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 8:15:00 AM

Project: Q3585

Lab ID: 25110874-006

Matrix: NON-POTABLE WATER

Client Sample ID: SW-3

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
11CI-PF3OUdS	ND	0.212	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
3:3 FTCA	ND	0.447	9.72	U	ng/L	1	11/21/2025 9:41:00 AM
4:2 FTS	ND	1.66	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
5:3 FTCA	ND	1.68	48.6	U	ng/L	1	11/21/2025 9:41:00 AM
6:2 FTS	ND	2.55	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
7:3 FTCA	ND	2.97	48.6	U	ng/L	1	11/21/2025 9:41:00 AM
8:2 FTS	ND	2.99	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
9CL-PF3ONS	ND	0.292	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
ADONA	ND	0.332	7.77	U	ng/L	1	11/21/2025 9:41:00 AM
HFPO-DA	1.88	0.405	7.77	J	ng/L	1	11/21/2025 9:41:00 AM
NEtFOSA	ND	0.355	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
NEtFOSAA	ND	0.987	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
NEtFOSE	ND	3.42	19.4	U	ng/L	1	11/21/2025 9:41:00 AM
NFDHA	ND	0.283	3.89	U	ng/L	1	11/21/2025 9:41:00 AM
NMeFOSA	ND	0.595	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
NMeFOSAA	ND	0.866	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
NMeFOSE	ND	3.67	19.4	U	ng/L	1	11/21/2025 9:41:00 AM
PFBA	3.21	0.798	7.77	J	ng/L	1	11/21/2025 9:41:00 AM
PFBS	1.59	0.473	1.94	J	ng/L	1	11/21/2025 9:41:00 AM
PFDA	ND	0.620	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFDoA	ND	0.558	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFDoS	ND	0.388	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFDS	ND	0.400	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFEESA	ND	0.235	3.89	U	ng/L	1	11/21/2025 9:41:00 AM
PFHpA	1.95	0.651	1.94		ng/L	1	11/21/2025 9:41:00 AM
PFHpS	ND	0.620	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFHxA	3.83	0.394	1.94		ng/L	1	11/21/2025 9:41:00 AM
PFHxS	1.67	1.19	1.94	J	ng/L	1	11/21/2025 9:41:00 AM
PFMBA	ND	0.270	3.89	U	ng/L	1	11/21/2025 9:41:00 AM
PFMPA	ND	0.129	3.89	U	ng/L	1	11/21/2025 9:41:00 AM
PFNA	1.48	0.368	1.94	J	ng/L	1	11/21/2025 9:41:00 AM
PFNS	ND	0.791	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFOA	9.21	0.520	1.94		ng/L	1	11/21/2025 9:41:00 AM
PFOS	5.43	1.27	1.94		ng/L	1	11/21/2025 9:41:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 8:15:00 AM

Project: Q3585

Lab ID: 25110874-006

Matrix: NON-POTABLE WATER

Client Sample ID: SW-3

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
PFOSA	0.826	0.287	1.94	J	ng/L	1	11/21/2025 9:41:00 AM
PFPeA	3.73	0.566	3.89	J	ng/L	1	11/21/2025 9:41:00 AM
PFPeS	ND	0.650	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFTeDA	ND	0.421	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFTrDA	ND	0.323	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
PFUnA	ND	0.696	1.94	U	ng/L	1	11/21/2025 9:41:00 AM
Surr: 13C2-PFDA	79.4	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 13C2-PFHxA	80.1	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 13C3-PFBA	51.9	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 13C4-PFOA	82.0	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 13C4-PFOS	80.5	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 13C5-PFNA	84.0	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM
Surr: 18O2-PFHxS	89.2	50-200	50-200		%Rec	1	11/21/2025 9:41:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/6/2025 3:25:00 PM

Project: Q3585

Lab ID: 25110874-007

Matrix: NON-POTABLE WATER

Client Sample ID: EB-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
11CI-PF3OUdS	ND	0.220	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
3:3 FTCA	0.544	0.464	10.1	J	ng/L	1	11/21/2025 9:56:00 AM
4:2 FTS	ND	1.72	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
5:3 FTCA	ND	1.74	50.4	U	ng/L	1	11/21/2025 9:56:00 AM
6:2 FTS	ND	2.64	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
7:3 FTCA	ND	3.08	50.4	U	ng/L	1	11/21/2025 9:56:00 AM
8:2 FTS	ND	3.10	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
9CL-PF3ONS	ND	0.303	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
ADONA	ND	0.344	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
HFPO-DA	ND	0.420	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
NEtFOSA	ND	0.368	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
NEtFOSAA	ND	1.02	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
NEtFOSE	ND	3.54	20.1	U	ng/L	1	11/21/2025 9:56:00 AM
NFDHA	ND	0.293	4.03	U	ng/L	1	11/21/2025 9:56:00 AM
NMeFOSA	ND	0.617	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
NMeFOSAA	ND	0.898	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
NMeFOSE	ND	3.80	20.1	U	ng/L	1	11/21/2025 9:56:00 AM
PFBA	ND	0.827	8.06	U	ng/L	1	11/21/2025 9:56:00 AM
PFBS	ND	0.491	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFDA	ND	0.643	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFDoA	ND	0.579	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFDoS	ND	0.403	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFDS	ND	0.415	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFEESA	ND	0.243	4.03	U	ng/L	1	11/21/2025 9:56:00 AM
PFHpA	ND	0.675	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFHpS	ND	0.643	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFHxA	ND	0.408	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFHxS	ND	1.23	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFMBA	ND	0.279	4.03	U	ng/L	1	11/21/2025 9:56:00 AM
PFMPA	ND	0.134	4.03	U	ng/L	1	11/21/2025 9:56:00 AM
PFNA	ND	0.381	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFNS	ND	0.820	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFOA	ND	0.539	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFOS	ND	1.32	2.01	U	ng/L	1	11/21/2025 9:56:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/6/2025 3:25:00 PM

Project: Q3585

Lab ID: 25110874-007

Matrix: NON-POTABLE WATER

Client Sample ID: EB-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
PFOSA	ND	0.297	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFPeA	ND	0.587	4.03	U	ng/L	1	11/21/2025 9:56:00 AM
PFPeS	ND	0.674	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFTeDA	ND	0.436	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFTrDA	ND	0.335	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
PFUnA	ND	0.721	2.01	U	ng/L	1	11/21/2025 9:56:00 AM
Surr: 13C2-PFDA	75.9	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 13C2-PFHxA	82.6	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 13C3-PFBA	51.6	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 13C4-PFOA	72.5	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 13C4-PFOS	79.7	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 13C5-PFNA	71.3	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM
Surr: 18O2-PFHxS	71.6	50-200	50-200		%Rec	1	11/21/2025 9:56:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 10:40:00 AM

Project: Q3585

Lab ID: 25110874-008

Matrix: NON-POTABLE WATER

Client Sample ID: SEEP-1

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
11CI-PF3OUdS	ND	0.216	7.92	U	ng/L	1	11/21/2025 10:11:00 AM
3:3 FTCA	31.6	0.456	9.90		ng/L	1	11/21/2025 8:45:00 PM
4:2 FTS	ND	1.69	7.92	U	ng/L	1	11/21/2025 10:11:00 AM
5:3 FTCA	267	1.71	49.5		ng/L	1	11/21/2025 10:11:00 AM
6:2 FTS	22.0	2.60	7.92		ng/L	1	11/21/2025 10:11:00 AM
7:3 FTCA	211	3.02	49.5		ng/L	1	11/21/2025 10:11:00 AM
8:2 FTS	10.8	3.05	7.92		ng/L	1	11/21/2025 10:11:00 AM
9CL-PF3ONS	ND	0.297	7.92	U	ng/L	1	11/21/2025 10:11:00 AM
ADONA	ND	0.338	7.92	U	ng/L	1	11/21/2025 10:11:00 AM
HFPO-DA	16.8	0.413	7.92		ng/L	1	11/21/2025 10:11:00 AM
NEtFOSA	5.76	0.362	1.98		ng/L	1	11/21/2025 10:11:00 AM
NEtFOSAA	152	1.01	1.98		ng/L	1	11/21/2025 10:11:00 AM
NEtFOSE	27.5	3.48	19.8		ng/L	1	11/21/2025 10:11:00 AM
NFDHA	ND	0.288	3.96	U	ng/L	1	11/21/2025 10:11:00 AM
NMeFOSA	1.82	0.606	1.98	J	ng/L	1	11/21/2025 10:11:00 AM
NMeFOSAA	26.1	0.883	1.98		ng/L	1	11/21/2025 10:11:00 AM
NMeFOSE	7.36	3.74	19.8	J	ng/L	1	11/21/2025 10:11:00 AM
PFBA	401	0.813	7.92		ng/L	1	11/21/2025 10:11:00 AM
PFBS	44.8	0.482	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFDA	13.5	0.632	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFDoA	0.723	0.569	1.98	J	ng/L	1	11/21/2025 10:11:00 AM
PFDoS	ND	0.396	1.98	U	ng/L	1	11/21/2025 10:11:00 AM
PFDS	0.901	0.408	1.98	J	ng/L	1	11/21/2025 10:11:00 AM
PFEESA	ND	0.239	3.96	U	ng/L	1	11/21/2025 10:11:00 AM
PFHpA	165	0.664	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFHpS	16.1	0.632	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFHxA	395	0.401	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFHxS	310	1.21	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFMBA	0.891	0.275	3.96	J	ng/L	1	11/21/2025 10:11:00 AM
PFMPA	ND	0.132	3.96	U	ng/L	1	11/21/2025 10:11:00 AM
PFNA	29.1	0.375	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFNS	ND	0.806	1.98	U	ng/L	1	11/21/2025 10:11:00 AM
PFOA	711	0.529	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFOS	459	1.30	1.98		ng/L	1	11/21/2025 10:11:00 AM

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date: 11/7/2025 10:40:00 AM

Project: Q3585

Lab ID: 25110874-008

Matrix: NON-POTABLE WATER

Client Sample ID: SEEP-1

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633		EPA 1633		Analyst: JJF
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)					DRAFT - REV. 4		
PFOSA	87.3	0.292	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFPeA	283	0.577	3.96		ng/L	1	11/21/2025 10:11:00 AM
PFPeS	32.3	0.662	1.98		ng/L	1	11/21/2025 10:11:00 AM
PFTeDA	ND	0.429	1.98	U	ng/L	1	11/21/2025 10:11:00 AM
PFTrDA	0.802	0.329	1.98	J	ng/L	1	11/21/2025 10:11:00 AM
PFUnA	1.76	0.709	1.98	J	ng/L	1	11/21/2025 10:11:00 AM
Surr: 13C2-PFDA	52.1	50-200	50-200		%Rec	1	11/21/2025 10:11:00 AM
Surr: 13C2-PFHxA	77.2	50-200	50-200		%Rec	1	11/21/2025 8:45:00 PM
Surr: 13C3-PFBA	50.6	50-200	50-200		%Rec	1	11/21/2025 8:45:00 PM
Surr: 13C4-PFOA	91.2	50-200	50-200		%Rec	1	11/21/2025 8:45:00 PM
Surr: 13C4-PFOS	81.7	50-200	50-200		%Rec	1	11/21/2025 10:11:00 AM
Surr: 13C5-PFNA	84.0	50-200	50-200		%Rec	1	11/21/2025 8:45:00 PM
Surr: 18O2-PFHxS	74.5	50-200	50-200		%Rec	1	11/21/2025 10:11:00 AM

Qualifiers:
H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date:

Project: Q3585

Lab ID: 25110874-009

Matrix: DI WATER

Client Sample ID: FB-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
11CI-PF3OUdS	ND	0.207	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
3:3 FTCA	ND	0.437	9.48	U	ng/L	1	12/1/2025 11:17:00 PM
4:2 FTS	ND	1.62	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
5:3 FTCA	ND	1.64	47.4	U	ng/L	1	12/1/2025 11:17:00 PM
6:2 FTS	ND	2.49	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
7:3 FTCA	ND	2.90	47.4	U	ng/L	1	12/1/2025 11:17:00 PM
8:2 FTS	ND	2.92	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
9CL-PF3ONS	ND	0.285	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
ADONA	ND	0.324	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
HFPO-DA	ND	0.395	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
NEtFOSA	ND	0.347	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
NEtFOSAA	ND	0.964	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
NEtFOSE	ND	3.34	19.0	U	ng/L	1	12/1/2025 11:17:00 PM
NFDHA	ND	0.276	3.79	U	ng/L	1	12/1/2025 11:17:00 PM
NMeFOSA	ND	0.581	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
NMeFOSAA	ND	0.845	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
NMeFOSE	ND	3.58	19.0	U	ng/L	1	12/1/2025 11:17:00 PM
PFBA	ND	0.779	7.59	U	ng/L	1	12/1/2025 11:17:00 PM
PFBS	ND	0.462	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFDA	ND	0.606	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFDoA	ND	0.545	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFDoS	ND	0.379	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFDS	ND	0.391	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFEESA	ND	0.229	3.79	U	ng/L	1	12/1/2025 11:17:00 PM
PFHpA	ND	0.636	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFHpS	ND	0.605	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFHxA	ND	0.384	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFHxS	ND	1.16	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFMBA	ND	0.263	3.79	U	ng/L	1	12/1/2025 11:17:00 PM
PFMPA	ND	0.126	3.79	U	ng/L	1	12/1/2025 11:17:00 PM
PFNA	ND	0.359	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFNS	ND	0.772	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFOA	ND	0.507	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFOS	ND	1.24	1.90	U	ng/L	1	12/1/2025 11:17:00 PM

Analyst: AEH

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Analytical Report

(consolidated)

WO#: 25110874

Date Reported: 12/3/2025

CLIENT: ATG - NEWARK LAB

Collection Date:

Project: Q3585

Lab ID: 25110874-009

Matrix: DI WATER

Client Sample ID: FB-2

Analyses	Result	MDL	RL	Qual	Units	DF	Date Analyzed
PFAS BY EPA 1633			EPA 1633			EPA 1633	
PFAS IN AQUEOUS SAMPLES (EPA METHOD 1633MOD)						DRAFT - REV. 4	
PFOSA	ND	0.280	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFPeA	ND	0.553	3.79	U	ng/L	1	12/1/2025 11:17:00 PM
PFPeS	ND	0.635	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFTeDA	ND	0.411	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFTrDA	ND	0.315	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
PFUnA	ND	0.679	1.90	U	ng/L	1	12/1/2025 11:17:00 PM
Surr: 13C2-PFDA	98.2	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 13C2-PFHxA	80.8	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 13C3-PFBA	79.9	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 13C4-PFOA	91.1	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 13C4-PFOS	87.8	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 13C5-PFNA	89.0	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM
Surr: 18O2-PFHxS	79.6	50-200	50-200		%Rec	1	12/1/2025 11:17:00 PM

Analyst: AEH

Qualifiers:
H Holding times for preparation or analysis exceeded
ND Not Detected
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
PL Permit Limit
U Samples with CalcVal < MDL
J Analyte detected below quantitation limits



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Accreditation Program Analytes Report

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

State: NJ

Program Name: NJ_DW_NPW_SC
M_NELAP

Test Name	Matrix	Analyte	Status
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	11Cl-PF3OUdS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	4:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	5:3 FTCA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	6:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	8:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	9CL-PF3ONS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	ADONA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	HFPO-DA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NEtFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NEtFOSAA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NEtFOSE	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NFDHA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NMeFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NMeFOSAA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	NMeFOSE	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFBA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFBS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFDoA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFDoS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFEESA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFHpA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFHpS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFHxA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFHxS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFMBA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFMPA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFNA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFNS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFOA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFOS	A



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Accreditation Program Analytes Report

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

State: NJ

Program Name: NJ_DW_NPW_SC
M_NELAP

Test Name	Matrix	Analyte	Status
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFPeA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFPeS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFTeDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFTTrDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	DI Water	PFUnA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	11Cl-PF3OUdS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	4:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	5:3 FTCA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	6:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	8:2 FTS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	9CL-PF3ONS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	ADONA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	HFPO-DA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NEtFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NEtFOSAA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NEtFOSE	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NFDHA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NMeFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NMeFOSAA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	NMeFOSE	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFBA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFBS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFDoA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFDoS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFEESA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFHpA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFHpS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFHxA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFHxS	A



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

Accreditation Program Analytes Report

WO#: 25110874
03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

State: NJ

Program Name: NJ_DW_NPW_SC
M_NELAP

Test Name	Matrix	Analyte	Status
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFMBA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFMPA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFNA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFNS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFOA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFOS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFOSA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFPeA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFPeS	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFTeDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFTTrDA	A
PFAS in Aqueous Samples (EPA Method 1633Mod)	Non-Potable Water	PFUnA	A



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 89954

Sample ID: MB-89954	SampType: MBLK	TestCode: SVOC-EPA16 Units: µg/L			Prep Date: 11/17/2025			RunNo: 221458			
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633		E1633	Analysis Date: 11/21/2025			SeqNo: 5906764			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11CI-PF3OUdS	ND	0.00800									U
3:3 FTCA	ND	0.0100									U
4:2 FTS	ND	0.00800									U
5:3 FTCA	ND	0.0500									U
6:2 FTS	ND	0.00800									U
7:3 FTCA	ND	0.0500									U
8:2 FTS	ND	0.00800									U
9CL-PF3ONS	ND	0.00800									U
ADONA	ND	0.00800									U
HFPO-DA	ND	0.00800									U
NEtFOSA	ND	0.00200									U
NEtFOSAA	ND	0.00200									U
NEtFOSE	ND	0.0200									U
NFDHA	ND	0.00400									U
NMeFOSA	ND	0.00200									U
NMeFOSAA	ND	0.00200									U
NMeFOSE	ND	0.0200									U
PFBA	ND	0.00800									U
PFBS	ND	0.00200									U
PFDA	ND	0.00200									U
PFDaA	ND	0.00200									U
PFDoS	ND	0.00200									U
PFDS	ND	0.00200									U

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 89954

Sample ID: MB-89954	SampType: MBLK	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/17/2025	RunNo: 221458						
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633	E1633	Analysis Date: 11/21/2025	SeqNo: 5906764						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFEESA	ND	0.00400									U
PFHpA	ND	0.00200									U
PFHpS	ND	0.00200									U
PFHxA	ND	0.00200									U
PFHxS	ND	0.00200									U
PFMBA	ND	0.00400									U
PFMPA	ND	0.00400									U
PFNA	ND	0.00200									U
PFNS	ND	0.00200									U
PFOA	ND	0.00200									U
PFOS	ND	0.00200									U
PFOSA	ND	0.00200									U
PFPeA	ND	0.00400									U
PFPeS	ND	0.00200									U
PFTeDA	ND	0.00200									U
PFTTrDA	ND	0.00200									U
PFUnA	ND	0.00200									U
Surr: 13C2-PFDA	0.0264		0.02500		106	50	200				
Surr: 13C2-PFHxA	0.0543		0.05000		109	50	200				
Surr: 13C3-PFBA	0.0903		0.1000		90.3	50	200				
Surr: 13C4-PFOA	0.0524		0.05000		105	50	200				
Surr: 13C4-PFOS	0.0490		0.04790		102	50	200				
Surr: 13C5-PFNA	0.0263		0.02500		105	50	200				

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 89954

Sample ID: MB-89954	SampType: MBLK	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/17/2025	RunNo: 221458						
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633	E1633	Analysis Date: 11/21/2025	SeqNo: 5906764						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Surr: 18O2-PFHxS	0.0492		0.04740		104	50	200				

Sample ID: LCS-89954	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/17/2025	RunNo: 221458						
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633	E1633	Analysis Date: 11/21/2025	SeqNo: 5906765						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11Cl-PF3OUdS	0.305	0.00800	0.3772	0	80.7	55	160				
3:3 FTCA	0.408	0.0100	0.5000	0	81.5	65	130				
4:2 FTS	0.408	0.00800	0.3748	0	109	70	145				
5:3 FTCA	2.67	0.0500	2.500	0	107	70	135				
6:2 FTS	0.369	0.00800	0.3804	0	97.0	65	155				
7:3 FTCA	2.48	0.0500	2.500	0	99.2	50	145				
8:2 FTS	0.340	0.00800	0.3840	0	88.5	60	150				
9CL-PF3ONS	0.373	0.00800	0.3732	0	100	70	155				
ADONA	0.410	0.00800	0.3780	0	108	65	145				
HFPO-DA	0.386	0.00800	0.4000	0	96.5	70	140				
NEtFOSA	0.100	0.00200	0.1000	0	100	65	145				
NEtFOSAA	0.104	0.00200	0.1000	0	104	70	145				
NEtFOSE	0.946	0.0200	1.000	0	94.6	70	135				
NFDHA	0.185	0.00400	0.2000	0	92.4	50	150				
NMeFOSA	0.104	0.00200	0.1000	0	104	60	150				
NMeFOSAA	0.108	0.00200	0.1000	0	108	50	140				

Qualifiers:	H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits	M	Manual Integration used to determine area response
	ND	Not Detected	PL	Permit Limit	RL	Reporting Detection Limit
	S	Spike Recovery outside accepted recovery limits	U	Samples with CalcVal < MDL	W	Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 89954

Sample ID: LCS-89954	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/17/2025	RunNo: 221458						
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633	E1633	Analysis Date: 11/21/2025	SeqNo: 5906765						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
NMeFOSE	0.978	0.0200	1.000	0	97.8	70	145				
PFBA	0.399	0.00800	0.4000	0	99.8	70	140				
PFBS	0.0912	0.00200	0.08880	0	103	60	145				
PFDA	0.0937	0.00200	0.1000	0	93.7	70	140				
PFDaA	0.0929	0.00200	0.1000	0	92.9	70	140				
PFDoS	0.0812	0.00200	0.09680	0	83.9	50	145				
PFDS	0.0945	0.00200	0.09640	0	98.0	60	145				
PFEESA	0.186	0.00400	0.1784	0	104	70	140				
PFHpA	0.0960	0.00200	0.1000	0	96.0	70	150				
PFHpS	0.102	0.00200	0.09520	0	108	70	150				
PFHxA	0.113	0.00200	0.1000	0	113	70	145				
PFHxS	0.0801	0.00200	0.09124	0	87.8	65	145				
PFMBA	0.185	0.00400	0.2000	0	92.4	60	150				
PFMPA	0.158	0.00400	0.2000	0	79.0	55	140				
PFNA	0.109	0.00200	0.1000	0	109	70	150				
PFNS	0.0976	0.00200	0.09600	0	102	56	145				
PFOA	0.0909	0.00200	0.1000	0	90.9	70	150				
PFOS	0.0898	0.00200	0.09280	0	96.7	55	150				
PFOSA	0.104	0.00200	0.1000	0	104	70	145				
PFPeA	0.204	0.00400	0.2000	0	102	65	135				
PFPeS	0.0832	0.00200	0.09400	0	88.5	65	140				
PFTeDA	0.0930	0.00200	0.1000	0	93.0	60	140				
PFTTrDA	0.108	0.00200	0.1000	0	108	65	140				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 89954

Sample ID: LCS-89954	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/17/2025	RunNo: 221458						
Client ID: BatchQC	Batch ID: 89954	TestNo: E1633	E1633	Analysis Date: 11/21/2025	SeqNo: 5906765						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFUnA	0.0929	0.00200	0.1000	0	92.9	70	145				
Surr: 13C2-PFDA	0.0224		0.02500		89.4	50	200				
Surr: 13C2-PFHxA	0.0397		0.05000		79.3	50	200				
Surr: 13C3-PFBA	0.0701		0.1000		70.1	50	200				
Surr: 13C4-PFOA	0.0421		0.05000		84.2	50	200				
Surr: 13C4-PFOS	0.0393		0.04790		82.0	50	200				
Surr: 13C5-PFNA	0.0203		0.02500		81.1	50	200				
Surr: 18O2-PFHxS	0.0377		0.04740		79.5	50	200				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 90195

Sample ID: MB-90195	SampType: MBLK	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920059						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11Cl-PF3OUdS	ND	0.00800									U
3:3 FTCA	ND	0.0100									U
4:2 FTS	ND	0.00800									U
5:3 FTCA	ND	0.0500									U
6:2 FTS	ND	0.00800									U
7:3 FTCA	ND	0.0500									U
8:2 FTS	ND	0.00800									U
9CL-PF3ONS	ND	0.00800									U
ADONA	ND	0.00800									U
HFPO-DA	ND	0.00800									U
NEtFOSA	ND	0.00200									U
NEtFOSAA	ND	0.00200									U
NEtFOSE	ND	0.0200									U
NFDHA	ND	0.00400									U
NMeFOSA	ND	0.00200									U
NMeFOSAA	ND	0.00200									U
NMeFOSE	ND	0.0200									U
PFBA	ND	0.00800									U
PFBS	ND	0.00200									U
PFDA	ND	0.00200									U
PFDaA	ND	0.00200									U
PFDoS	ND	0.00200									U
PFDS	ND	0.00200									U

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 90195

Sample ID: MB-90195	SampType: MBLK	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920059						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFEESA	ND	0.00400									U
PFHpA	ND	0.00200									U
PFHpS	ND	0.00200									U
PFHxA	ND	0.00200									U
PFHxS	ND	0.00200									U
PFMBA	ND	0.00400									U
PFMPA	ND	0.00400									U
PFNA	ND	0.00200									U
PFNS	ND	0.00200									U
PFOA	ND	0.00200									U
PFOS	ND	0.00200									U
PFOSA	ND	0.00200									U
PFPeA	ND	0.00400									U
PFPeS	ND	0.00200									U
PFTeDA	ND	0.00200									U
PFTTrDA	ND	0.00200									U
PFUnA	ND	0.00200									U
Surr: 13C2-PFDA	0.0218		0.02500		87.0	50	200				
Surr: 13C2-PFHxA	0.0515		0.05000		103	50	200				
Surr: 13C3-PFBA	0.0988		0.1000		98.8	50	200				
Surr: 13C4-PFOA	0.0513		0.05000		103	50	200				
Surr: 13C4-PFOS	0.0430		0.04790		89.8	50	200				
Surr: 13C5-PFNA	0.0242		0.02500		97.0	50	200				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 90195

Sample ID: MB-90195	SampType: MBLK	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920059						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Surr: 18O2-PFHxS	0.0460		0.04740		97.0	50	200				

Sample ID: LCS-90195	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920060						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11Cl-PF3OUdS	0.318	0.00800	0.3772	0	84.2	55	160				
3:3 FTCA	0.382	0.0100	0.5000	0	76.3	65	130				
4:2 FTS	0.342	0.00800	0.3748	0	91.2	70	145				
5:3 FTCA	2.31	0.0500	2.500	0	92.5	70	135				
6:2 FTS	0.337	0.00800	0.3804	0	88.7	65	155				
7:3 FTCA	2.25	0.0500	2.500	0	89.9	50	145				
8:2 FTS	0.370	0.00800	0.3840	0	96.3	60	150				
9CL-PF3ONS	0.344	0.00800	0.3732	0	92.3	70	155				
ADONA	0.370	0.00800	0.3780	0	97.9	65	145				
HFPO-DA	0.370	0.00800	0.4000	0	92.5	70	140				
NEtFOSA	0.0953	0.00200	0.1000	0	95.3	65	145				
NEtFOSAA	0.0944	0.00200	0.1000	0	94.4	70	145				
NEtFOSE	0.971	0.0200	1.000	0	97.1	70	135				
NFDHA	0.184	0.00400	0.2000	0	92.1	50	150				
NMeFOSA	0.100	0.00200	0.1000	0	100	60	150				
NMeFOSAA	0.100	0.00200	0.1000	0	100	50	140				

Qualifiers:	H	Holding times for preparation or analysis exceeded	J	Analyte detected below quantitation limits	M	Manual Integration used to determine area response
	ND	Not Detected	PL	Permit Limit	RL	Reporting Detection Limit
	S	Spike Recovery outside accepted recovery limits	U	Samples with CalcVal < MDL	W	Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 90195

Sample ID: LCS-90195	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920060						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
NMeFOSE	0.910	0.0200	1.000	0	91.0	70	145				
PFBA	0.373	0.00800	0.4000	0	93.2	70	140				
PFBS	0.0828	0.00200	0.08880	0	93.3	60	145				
PFDA	0.0857	0.00200	0.1000	0	85.7	70	140				
PFDaA	0.0921	0.00200	0.1000	0	92.1	70	140				
PFDoS	0.0755	0.00200	0.09680	0	78.0	50	145				
PFDS	0.0815	0.00200	0.09640	0	84.6	60	145				
PFEESA	0.173	0.00400	0.1784	0	97.0	70	140				
PFHpA	0.0853	0.00200	0.1000	0	85.3	70	150				
PFHpS	0.0846	0.00200	0.09520	0	88.8	70	150				
PFHxA	0.0899	0.00200	0.1000	0	89.9	70	145				
PFHxS	0.0837	0.00200	0.09124	0	91.7	65	145				
PFMBA	0.160	0.00400	0.2000	0	79.8	60	150				
PFMPA	0.154	0.00400	0.2000	0	77.1	55	140				
PFNA	0.0845	0.00200	0.1000	0	84.5	70	150				
PFNS	0.0931	0.00200	0.09600	0	96.9	56	145				
PFOA	0.0794	0.00200	0.1000	0	79.4	70	150				
PFOS	0.0758	0.00200	0.09280	0	81.7	55	150				
PFOSA	0.0925	0.00200	0.1000	0	92.5	70	145				
PFPeA	0.181	0.00400	0.2000	0	90.3	65	135				
PFPeS	0.0857	0.00200	0.09400	0	91.2	65	140				
PFTeDA	0.0861	0.00200	0.1000	0	86.1	60	140				
PFTTrDA	0.0950	0.00200	0.1000	0	95.0	65	140				

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: 90195

Sample ID: LCS-90195	SampType: LCS	TestCode: SVOC-EPA16	Units: µg/L	Prep Date: 11/25/2025	RunNo: 221976						
Client ID: BatchQC	Batch ID: 90195	TestNo: E1633	E1633	Analysis Date: 12/1/2025	SeqNo: 5920060						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFUnA	0.101	0.00200	0.1000	0	101	70	145				
Surr: 13C2-PFDA	0.0288		0.02500		115	50	200				
Surr: 13C2-PFHxA	0.0502		0.05000		100	50	200				
Surr: 13C3-PFBA	0.104		0.1000		104	50	200				
Surr: 13C4-PFOA	0.0526		0.05000		105	50	200				
Surr: 13C4-PFOS	0.0516		0.04790		108	50	200				
Surr: 13C5-PFNA	0.0277		0.02500		111	50	200				
Surr: 18O2-PFHxS	0.0470		0.04740		99.1	50	200				

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: R221008

Sample ID: MB-R221008	SampType: MBLK	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:			RunNo: 221008			
Client ID: BatchQC	Batch ID: R221008	TestNo: A2540D				Analysis Date: 11/14/2025			SeqNo: 5893869			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	ND	2.50										U

Sample ID: LCS-R221008	SampType: LCS	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:			RunNo: 221008			
Client ID: BatchQC	Batch ID: R221008	TestNo: A2540D				Analysis Date: 11/14/2025			SeqNo: 5893870			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	97.5	25.0	100.0	0	97.5	85	115					

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110874

03-Dec-25

Client: ATG - NEWARK LAB

Project: Q3585

BatchID: R221554

Sample ID: MB-R221554	SampType: MBLK	TestCode: TSS_NPW(25 Units: mg/L			Prep Date:			RunNo: 221554			
Client ID: BatchQC	Batch ID: R221554	TestNo: A2540D			Analysis Date: 11/21/2025			SeqNo: 5909101			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
TSS	ND	2.50									U

Sample ID: LCS-R221554	SampType: LCS	TestCode: TSS_NPW(25 Units: mg/L			Prep Date:			RunNo: 221554			
Client ID: BatchQC	Batch ID: R221554	TestNo: A2540D			Analysis Date: 11/21/2025			SeqNo: 5909102			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
TSS	95.5	25.0	100.0	0	95.5	85	115				

Sample ID: 25111516-001ADUP	SampType: DUP	TestCode: TSS_NPW(25 Units: mg/L			Prep Date:			RunNo: 221554			
Client ID: BatchQC	Batch ID: R221554	TestNo: A2540D			Analysis Date: 11/21/2025			SeqNo: 5909104			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
TSS	1540	25.0						1518	1.24	5	

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

Original

CLIENT INFORMATION			CLIENT PROJECT INFORMATION			CLIENT BILLING INFORMATION											
REPORT TO BE SENT TO: COMPANY: Remington + Vernice Engineers ADDRESS: 2059 Springdale Rd CITY: Cherry Hill NJ STATE: NJ ZIP: 08003 ATTENTION: Kyle Carlson @rve.com PHONE: 609-682-3049 FAX:			PROJECT NAME: Edison Landfill PROJECT NO.: 12057007 LOCATION: Edison, NJ PROJECT MANAGER: Leslie Hartzell e-mail: Leslie.Hartzell@rve.com PHONE: 610-663-4665 FAX:			BILL TO: APinvoices@rve.com PO#: 12057007 ADDRESS: 2059 Springdale Rd CITY: Cherry Hill STATE: NJ ZIP: 08003 ATTENTION: Leslie Hartzell PHONE: 610-663-4665											
DATA TURNAROUND INFORMATION FAX (RUSH) _____ DAYS* HARDCOPY (DATA PACKAGE): Standard DAYS* EDD: _____ DAYS* *TO BE APPROVED BY CHEMTECH STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS			DATA DELIVERABLE INFORMATION ☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data) ☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP ☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B + Raw Data ☐ Other _____ ☒ EDD FORMAT NJ HAZ SITE			ANALYSIS											
ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	SED-1	SOIL		X	11/6/25	14:15	3	3									
2.	SED-2	SOIL		X	11/6/25	12:40	3	3									
3.	SED-3	SOIL		X	11/6/25	11:55	3	3									
4.	SW-1	SURF. WATER		X	11/6/25	14:15	3	3									
5.	SW-2	SURF. WATER		X	11/7/25	8:30	3	3									
6.	SW-3	SURF. WATER		X	11/7/25	8:15	3	3									
7.	EB-2	(aq) Blank			11/6/25	15:25	3	3									
8.	FB-2	(aq) Blank			11/7/25	8:55	3	3									
9.	SEEP-1	GW		X	11/7/25	10:40	3	3									
10.																	
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY																	
RELINQUISHED BY SAMPLER:		DATE/TIME:		RECEIVED BY:		Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.5 + 0 = 3.5 °C Comments:											
1. Anna Maier		11/7/25		1.													
RELINQUISHED BY SAMPLER:		DATE/TIME:		RECEIVED BY:													
2.		11-7-25		2. JT													
RELINQUISHED BY SAMPLER:		DATE/TIME:		RECEIVED BY:													
3. JT		11-7-25		3.													
Page 1 of 1						CLIENT: ☐ Hand Delivered ☐ Other						Shipment Complete ☐ YES ☐ NO					



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
WWW.CHEMTECH.NET

25110874

CHAIN OF CUSTODY RECORD

Sub Lab INFORMATION	CLIENT PROJECT INFORMATION	CLIENT BILLING INFORMATION
COMPANY : Alliance Technical Group - Akron	ORDER ID : Q3585	BILL TO: CHEMTECH PO# : q3585
ADDRESS : 3310 Win Street	PROJECT ID: Edison Landfill	ADDRESS : 284, Sheffield Street
CITY: Cuyahoga Fal State : OH ZIP : 44223	PROJECT MANAGER YAZMEEN	CITY: Mountainside State : NJ ZIP : 07092
E-mail : jennifer.woolf@alliancetg.com	E-mail : YAZMEEN.GOMEZ@AllianceTG.com	ATTENTION : YAZMEE
PHONE : 330-253-8211	PHONE : (908) 789 8900 FAX: (908) 789 8922	PHONE : (908) 789 8900 FAX : (908) 789 8922

EDD : Excel NJ	Report : NJ Reduced	Comment : Grabs. NJ.
----------------	---------------------	----------------------

ID	CLIENT SAMPLE IDENTIFICATION	SAMPLE MATRIX	ANALYSIS	Preservative	Method	SAMPLE COLLECTION		# OF BOTTLES	TAT DAYS
						DATE	TIME		
01	SED-1	Water	PFAS	Cool 4 deg C	1633	11/06/2025	14:15:00	1	10
02	SED-2	Water	PFAS	Cool 4 deg C	1633	11/06/2025	12:40:00	1	10
03	SED-3	Water	PFAS	Cool 4 deg C	1633	11/06/2025	11:55:00	1	10
04	SW-1	Water	PFAS	Cool 4 deg C	1633	11/06/2025	14:15:00	1	10
05	SW-2	Water	PFAS	Cool 4 deg C	1633	11/07/2025	08:30:00	1	10
06	SW-3	Water	PFAS	Cool 4 deg C	1633	11/07/2025	08:15:00	1	10
07	EB-2	Water	PFAS	Cool 4 deg C	1633	11/06/2025	15:25:00	1	10
08		Water	PFAS	Cool 4 deg C	1633			1	10
09	SEEP-1	Water	PFAS	Cool 4 deg C	1633	11/07/2025	10:40:00	1	10

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGES POSSESSION INCLUDING COURIER DELIVERY				
RELINQUISHED BY SAMPLER:	DATETIME:	RECEIVED BY:	Conditions of bottles or Coolers at receipt:	
1. <i>[Signature]</i>	11/10/25 1:40	1. <i>[Signature]</i> 11/12/25 10:40	☐ Compliant ☐ Non Compliant Cooler Temp _____ Ice or Cooler? _____	
RELINQUISHED BY:	DATETIME:	RECEIVED BY:	<i>5.8 + 0.2 = 6.8</i>	
2.		2.		
RELINQUISHED BY:	DATETIME:	RECEIVED BY:	☐ OVERNIGHT Shipment Complete:	
3.		3.	☐ OVERNIGHT ☐ YES ☐ NO	

Sample Log-In Check List

Client Name: **CHE-NJ-07972**

Work Order Number: **25110874**

RcptNo: **1**

Logged by: **Spencer M. Hartwell** **11/12/2025 10:40:00 AM**

Spencer M. Hartwell

Completed By: **Spencer M. Hartwell** **11/12/2025 6:19:51 PM**

Spencer M. Hartwell

Reviewed By: **Jennifer Woolf** **11/25/2025 11:41:02 AM**

Jennifer M. Woolf

Chain of Custody

1. Is Chain of Custody complete? Yes ☐ No ☒ Not Present ☐
2. How was the sample delivered? UPS

Log In

3. Coolers are present? Yes ☒ No ☐ NA ☐
4. Shipping container/cooler in good condition? Yes ☒ No ☐
Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒ NA ☐
No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes ☒ No ☐ NA ☐
6. Were all samples received at a temperature of >0° C to 6.0°C Yes ☒ No ☐ NA ☐
7. Sample(s) in proper container(s)? Yes ☒ No ☐
8. Sufficient sample volume for indicated test(s)? Yes ☒ No ☐
9. Are samples (except VOA and ONG) properly preserved? Yes ☒ No ☐
10. Was preservative added to bottles? Yes ☐ No ☒ NA ☐
11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes ☐ No ☐ No VOA Vials ☒
12. Were any sample containers received broken? Yes ☐ No ☒
13. Does paperwork match bottle labels? Yes ☒ No ☐
(Note discrepancies on chain of custody)
14. Are matrices correctly identified on Chain of Custody? Yes ☒ No ☐
15. Is it clear what analyses were requested? Yes ☒ No ☐
16. Were all holding times able to be met? Yes ☒ No ☐
(If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes ☐ No ☐ NA ☒

Person Notified: Date:
By Whom: Via: ☐ eMail ☐ Phone ☐ Fax ☐ In Person
Regarding:
Client Instructions:

18. Additional remarks:

Received FB-2 not on chain:
Analyze per email
SED-1, SED-2, SED-3 analyzed on 25110871

Cooler Information



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Sample Log-In Check List

Client Name: **CHE-NJ-07972**

Work Order Number: **25110874**

RcptNo: **1**

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
1	6					