

Laboratory Control Sample Summary

Client: Remington & Vernick Engineers

SDG No.: Q3586

Project: Edison Landfill

Run No.: LB137839

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170464BS							
Hexavalent Chromium	mg/Kg	20	19.8		99	1	84-110	11/10/2025

Laboratory Control Sample Summary

Client: Remington & Vernick Engineers

SDG No.: Q3586

Project: Edison Landfill

Run No.: LB137840

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170470BS							
Cyanide	mg/Kg	5	4.90		98	1	85-115	11/10/2025