

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3586

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB170491BL		SOLID		Batch Number:	PB170491		Prep Date:	11/11/2025	
	Mercury	0.0080	<0.014	U	0.014	CV	11/11/2025	14:40	LB137851

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3586

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB170471BL	SOLID			Batch Number:	PB170471		Prep Date:	11/10/2025	
	Aluminum	0.84	<2.5	U	5.00	P	11/14/2025	15:00	LB137927
	Antimony	0.22	<1.25	U	2.50	P	11/14/2025	15:00	LB137927
	Arsenic	0.19	<0.5	U	1.00	P	11/14/2025	15:00	LB137927
	Barium	0.73	<2.5	U	5.00	P	11/14/2025	15:00	LB137927
	Beryllium	0.025	<0.15	U	0.30	P	11/14/2025	15:00	LB137927
	Cadmium	0.024	<0.15	U	0.30	P	11/14/2025	15:00	LB137927
	Calcium	11.1	<50	U	100	P	11/14/2025	15:00	LB137927
	Chromium	0.075	<0.25	J	0.50	P	11/14/2025	15:00	LB137927
	Cobalt	0.10	<0.75	U	1.50	P	11/14/2025	15:00	LB137927
	Copper	0.22	<0.5	U	1.00	P	11/14/2025	15:00	LB137927
	Iron	3.99	<2.5	U*	5.00	P	11/14/2025	15:00	LB137927
	Lead	0.13	<0.3	U	0.60	P	11/14/2025	15:00	LB137927
	Magnesium	12.0	<50	U	100	P	11/14/2025	15:00	LB137927
	Manganese	0.15	<0.5	J	1.00	P	11/14/2025	15:00	LB137927
	Nickel	0.13	<1	U	2.00	P	11/14/2025	15:00	LB137927
	Potassium	27.7	<50	U	100	P	11/14/2025	15:00	LB137927
	Selenium	0.26	<0.5	U	1.00	P	11/14/2025	15:00	LB137927
	Silver	0.12	<0.25	U	0.50	P	11/14/2025	15:00	LB137927
	Sodium	17.8	<50	U	100	P	11/14/2025	15:00	LB137927
	Thallium	0.23	<1	U	2.00	P	11/14/2025	15:00	LB137927
	Vanadium	0.25	<1	U	2.00	P	11/14/2025	15:00	LB137927
	Zinc	0.23	<1	U	2.00	P	11/14/2025	15:00	LB137927