

Continuing Calibration Report for SequenceID : FF111925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016745.D

Aromatic C10-C12	5063875.000	40.000	4.501	6.378	126596.875	160369.151	21.059
Aromatic C12-C16	7670696.000	60.000	6.379	9.081	127844.933	160326.403	20.260
Aromatic C16-C21	9715936.000	80.000	9.082	13.401	121449.200	156685.093	22.488
Aromatic C21-C36	20842553.000	180.000	13.402	18.862	115791.961	145594.493	20.470
Aromatic EPH	43293060.000	360.000	4.501	18.862	120258.500	152156.018	20.964