

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG NO.: Q3586

Instrument ID: ECD_O Date(s) Analyzed: 11/04/2025 11/04/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.55	4.45	4.65	84124000
		2	4.63	4.53	4.73	60183600
		3	4.71	4.61	4.81	193335000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.71	4.61	4.81	150015000
		2	5.22	5.12	5.32	72567400
		3	5.51	5.41	5.61	158675000
		4	5.67	5.57	5.77	73795600
		5	5.76	5.66	5.86	51580600
Aroclor-1262	500	1	7.70	7.60	7.80	453404000
		2	8.23	8.13	8.33	707198000
		3	8.53	8.43	8.63	451420000
		4	8.62	8.52	8.72	336186000
		5	9.24	9.14	9.34	241816000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG NO.: Q3586

Instrument ID: ECD_O Date(s) Analyzed: 11/04/2025 11/04/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.86	3.76	3.96	61799200
		2	3.95	3.85	4.05	45637000
		3	4.02	3.92	4.12	140257000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.02	3.92	4.12	111493000
		2	4.75	4.65	4.85	112095000
		3	4.92	4.82	5.02	59654200
		4	5.01	4.91	5.11	50889000
		5	5.18	5.08	5.28	56587600
Aroclor-1262	500	1	6.76	6.66	6.86	494206000
		2	7.26	7.16	7.36	795090000
		3	7.54	7.44	7.64	329028000
		4	7.61	7.51	7.71	538192000
		5	8.10	8.00	8.20	238020000