

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI02
SDG NO.: Q3586

Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4150980000	4075270000	4055470000	3840760000	4148570000	4054210000	3
4,4'-DDE	5435610000	5370750000	5317610000	5053060000	5456160000	5326640000	3
4,4'-DDT	4256810000	4152600000	4117930000	3927880000	4097130000	4110470000	3
Aldrin	6694230000	6562230000	6516940000	6238140000	6724820000	6547270000	3
alpha-BHC	7752530000	7402380000	7252210000	6743070000	6735520000	7177140000	6
alpha-Chlordane	5966070000	5905470000	5938650000	6170460000	6920630000	6180250000	7
beta-BHC	2515210000	2489310000	2540370000	2560620000	2992650000	2619630000	8
Decachlorobiphenyl	4361180000	4315270000	4506230000	4707060000	5500070000	4677960000	10
delta-BHC	7169260000	6905720000	6771510000	6342450000	6619680000	6761720000	5
Dieldrin	5929890000	5841940000	5818180000	5550330000	6016400000	5831350000	3
Endosulfan I	5452270000	5404490000	5449050000	5348660000	6028770000	5536650000	5
Endosulfan II	4873230000	4835950000	4885080000	4789750000	5615800000	4999960000	7
Endosulfan sulfate	4557060000	4504850000	4556170000	4479960000	5092920000	4638190000	6
Endrin	5035940000	4962910000	4938250000	4711180000	5120260000	4953710000	3
Endrin aldehyde	3505880000	3507960000	3571680000	3579710000	4148280000	3662700000	7
Endrin ketone	4819770000	4736510000	4782320000	4695250000	5260400000	4858850000	5
gamma-BHC (Lindane)	7185880000	6905860000	6799650000	6417980000	6684750000	6798820000	4
gamma-Chlordane	5972450000	5894040000	5855640000	5592670000	6341090000	5931180000	5
Heptachlor	5757910000	5597510000	5545540000	5288860000	5757970000	5589560000	3
Heptachlor epoxide	5795570000	5708170000	5718680000	5584980000	6137920000	5789060000	4
Methoxychlor	2050060000	2015340000	2078640000	2094570000	2389310000	2125590000	7
Tetrachloro-m-xylene	3010860000	2951180000	3030100000	3124300000	3454390000	3114160000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: REMI02
SDG NO.: Q3586

Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	33708900000	34309400000	35560900000	36237500000	42373300000	36438000000	10
4,4'-DDE	40500800000	41191000000	42548600000	43577200000	51373700000	43838300000	10
4,4'-DDT	33888000000	33982600000	34868200000	34959100000	37667600000	35073100000	4
Aldrin	43083800000	43574800000	45033800000	46172400000	53729700000	46318900000	9
alpha-BHC	49118000000	48392000000	49683700000	50500300000	57784900000	51095800000	7
alpha-Chlordane	40004000000	40550200000	41893000000	42708900000	50532500000	43137700000	10
beta-BHC	18296700000	18365500000	19080800000	19804100000	23600800000	19829500000	11
Decachlorobiphenyl	28856800000	28247200000	29308000000	30057200000	34987600000	30291400000	9
delta-BHC	45324900000	45220200000	46466400000	47155800000	54416000000	47716700000	8
Dieldrin	40601100000	41425500000	43096200000	44203200000	51593900000	44184000000	10
Endosulfan I	35728600000	36586400000	38185700000	39478600000	46972400000	39390300000	11
Endosulfan II	34269700000	34968100000	36432200000	37520200000	44587200000	37555500000	11
Endosulfan sulfate	32963400000	33538900000	34960800000	35861000000	41900900000	35845000000	10
Endrin	36902000000	37827400000	39400900000	40693000000	47646500000	40494000000	11
Endrin aldehyde	25110900000	25739400000	26893100000	27767400000	32953200000	27692800000	11
Endrin ketone	34460000000	35020500000	36550700000	37816700000	44740900000	37717800000	11
gamma-BHC (Lindane)	44964800000	44604900000	45972800000	46808300000	53893500000	47248900000	8
gamma-Chlordane	41920800000	42422300000	43748100000	44346600000	52306700000	44948900000	9
Heptachlor	41272300000	41810600000	43239300000	44502900000	52189000000	44602800000	10
Heptachlor epoxide	38203400000	38891500000	40672900000	42374400000	53052500000	42639000000	14
Methoxychlor	16116900000	16316400000	17198900000	17735100000	19771700000	17427800000	8
Tetrachloro-m-xylene	27903400000	27689100000	28617400000	29726500000	34949400000	29777100000	10