

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091052.D **Time Analyzed:** 09:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	47.930	50.000	-4.1
4,4'-DDE	6.189	6.090	6.290	45.680	50.000	-8.6
4,4'-DDT	7.015	6.915	7.115	45.970	50.000	-8.1
Aldrin	5.263	5.164	5.364	46.770	50.000	-6.5
alpha-BHC	3.993	3.894	4.094	48.020	50.000	-4.0
alpha-Chlordane	6.020	5.921	6.121	44.830	50.000	-10.3
beta-BHC	4.512	4.412	4.612	46.640	50.000	-6.7
Decachlorobiphenyl	9.065	8.964	9.164	40.770	50.000	-18.5
delta-BHC	4.760	4.660	4.860	47.520	50.000	-5.0
Dieldrin	6.340	6.240	6.440	46.640	50.000	-6.7
Endosulfan I	6.067	5.968	6.168	46.250	50.000	-7.5
Endosulfan II	6.781	6.681	6.881	44.170	50.000	-11.7
Endosulfan sulfate	7.145	7.045	7.245	44.560	50.000	-10.9
Endrin	6.568	6.468	6.668	44.690	50.000	-10.6
Endrin aldehyde	6.910	6.810	7.010	44.860	50.000	-10.3
Endrin ketone	7.625	7.525	7.725	45.420	50.000	-9.2
gamma-BHC (Lindane)	4.324	4.225	4.425	47.370	50.000	-5.3
gamma-Chlordane	5.939	5.840	6.040	45.910	50.000	-8.2
Heptachlor	4.922	4.823	5.023	56.770	50.000	13.5
Heptachlor epoxide	5.684	5.584	5.784	46.640	50.000	-6.7
Methoxychlor	7.488	7.388	7.588	45.160	50.000	-9.7
Tetrachloro-m-xylene	3.544	3.445	3.645	46.290	50.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091052.D **Time Analyzed:** 09:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	52.740	50.000	5.5
4,4'-DDE	5.364	5.266	5.466	51.460	50.000	2.9
4,4'-DDT	6.173	6.074	6.274	51.330	50.000	2.7
Aldrin	4.359	4.261	4.461	51.960	50.000	3.9
alpha-BHC	3.385	3.287	3.487	52.820	50.000	5.6
alpha-Chlordane	5.180	5.081	5.281	51.460	50.000	2.9
beta-BHC	4.018	3.919	4.119	51.530	50.000	3.1
Decachlorobiphenyl	8.060	7.962	8.162	48.670	50.000	-2.7
delta-BHC	4.254	4.155	4.355	52.350	50.000	4.7
Dieldrin	5.502	5.404	5.604	51.890	50.000	3.8
Endosulfan I	5.237	5.138	5.338	51.410	50.000	2.8
Endosulfan II	6.071	5.972	6.172	51.000	50.000	2.0
Endosulfan sulfate	6.472	6.374	6.574	50.850	50.000	1.7
Endrin	5.779	5.681	5.881	50.610	50.000	1.2
Endrin aldehyde	6.249	6.150	6.350	50.690	50.000	1.4
Endrin ketone	6.982	6.883	7.083	51.750	50.000	3.5
gamma-BHC (Lindane)	3.722	3.623	3.823	52.540	50.000	5.1
gamma-Chlordane	5.115	5.017	5.217	51.560	50.000	3.1
Heptachlor	4.073	3.975	4.175	53.950	50.000	7.9
Heptachlor epoxide	4.863	4.764	4.964	50.620	50.000	1.2
Methoxychlor	6.743	6.645	6.845	50.330	50.000	0.7
Tetrachloro-m-xylene	2.874	2.775	2.975	52.220	50.000	4.4