

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091069.D **Time Analyzed:** 15:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	46.770	50.000	-6.5
4,4'-DDE	6.190	6.090	6.290	44.410	50.000	-11.2
4,4'-DDT	7.016	6.915	7.115	44.490	50.000	-11.0
Aldrin	5.263	5.164	5.364	45.710	50.000	-8.6
alpha-BHC	3.993	3.894	4.094	47.310	50.000	-5.4
alpha-Chlordane	6.020	5.921	6.121	42.890	50.000	-14.2
beta-BHC	4.511	4.412	4.612	45.720	50.000	-8.6
Decachlorobiphenyl	9.066	8.964	9.164	42.180	50.000	-15.6
delta-BHC	4.759	4.660	4.860	46.470	50.000	-7.1
Dieldrin	6.341	6.240	6.440	44.580	50.000	-10.8
Endosulfan I	6.068	5.968	6.168	44.380	50.000	-11.2
Endosulfan II	6.781	6.681	6.881	42.760	50.000	-14.5
Endosulfan sulfate	7.145	7.045	7.245	43.560	50.000	-12.9
Endrin	6.568	6.468	6.668	43.400	50.000	-13.2
Endrin aldehyde	6.910	6.810	7.010	43.990	50.000	-12.0
Endrin ketone	7.626	7.525	7.725	44.810	50.000	-10.4
gamma-BHC (Lindane)	4.324	4.225	4.425	46.490	50.000	-7.0
gamma-Chlordane	5.939	5.840	6.040	44.130	50.000	-11.7
Heptachlor	4.921	4.823	5.023	55.010	50.000	10.0
Heptachlor epoxide	5.682	5.584	5.784	45.090	50.000	-9.8
Methoxychlor	7.489	7.388	7.588	44.120	50.000	-11.8
Tetrachloro-m-xylene	3.543	3.445	3.645	46.120	50.000	-7.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091069.D **Time Analyzed:** 15:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	49.330	50.000	-1.3
4,4'-DDE	5.364	5.266	5.466	47.770	50.000	-4.5
4,4'-DDT	6.173	6.074	6.274	47.280	50.000	-5.4
Aldrin	4.358	4.261	4.461	48.890	50.000	-2.2
alpha-BHC	3.385	3.287	3.487	50.250	50.000	0.5
alpha-Chlordane	5.180	5.081	5.281	47.750	50.000	-4.5
beta-BHC	4.018	3.919	4.119	49.010	50.000	-2.0
Decachlorobiphenyl	8.060	7.962	8.162	45.830	50.000	-8.3
delta-BHC	4.254	4.155	4.355	49.490	50.000	-1.0
Dieldrin	5.502	5.404	5.604	48.120	50.000	-3.8
Endosulfan I	5.237	5.138	5.338	47.550	50.000	-4.9
Endosulfan II	6.070	5.972	6.172	47.490	50.000	-5.0
Endosulfan sulfate	6.472	6.374	6.574	47.450	50.000	-5.1
Endrin	5.779	5.681	5.881	47.840	50.000	-4.3
Endrin aldehyde	6.249	6.150	6.350	47.200	50.000	-5.6
Endrin ketone	6.981	6.883	7.083	47.690	50.000	-4.6
gamma-BHC (Lindane)	3.721	3.623	3.823	49.700	50.000	-0.6
gamma-Chlordane	5.115	5.017	5.217	47.830	50.000	-4.3
Heptachlor	4.073	3.975	4.175	51.050	50.000	2.1
Heptachlor epoxide	4.862	4.764	4.964	47.460	50.000	-5.1
Methoxychlor	6.743	6.645	6.845	47.560	50.000	-4.9
Tetrachloro-m-xylene	2.873	2.775	2.975	49.860	50.000	-0.3