

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091083.D **Time Analyzed:** 09:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	50.700	50.000	1.4
4,4'-DDE	6.190	6.090	6.290	47.140	50.000	-5.7
4,4'-DDT	7.016	6.915	7.115	46.930	50.000	-6.1
Aldrin	5.264	5.164	5.364	47.990	50.000	-4.0
alpha-BHC	3.993	3.894	4.094	49.100	50.000	-1.8
alpha-Chlordane	6.020	5.921	6.121	46.460	50.000	-7.1
beta-BHC	4.512	4.412	4.612	47.120	50.000	-5.8
Decachlorobiphenyl	9.066	8.964	9.164	43.920	50.000	-12.2
delta-BHC	4.760	4.660	4.860	48.430	50.000	-3.1
Dieldrin	6.341	6.240	6.440	48.050	50.000	-3.9
Endosulfan I	6.068	5.968	6.168	47.580	50.000	-4.8
Endosulfan II	6.781	6.681	6.881	45.840	50.000	-8.3
Endosulfan sulfate	7.145	7.045	7.245	46.880	50.000	-6.2
Endrin	6.568	6.468	6.668	45.730	50.000	-8.5
Endrin aldehyde	6.910	6.810	7.010	47.220	50.000	-5.6
Endrin ketone	7.625	7.525	7.725	48.450	50.000	-3.1
gamma-BHC (Lindane)	4.324	4.225	4.425	48.390	50.000	-3.2
gamma-Chlordane	5.939	5.840	6.040	47.260	50.000	-5.5
Heptachlor	4.923	4.823	5.023	57.890	50.000	15.8
Heptachlor epoxide	5.684	5.584	5.784	48.040	50.000	-3.9
Methoxychlor	7.489	7.388	7.588	46.930	50.000	-6.1
Tetrachloro-m-xylene	3.544	3.445	3.645	47.460	50.000	-5.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3586
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091083.D **Time Analyzed:** 09:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	57.160	50.000	14.3
4,4'-DDE	5.364	5.266	5.466	53.980	50.000	8.0
4,4'-DDT	6.172	6.074	6.274	53.820	50.000	7.6
Aldrin	4.359	4.261	4.461	53.990	50.000	8.0
alpha-BHC	3.385	3.287	3.487	54.770	50.000	9.5
alpha-Chlordane	5.180	5.081	5.281	53.940	50.000	7.9
beta-BHC	4.018	3.919	4.119	53.360	50.000	6.7
Decachlorobiphenyl	8.060	7.962	8.162	55.600	50.000	11.2
delta-BHC	4.254	4.155	4.355	54.010	50.000	8.0
Dieldrin	5.502	5.404	5.604	54.550	50.000	9.1
Endosulfan I	5.236	5.138	5.338	53.860	50.000	7.7
Endosulfan II	6.070	5.972	6.172	54.820	50.000	9.6
Endosulfan sulfate	6.472	6.374	6.574	54.680	50.000	9.4
Endrin	5.779	5.681	5.881	52.800	50.000	5.6
Endrin aldehyde	6.249	6.150	6.350	54.950	50.000	9.9
Endrin ketone	6.981	6.883	7.083	57.120	50.000	14.2
gamma-BHC (Lindane)	3.721	3.623	3.823	54.230	50.000	8.5
gamma-Chlordane	5.115	5.017	5.217	54.010	50.000	8.0
Heptachlor	4.073	3.975	4.175	55.610	50.000	11.2
Heptachlor epoxide	4.862	4.764	4.964	52.690	50.000	5.4
Methoxychlor	6.743	6.645	6.845	53.940	50.000	7.9
Tetrachloro-m-xylene	2.874	2.775	2.975	53.990	50.000	8.0