

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3586</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Remington & Vernick Engineers</u>	Datafile :	<u>PD091085.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170476BS (Column 1)	alpha-BHC	16.66	15.4	ug/kg	92				40 (84)	140 (123)	
	beta-BHC	16.66	14.9	ug/kg	89				40 (82)	140 (123)	
	delta-BHC	16.66	15.6	ug/kg	94				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.66	15.2	ug/kg	91				40 (83)	140 (125)	
	Heptachlor	16.66	18.6	ug/kg	112				40 (83)	140 (122)	
	Aldrin	16.66	15.2	ug/kg	91				40 (82)	140 (124)	
	Heptachlor epoxide	16.66	16.6	ug/kg	100				40 (83)	140 (120)	
	Endosulfan I	16.66	15.3	ug/kg	92				40 (81)	140 (124)	
	Dieldrin	16.66	15.4	ug/kg	92				40 (85)	140 (121)	
	4,4'-DDE	16.66	14.6	ug/kg	88				40 (81)	140 (123)	
	Endrin	16.66	14.8	ug/kg	89				40 (76)	140 (130)	
	Endosulfan II	16.66	17.0	ug/kg	102				40 (80)	140 (125)	
	4,4'-DDD	16.66	16.3	ug/kg	98				40 (80)	140 (131)	
	Endosulfan sulfate	16.66	15.2	ug/kg	91				40 (81)	140 (122)	
	4,4'-DDT	16.66	16.2	ug/kg	97				40 (70)	140 (129)	
	Methoxychlor	16.66	18.3	ug/kg	110				40 (60)	140 (119)	
	Endrin ketone	16.66	15.8	ug/kg	95				40 (77)	140 (132)	
	Endrin aldehyde	16.66	16.4	ug/kg	98				40 (79)	140 (124)	
	alpha-Chlordane	16.66	15.1	ug/kg	91				40 (84)	140 (120)	
	gamma-Chlordane	16.66	15.1	ug/kg	91				40 (83)	140 (122)	
PB170476BS (Column 2)	alpha-BHC	16.66	17.4	ug/kg	104				40 (84)	140 (123)	
	beta-BHC	16.66	16.6	ug/kg	100				40 (82)	140 (123)	
	delta-BHC	16.66	17.2	ug/kg	103				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.66	17.2	ug/kg	103				40 (83)	140 (125)	
	Heptachlor	16.66	17.9	ug/kg	107				40 (83)	140 (122)	
	Aldrin	16.66	17.4	ug/kg	104				40 (82)	140 (124)	
	Heptachlor epoxide	16.66	16.7	ug/kg	100				40 (83)	140 (120)	
	Endosulfan I	16.66	13.5	ug/kg	81				40 (81)	140 (124)	
	Dieldrin	16.66	17.4	ug/kg	104				40 (85)	140 (121)	
	4,4'-DDE	16.66	17.2	ug/kg	103				40 (81)	140 (123)	
	Endrin	16.66	19.9	ug/kg	119				40 (76)	140 (130)	
	Endosulfan II	16.66	17.5	ug/kg	105				40 (80)	140 (125)	
	4,4'-DDD	16.66	18.1	ug/kg	109				40 (80)	140 (131)	
	Endosulfan sulfate	16.66	17.4	ug/kg	104				40 (81)	140 (122)	
	4,4'-DDT	16.66	17.6	ug/kg	106				40 (70)	140 (129)	
	Methoxychlor	16.66	17.6	ug/kg	106				40 (60)	140 (119)	
	Endrin ketone	16.66	21.5	ug/kg	129				40 (77)	140 (132)	
	Endrin aldehyde	16.66	17.8	ug/kg	107				40 (79)	140 (124)	
	alpha-Chlordane	16.66	17.1	ug/kg	103				40 (84)	140 (120)	
	gamma-Chlordane	16.66	17.2	ug/kg	103				40 (83)	140 (122)	