

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG No.: Q3586

Instrument ID: BNA_F

Calibration Date(s): 11/05/2025 11/05/2025

Calibration Time(s): 12:50 16:49

LAB FILE ID:		RRF2.5 = BF144169.D			RRF005 = BF144170.D		RRF010 = BF144171.D	
		RRF020 = BF144172.D			RRF040 = BF144173.D		RRF050 = BF144174.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.332	1.303	1.316	1.315	1.157	1.278	5.6
Benzaldehyde			1.033	0.917	0.816	0.672	0.818	17.0
Phenol-d6		1.566	1.518	1.511	1.460	1.281	1.448	7.1
Phenol		1.845	1.896	1.840	1.858	1.562	1.769	7.6
bis(2-Chloroethyl)ether		1.278	1.333	1.342	1.316	1.155	1.289	5.2
2-Chlorophenol		1.242	1.309	1.292	1.281	1.152	1.253	4.6
2-Methylphenol		1.031	1.054	1.007	1.033	0.895	0.997	5.4
2,2-oxybis(1-Chloropropane)		2.408	2.470	2.496	2.418	2.147	2.374	5.5
Acetophenone		0.479	0.464	0.434	0.434	0.380	0.429	8.3
3+4-Methylphenols			1.330	1.223	1.215	1.030	1.175	9.2
n-Nitroso-di-n-propylamine	0.998	1.010	1.023	0.996	0.950	0.813	0.953	7.4
Nitrobenzene-d5		0.343	0.352	0.349	0.360	0.321	0.346	4.1
Hexachloroethane		0.512	0.513	0.532	0.532	0.467	0.515	4.9
Nitrobenzene		0.370	0.390	0.376	0.388	0.342	0.376	4.9
Isophorone		0.659	0.668	0.669	0.673	0.581	0.655	5.1
2-Nitrophenol		0.167	0.187	0.191	0.197	0.174	0.186	6.5
2,4-Dimethylphenol		0.281	0.278	0.265	0.298	0.260	0.279	5.1
bis(2-Chloroethoxy)methane		0.426	0.417	0.412	0.423	0.372	0.409	5.0
2,4-Dichlorophenol		0.334	0.342	0.330	0.331	0.289	0.322	6.2
Naphthalene		1.019	1.008	0.982	0.969	0.850	0.951	6.7
4-Chloroaniline		0.358	0.364	0.355	0.362	0.309	0.347	6.1
Hexachlorobutadiene		0.256	0.250	0.252	0.259	0.232	0.249	4.0
Caprolactam			0.094	0.093	0.092	0.075	0.088	8.1
4-Chloro-3-methylphenol		0.283	0.307	0.304	0.297	0.259	0.288	5.8
2-Methylnaphthalene		0.702	0.672	0.665	0.647	0.559	0.636	8.0
Hexachlorocyclopentadiene			0.121	0.142	0.200	0.204	0.195	27.8
2,4,6-Trichlorophenol		0.440	0.449	0.449	0.456	0.409	0.446	5.5
2-Fluorobiphenyl		1.598	1.428	1.361	1.350	1.177	1.354	10.7
2,4,5-Trichlorophenol		0.497	0.489	0.497	0.489	0.438	0.481	5.1
1,1-Biphenyl		1.535	1.446	1.425	1.408	1.239	1.396	7.5
2-Chloronaphthalene		1.303	1.216	1.206	1.204	1.068	1.191	7.1
2-Nitroaniline		0.302	0.347	0.349	0.361	0.326	0.344	6.8
Dimethylphthalate		1.437	1.343	1.359	1.345	1.168	1.325	6.7
Acenaphthylene		1.727	1.686	1.677	1.645	1.440	1.623	6.6

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG No.: Q3586

Instrument ID: BNA_F

Calibration Date(s): 11/05/2025 11/05/2025

Calibration Time(s): 12:50 16:49

LAB FILE ID:		RRF2.5 = BF144169.D			RRF005 = BF144170.D		RRF010 = BF144171.D	
		RRF020 = BF144172.D			RRF040 = BF144173.D		RRF050 = BF144174.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.251	0.271	0.271	0.276	0.252	0.268	5.2
3-Nitroaniline		0.271	0.301	0.302	0.315	0.264	0.293	7.1
Acenaphthene		1.138	1.162	1.120	1.096	0.956	1.085	6.9
2,4-Dinitrophenol			0.202	0.166	0.140	0.133	0.162	15.0
4-Nitrophenol			0.273	0.202	0.204	0.185	0.212	14.6
Dibenzofuran		1.644	1.600	1.586	1.551	1.337	1.520	8.0
2,4-Dinitrotoluene		0.322	0.377	0.372	0.385	0.328	0.359	7.1
Diethylphthalate		1.294	1.273	1.254	1.254	1.075	1.221	6.3
4-Chlorophenyl-phenylether		0.711	0.701	0.678	0.685	0.570	0.658	7.9
Fluorene		1.295	1.223	1.197	1.180	0.997	1.156	8.9
4-Nitroaniline		0.273	0.290	0.293	0.288	0.254	0.279	5.1
4,6-Dinitro-2-methylphenol			0.147	0.138	0.130	0.120	0.136	7.0
n-Nitrosodiphenylamine		0.695	0.676	0.635	0.646	0.577	0.643	6.1
2,4,6-Tribromophenol		0.242	0.257	0.257	0.265	0.227	0.251	5.3
4-Bromophenyl-phenylether		0.264	0.260	0.255	0.263	0.228	0.255	5.2
Hexachlorobenzene		0.306	0.306	0.305	0.300	0.273	0.301	4.7
Atrazine		0.215	0.211	0.215	0.206	0.180	0.205	6.3
Pentachlorophenol			0.135	0.144	0.156	0.137	0.150	8.8
Phenanthrene		1.103	1.042	0.986	1.009	0.892	0.996	6.9
Anthracene		1.120	1.059	1.027	1.022	0.897	1.013	7.4
Carbazole		0.938	0.917	0.882	0.878	0.762	0.861	7.4
Di-n-butylphthalate		1.075	1.101	1.077	1.070	0.917	1.041	6.3
Fluoranthene		1.117	1.066	1.053	1.054	0.910	1.029	6.8
Pyrene		1.763	1.745	1.631	1.704	1.372	1.613	8.8
Terphenyl-d14		1.403	1.415	1.274	1.317	1.042	1.259	10.7
Butylbenzylphthalate		0.565	0.590	0.550	0.590	0.499	0.559	5.6
3,3-Dichlorobenzidine			0.453	0.461	0.482	0.441	0.475	6.8
Benzo (a) anthracene		1.455	1.340	1.330	1.430	1.290	1.386	5.4
Chrysene		1.276	1.245	1.279	1.403	1.164	1.280	5.7
Bis (2-ethylhexyl) phthalate		0.773	0.739	0.755	0.816	0.705	0.765	4.9
Di-n-octyl phthalate			1.163	1.272	1.410	1.269	1.320	7.9
Benzo (b) fluoranthene		1.199	1.133	1.114	1.233	1.002	1.137	6.4
Benzo (k) fluoranthene		1.077	1.094	1.083	1.041	0.999	1.064	4.4
Benzo (a) pyrene		1.064	1.050	1.058	1.077	0.963	1.049	4.2
Indeno (1,2,3-cd) pyrene		1.450	1.411	1.447	1.485	1.311	1.436	4.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG No.: Q3586

Instrument ID: BNA_F

Calibration Date(s): 11/05/2025 11/05/2025

Calibration Time(s): 12:50 16:49

LAB FILE ID:		RRF2.5 = BF144169.D		RRF005 = BF144170.D		RRF010 = BF144171.D		
		RRF020 = BF144172.D		RRF040 = BF144173.D		RRF050 = BF144174.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		1.155	1.155	1.175	1.198	1.040	1.153	4.8
Benzo(g,h,i)perylene		1.095	1.125	1.153	1.191	1.028	1.139	5.7
1,2,4,5-Tetrachlorobenzene		0.677	0.662	0.659	0.675	0.588	0.655	6.6
1,4-Dioxane		0.526	0.487	0.518	0.554	0.500	0.528	5.7
2,3,4,6-Tetrachlorophenol		0.308	0.347	0.336	0.358	0.316	0.340	6.6