

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111125\
 Data File : BF144199.D
 Acq On : 11 Nov 2025 13:07
 Operator : RC/JU
 Sample : PB170478BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB170478BL

Quant Time: Nov 11 13:46:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 18:37:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	80148	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	308933	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	180668	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	336603	20.000	ng	0.00
76) Chrysene-d12	14.057	240	225293	20.000	ng	0.00
86) Perylene-d12	15.545	264	226985	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	588374	114.879	ng	0.01
7) Phenol-d6	6.504	99	684658	117.978	ng	0.00
23) Nitrobenzene-d5	7.434	82	443974	83.001	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	284869	125.649	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	975916	79.780	ng	0.00
79) Terphenyl-d14	12.998	244	1198876	84.526	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111125\
Data File : BF144199.D
Acq On : 11 Nov 2025 13:07
Operator : RC/JU
Sample : PB170478BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB170478BL


Quant Time: Nov 11 13:46:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 05 18:37:33 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 796
Delta R.T. -0.000 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07


Instrument :
BNA_F
ClientSampleId :
PB170478BL





Tgt Ion:152 Resp: 80148
Ion Ratio Lower Upper
152 100
150 156.6 127.4 191.0
115 61.5 49.5 74.3




#5
2-Fluorophenol
Concen: 114.879 ng
RT: 5.504 min Scan# 563
Delta R.T. 0.012 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07



Tgt Ion:112 Resp: 588374
Ion Ratio Lower Upper
112 100
64 61.9 47.2 70.8
63 32.3 24.4 36.6









#7
Phenol-d6
Concen: 117.978 ng
RT: 6.504 min Scan# 733
Delta R.T. 0.006 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07




Instrument :
BNA_F
ClientSampleId :
PB170478BL


Tgt Ion: 99 Resp: 684658
Ion Ratio Lower Upper
99 100
42 27.5 20.9 31.3
71 39.9 31.4 47.2



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07

Tgt Ion: 136 Resp: 308933
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 8.6 7.0 10.6
68 5.6 4.3 6.5





#23

Nitrobenzene-d5

Concen: 83.001 ng

RT: 7.434 min Scan# 891

Delta R.T. 0.000 min

Lab File: BF144199.D

Acq: 11 Nov 2025 13:07

Instrument :

BNA_F

ClientSampleId :

PB170478BL





Tgt Ion: 82 Resp: 443974

Ion Ratio Lower Upper

82 100

128 41.6 34.7 52.1

54 65.2 49.5 74.3



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.916 min Scan# 1313

Delta R.T. -0.000 min

Lab File: BF144199.D

Acq: 11 Nov 2025 13:07




Tgt Ion:164 Resp: 180668


Ion Ratio Lower Upper

164 100

162 101.5 81.1 121.7


160 44.6 34.5 51.7








#42
2,4,6-Tribromophenol
Concen: 125.649 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.006 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07

Instrument :
BNA_F
ClientSampleId :
PB170478BL







Tgt Ion: 330 Resp: 284869
Ion Ratio Lower Upper
330 100
332 97.7 76.7 115.1
141 22.2 17.8 26.8



#45
2-Fluorobiphenyl
Concen: 79.780 ng
RT: 9.234 min Scan# 1197
Delta R.T. 0.006 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07

Tgt Ion: 172 Resp: 975916
Ion Ratio Lower Upper
172 100
171 34.0 28.1 42.1
170 23.1 18.6 28.0









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07




Instrument :
BNA_F
ClientSampleId :
PB170478BL


Tgt Ion:188 Resp: 336603
Ion Ratio Lower Upper
188 100
94 7.4 6.2 9.2
80 7.6 6.9 10.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BF144199.D
Acq: 11 Nov 2025 13:07


Tgt Ion:240 Resp: 225293
Ion Ratio Lower Upper
240 100
120 7.7 6.6 10.0
236 25.2 21.4 32.2








#79
 Terphenyl-d14
 Concen: 84.526 ng
 RT: 12.998 min Scan# 1836
 Delta R.T. 0.006 min
 Lab File: BF144199.D
 Acq: 11 Nov 2025 13:07

Instrument :
 BNA_F
 ClientSampleId :
 PB170478BL



Tgt Ion:244 Resp: 1198876
 Ion Ratio Lower Upper
 244 100
 212 6.4 5.3 7.9
 122 6.8 5.6 8.4



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.545 min Scan# 2270
 Delta R.T. 0.006 min
 Lab File: BF144199.D
 Acq: 11 Nov 2025 13:07

Tgt Ion:264 Resp: 226985
 Ion Ratio Lower Upper
 264 100
 260 21.5 17.1 25.7
 265 21.4 17.4 26.0

