

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3586

Client ID : SSTDCCC040

Date Analyzed: 11/11/2025

Lab File ID: BP026090.D

Time Analyzed: 14:44

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	341127	7.678	1393560	10.45	901803	14.32
UPPER LIMIT	682254	8.178	2787120	10.949	1803610	14.819
LOWER LIMIT	170564	7.178	696780	9.949	450902	13.819
EPA SAMPLE NO.						
01 SED-2	249884	7.68	984033	10.44	622887	14.31
02 SED-3	318405	7.68	1214570	10.46	729423	14.32
03 SED-3MS	296222	7.68	1166340	10.45	745428	14.31
04 SED-3MSD	270854	7.67	1115780	10.44	711326	14.32

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.