

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3586</u>
Lab File ID:	<u>VW032415.D</u>	Date Analyzed:	<u>11/11/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>08:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	445777	7.96	817849	8.85	708146	11.63
UPPER LIMIT	891554	8.459	1635700	9.349	1416290	12.129
LOWER LIMIT	222889	7.459	408925	8.349	354073	11.129
EPA SAMPLE NO.						
SED-1	286762	7.97	660595	8.86	665967	11.64
SED-2	283259	7.96	670411	8.86	673114	11.64
VW1111SBL01	277247	7.96	601287	8.85	633123	11.63
VW1111SBS01	436360	7.96	782865	8.85	700699	11.63
VW1111SBSD01	419398	7.97	800811	8.85	702782	11.64

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.