

### Duplicate Sample Summary

|                               |                  |                  |
|-------------------------------|------------------|------------------|
| <b>Lab Sample Id :</b>        | Q3594-01DUP      | Q3594-01         |
| <b>Client Id :</b>            | IA1ADUP          | IA1A             |
| <b>DF :</b>                   | 1                | 1                |
| <b>Datafile :</b>             | VL043184.D       | VL043183.D       |
| <b>Anal Date &amp; Time :</b> | 11/11/2025 11:09 | 11/11/2025 10:35 |

| Parameter              | Result | Result | RPD    |
|------------------------|--------|--------|--------|
| 1,1,1-Trichloroethane  | 0      | 0      | 0      |
| 1,1-Dichloroethane     | 0      | 0      | 0      |
| 1,2,4-Trimethylbenzene | 0.71   | 0.46   | 42.7 * |
| 1,3,5-Trimethylbenzene | 0.2    | 0      | 200 *  |
| 2,2,4-Trimethylpentane | 1.4    | 0.96   | 37.3 * |
| Benzene                | 1.5    | 1.1    | 30.8 * |
| cis-1,2-Dichloroethene | 0      | 0      | 0      |
| Cyclohexane            | 0.55   | 0.41   | 29.2 * |
| Ethyl Benzene          | 0.72   | 0.49   | 38 *   |
| Heptane                | 1.1    | 0.79   | 32.8 * |
| Hexane                 | 19.6   | 18.2   | 7.4    |
| m/p-Xylene             | 2.4    | 1.6    | 40 *   |
| Naphthalene            | 0      | 0      | 0      |
| o-Xylene               | 0.88   | 0.56   | 44.4 * |
| Tetrachloroethene      | 0.04   | 0.05   | 22.2 * |
| Toluene                | 3.8    | 2.6    | 37.5 * |
| Trichloroethene        | 0      | 0      | 0      |
| Vinyl Chloride         | 0      | 0      | 0      |