

Duplicate Sample Summary

Client:	Homeowner	SDG No.:	Q3603
Project:	Short Hills - Drinking Water	Sample ID:	Q3603-01
Client ID:	SHORT-HILLS-DWDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		11/11/2025

Duplicate Sample Summary

Client:	Homeowner	SDG No.:	Q3603
Project:	Short Hills - Drinking Water	Sample ID:	Q3603-01
Client ID:	SHORT-HILLS-DWMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.98		0.97		2	1.23		11/11/2025
Nitrate	mg/L	+/-20	2.50		2.50		1	0		11/11/2025
Chloride	mg/L	+/-20	77.1	OR	77.1	OR	1	0		11/11/2025
Sulfate	mg/L	+/-20	36.0		35.5		1	1		11/11/2025
Orthophosphate as P	mg/L	+/-20	5.20		5.10		1	2		11/11/2025
Bromide	mg/L	+/-20	10.3		10.0		1	3		11/11/2025
Nitrite	mg/L	+/-20	3.10		3.00		1	3		11/11/2025
Fluoride	mg/L	+/-20	2.20		2.10		1	5		11/11/2025