



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB124	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	12:32	LB138004

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB22	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	12:36	LB138004
CCB23	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	13:04	LB138004
CCB24	Mercury	0.027	+/-0.2	U	0.20	CV	11/21/2025	13:16	LB138004

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	28.4	+/-50	J	100	P	11/12/2025	17:35	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	17:35	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	17:35	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	17:35	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	17:35	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	17:35	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	17:35	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	17:35	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	17:35	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	17:35	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	17:35	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	17:35	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	17:35	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	17:35	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	17:35	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	17:35	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	17:35	LB137883
	Silver	1.62	+/-5	U	10.0	P	11/12/2025	17:35	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	17:35	LB137883
	Thallium	4.38	+/-20	U	40.0	P	11/12/2025	17:35	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	17:35	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	17:35	LB137883

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	22.3	+/-50	J	100	P	11/12/2025	18:13	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	18:13	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	18:13	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	18:13	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	18:13	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	18:13	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	18:13	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	18:13	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	18:13	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	18:13	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	18:13	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	18:13	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	18:13	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	18:13	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	18:13	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	18:13	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	18:13	LB137883
	Silver	1.68	+/-5	J	10.0	P	11/12/2025	18:13	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	18:13	LB137883
	Thallium	8.50	+/-20	J	40.0	P	11/12/2025	18:13	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	18:13	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	18:13	LB137883
CCB02	Aluminum	11.3	+/-50	U	100	P	11/12/2025	19:24	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	19:24	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	19:24	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	19:24	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	19:24	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	19:24	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	19:24	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	19:24	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	19:24	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	19:24	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	19:24	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	19:24	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	19:24	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	19:24	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	19:24	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	19:24	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	19:24	LB137883

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/12/2025	19:24	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	19:24	LB137883
	Thallium	4.38	+/-20	U	40.0	P	11/12/2025	19:24	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	19:24	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	19:24	LB137883
CCB03	Aluminum	19.7	+/-50	J	100	P	11/12/2025	20:37	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	20:37	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	20:37	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	20:37	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	20:37	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	20:37	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	20:37	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	20:37	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	20:37	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	20:37	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	20:37	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	20:37	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	20:37	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	20:37	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	20:37	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	20:37	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	20:37	LB137883
	Silver	1.62	+/-5	U	10.0	P	11/12/2025	20:37	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	20:37	LB137883
	Thallium	7.39	+/-20	J	40.0	P	11/12/2025	20:37	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	20:37	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	20:37	LB137883
CCB04	Aluminum	11.3	+/-50	U	100	P	11/12/2025	22:03	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	22:03	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	22:03	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	22:03	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	22:03	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	22:03	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	22:03	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	22:03	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	22:03	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	22:03	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	22:03	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	22:03	LB137883

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/12/2025	22:03	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	22:03	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	22:03	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	22:03	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	22:03	LB137883
	Silver	1.62	+/-5	U	10.0	P	11/12/2025	22:03	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	22:03	LB137883
	Thallium	8.31	+/-20	J	40.0	P	11/12/2025	22:03	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	22:03	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	22:03	LB137883
CCB05	Aluminum	23.3	+/-50	J	100	P	11/12/2025	22:20	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	22:20	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	22:20	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	22:20	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	22:20	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	22:20	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	22:20	LB137883
	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	22:20	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	22:20	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	22:20	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	22:20	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	22:20	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	22:20	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	22:20	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	22:20	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	22:20	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	22:20	LB137883
	Silver	1.62	+/-5	U	10.0	P	11/12/2025	22:20	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	22:20	LB137883
	Thallium	4.38	+/-20	U	40.0	P	11/12/2025	22:20	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	22:20	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	22:20	LB137883
CCB06	Aluminum	40.9	+/-50	J	100	P	11/12/2025	22:57	LB137883
	Antimony	6.76	+/-25	U	50.0	P	11/12/2025	22:57	LB137883
	Arsenic	5.12	+/-10	U	20.0	P	11/12/2025	22:57	LB137883
	Barium	14.6	+/-50	U	100	P	11/12/2025	22:57	LB137883
	Beryllium	0.56	+/-3	U	6.00	P	11/12/2025	22:57	LB137883
	Cadmium	0.50	+/-3	U	6.00	P	11/12/2025	22:57	LB137883
	Calcium	234	+/-1000	U	2000	P	11/12/2025	22:57	LB137883

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/12/2025	22:57	LB137883
	Cobalt	2.26	+/-15	U	30.0	P	11/12/2025	22:57	LB137883
	Copper	4.60	+/-10	U	20.0	P	11/12/2025	22:57	LB137883
	Iron	23.4	+/-50	U	100	P	11/12/2025	22:57	LB137883
	Lead	2.30	+/-6	U	12.0	P	11/12/2025	22:57	LB137883
	Magnesium	244	+/-1000	U	2000	P	11/12/2025	22:57	LB137883
	Manganese	5.94	+/-10	U	20.0	P	11/12/2025	22:57	LB137883
	Nickel	3.06	+/-20	U	40.0	P	11/12/2025	22:57	LB137883
	Potassium	918	+/-1000	U	2000	P	11/12/2025	22:57	LB137883
	Selenium	9.64	+/-10	U	20.0	P	11/12/2025	22:57	LB137883
	Silver	1.62	+/-5	U	10.0	P	11/12/2025	22:57	LB137883
	Sodium	868	+/-1000	U	2000	P	11/12/2025	22:57	LB137883
	Thallium	7.26	+/-20	J	40.0	P	11/12/2025	22:57	LB137883
	Vanadium	6.26	+/-20	U	40.0	P	11/12/2025	22:57	LB137883
	Zinc	16.7	+/-20	U	40.0	P	11/12/2025	22:57	LB137883

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------