

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170508BS							
Aluminum	ug/L	1000	982		98	85 - 115	P
Antimony	ug/L	400	383		96	85 - 115	P
Arsenic	ug/L	400	391		98	85 - 115	P
Barium	ug/L	100	93.6		94	85 - 115	P
Beryllium	ug/L	100	99.6		100	85 - 115	P
Cadmium	ug/L	100	91.6		92	85 - 115	P
Calcium	ug/L	500	501	J	100	85 - 115	P
Chromium	ug/L	200	198		99	85 - 115	P
Cobalt	ug/L	100	94.1		94	85 - 115	P
Copper	ug/L	150	149		99	85 - 115	P
Iron	ug/L	1500	1460		97	85 - 115	P
Lead	ug/L	500	446		89	85 - 115	P
Magnesium	ug/L	1000	977	J	98	85 - 115	P
Manganese	ug/L	100	93.2		93	85 - 115	P
Nickel	ug/L	250	234		94	85 - 115	P
Potassium	ug/L	5000	4910		98	85 - 115	P
Selenium	ug/L	1000	940		94	85 - 115	P
Silver	ug/L	37.5	36.1		96	85 - 115	P
Sodium	ug/L	1500	1620		108	85 - 115	P
Thallium	ug/L	1000	1010		101	85 - 115	P
Vanadium	ug/L	150	141		94	85 - 115	P
Zinc	ug/L	100	101		101	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Homeowner

SDG No.: Q3603

Contract: HOME01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170687BS Mercury	ug/L	4.0	4.41		110	85 - 115	CV