

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>HOME01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3603</u>
Lab File ID:	<u>VX048563.D</u>	Date Analyzed:	<u>11/12/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:11</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	123894	12.00€				
UPPER LIMIT	247788	12.50€				
LOWER LIMIT	61947	11.50€				
EPA SAMPLE NO.						
SHORT-HILLS-DW	211749	12.01				
TB	181468	12.01				
VX1112WBL01	232670	12.01				
VX1112WBS01	136753	12.01				
VX1112WBSD01	125743	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.