

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137876

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/12/2025	14:00
2	CAL2	1	Water	NA	NA	20.2	7.01	11/12/2025	14:02
3	CAL3	1	Water	NA	NA	20.3	10.02	11/12/2025	14:05
4	ICV	1	Water	NA	NA	20.3	7.01	11/12/2025	14:10
5	CCV1	1	Water	NA	NA	20.2	2.01	11/12/2025	14:15
6	Q3604-01	1	Solid	20.02	20	20.6	7.70	11/13/2025	14:25
7	Q3604-01DUP	1	Solid	20.03	20	20.8	7.72	11/12/2025	14:30
8	Q3604-02	1	Solid	20.02	20	20.1	8.66	11/12/2025	14:37
9	Q3609-03	1	Solid	20.03	20	21.0	7.53	11/12/2025	14:45
10	Q3609-06	1	Solid	20.02	20	21.9	6.41	11/12/2025	14:50
11	Q3609-09	1	Solid	20.03	20	21.7	6.35	11/12/2025	15:00
12	Q3609-12	1	Solid	20.04	20	21.2	5.83	11/12/2025	15:05
13	Q3609-15	1	Solid	20.02	20	21.2	5.75	11/12/2025	15:15
14	CCV2	1	Water	NA	NA	20.3	12.02	11/12/2025	15:25

WORKLIST(Hardcopy Internal Chain)

137876

WorkList Name : corrsivity q3609 WorkList ID : 193076 Department : Wet-Chemistry Date : 11-12-2025 13:03:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3604-01	S-1	Solid	Corrosivity	Cool 4 deg C	ROMA02	D41	11/10/2025	9045D
Q3604-02	S-2	Solid	Corrosivity	Cool 4 deg C	ROMA02	D41	11/10/2025	9045D
Q3609-03	AU-713-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/11/2025	9045D
Q3609-06	AU-713-COMP-02	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/11/2025	9045D
Q3609-09	AU-713-COMP-03	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/11/2025	9045D
Q3609-12	AU-713-COMP-04	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/11/2025	9045D
Q3609-15	AU-713-COMP-05	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/11/2025	9045D

Date/Time 11-12-25 13:30
 Raw Sample Received by: Se Wei
 Raw Sample Relinquished by: OP

Date/Time 11-12-25 17:40
 Raw Sample Received by: OP
 Raw Sample Relinquished by: Se Wei