

Initial and Continuing Calibration Verification

Client: Roman E&G Corp

SDG No.: Q3604

Project: MCUA - New Brunswick

RunNo.: LB137874

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	11/12/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/12/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/12/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/12/2025

Initial and Continuing Calibration Verification

Client: Roman E&G Corp

SDG No.: Q3604

Project: MCUA - New Brunswick

RunNo.: LB137876

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Corrosivity	pH	7.01	7	100	90-110	11/12/2025
Sample ID: CCV1 Corrosivity	pH	2.01	2.00	101	90-110	11/12/2025
Sample ID: CCV2 Corrosivity	pH	12.02	12.00	100	90-110	11/12/2025