

Initial Calibration Report for SequenceID : FE103025AL

AreaCount

Parameter Range	FE056575.D	FE056576.D	FE056577.D	FE056578.D	FE056579.D	
Aliphatic C9-C12	47378340.000	25372826.000	10134874.000	5323968.000	2735296.000	
Aliphatic C12-C16	32958436.000	17867364.000	7039199.000	3707774.000	1889371.000	
Aliphatic C16-C21	50634896.000	27807189.000	10962966.000	5844076.000	3012669.000	
Aliphatic C21-C28	65391355.000	36849784.000	14322959.000	8150425.000	4666116.000	
Aliphatic C28-C40	76863576.000	43549713.000	17723612.000	10141838.000	6197225.000	
Aliphatic EPH	273226603.000	151446876.000	60183610.000	33168081.000	18500677.000	

AVG Response Factor

Parameter Range	AVG RF	% RSD				
Aliphatic C9-C12	171162.641333	5.459				
Aliphatic C12-C16	178754.319	5.236				
Aliphatic C16-C21	186505.4959998	6.59				
Aliphatic C21-C28	192766.144	13.927				
Aliphatic C28-C40	159314.647333	18.927				
Aliphatic EPH	175414.750555	11.628				

Concentration

Parameter Range	FE056575.D	FE056576.D	FE056577.D	FE056578.D	FE056579.D	
Aliphatic C9-C12	300.000	150.000	60.000	30.000	15.000	
Aliphatic C12-C16	200.000	100.000	40.000	20.000	10.000	
Aliphatic C16-C21	300.000	150.000	60.000	30.000	15.000	
Aliphatic C21-C28	400.000	200.000	80.000	40.000	20.000	
Aliphatic C28-C40	600.000	300.000	120.000	60.000	30.000	
Aliphatic EPH	1800.000	900.000	360.000	180.000	90.000	

Response Factor

Parameter Range	FE056575.D	FE056576.D	FE056577.D	FE056578.D	FE056579.D	
Aliphatic C9-C12	157927.800000	169152.173333	168914.566666	177465.600000	182353.066666	
Aliphatic C12-C16	164792.180000	178673.640000	175979.975000	185388.700000	188937.100000	
Aliphatic C16-C21	168782.986666	185381.260000	182716.100000	194802.533333	200844.600000	

Initial Calibration Report for SequenceID : FE103025AL

Aliphatic C21-C28	163478.387500	184248.920000	179036.987500	203760.625000	233305.800000	
Aliphatic C28-C40	128105.960000	145165.710000	147696.766666	169030.633333	206574.166666	
Aliphatic EPH	151792.557222	168274.306666	167176.694444	184267.116666	205563.077777	

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056575.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 10:20
 Operator : YP\AJ
 Sample : 100 PPM ALIPHATIC HC STD1
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
 Quant Time: Oct 30 12:03:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 11:59:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.992	19059851	96.235 ug/ml
Spiked Amount 50.000		Recovery =	192.47%
12) S 1-chlorooctadecane (S...	13.427	14749289	96.361 ug/ml
Spiked Amount 50.000		Recovery =	192.72%
Target Compounds			
1) T n-Nonane (C9)	3.337	15510371	96.767 ug/ml
2) T n-Decane (C10)	4.591	15797359	96.498 ug/ml
3) T A~Naphthalene (C11.7)	6.308	17418486	97.460 ug/ml
4) T n-Dodecane (C12)	6.765	16070610	96.653 ug/ml
5) T A~2-methylnaphthalene...	7.409	16955970	98.036 ug/ml
6) T n-Tetradecane (C14)	8.601	16108586	96.787 ug/ml
7) T n-Hexadecane (C16)	10.215	16849850	96.650 ug/ml
8) T n-Octadecane (C18)	11.666	17249418	96.237 ug/ml
10) T n-Eicosane (C20)	12.979	16787284	95.930 ug/ml
11) T n-Heneicosane (C21)	13.593	16598194	95.935 ug/ml
13) T n-Docosane (C22)	14.181	16594646	95.890 ug/ml
14) T n-Tetracosane (C24)	15.288	16611261	95.631 ug/ml
15) T n-Hexacosane (C26)	16.313	16325668	95.495 ug/ml
16) T n-Octacosane (C28)	17.266	15859780	94.791 ug/ml
17) T n-Tricontane (C30)	18.155	15787066	93.066 ug/ml
18) T n-Dotriacontane (C32)	18.990	14882432	91.297 ug/ml
19) T n-Tetratriacontane (C34)	19.776	13398021	91.103 ug/ml
20) T n-Hexatriacontane (C36)	20.516	11781393	92.518 ug/ml
21) T n-Octatriacontane (C38)	21.272	10906555	95.307 ug/ml
22) T n-Tetracontane (C40)	22.223	10108109	95.430 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056575.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 10:20
Operator : YP\AJ
Sample : 100 PPM ALIPHATIC HC STD1
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
Quant Time: Oct 30 12:03:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 11:59:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.337 min
Delta R.T.: 0.000 min
Response: 15510371
Conc: 96.77 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#2 n-Decane (C10)

R.T.: 4.591 min
Delta R.T.: 0.000 min
Response: 15797359
Conc: 96.50 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 17418486
Conc: 97.46 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.765 min
Delta R.T.: 0.000 min
Response: 16070610
Conc: 96.65 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.409 min
Delta R.T.: 0.000 min
Response: 16955970
Conc: 98.04 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#6 n-Tetradecane (C14)

R.T.: 8.601 min
Delta R.T.: 0.000 min
Response: 16108586
Conc: 96.79 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.215 min
Delta R.T.: 0.000 min
Response: 16849850
Conc: 96.65 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.666 min
Delta R.T.: 0.000 min
Response: 17249418
Conc: 96.24 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.992 min
Delta R.T.: 0.000 min
Response: 19059851
Conc: 96.24 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#10 n-Eicosane (C20)

R.T.: 12.979 min
Delta R.T.: 0.000 min
Response: 16787284
Conc: 95.93 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.593 min
Delta R.T.: 0.000 min
Response: 16598194
Conc: 95.93 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.427 min
Delta R.T.: 0.000 min
Response: 14749289
Conc: 96.36 ug/ml



#13 n-Docosane (C22)


R.T.: 14.181 min
Delta R.T.: 0.000 min
Response: 16594646
Conc: 95.89 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#14 n-Tetracosane (C24)

R.T.: 15.288 min
Delta R.T.: 0.000 min
Response: 16611261
Conc: 95.63 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.313 min
Delta R.T.: 0.000 min
Response: 16325668
Conc: 95.50 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.266 min
Delta R.T.: 0.000 min
Response: 15859780
Conc: 94.79 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.155 min
Delta R.T.: 0.000 min
Response: 15787066
Conc: 93.07 ug/ml

Instrument :
FID_E
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#18 n-Dotriacontane (C32)

R.T.: 18.990 min
Delta R.T.: 0.000 min
Response: 14882432
Conc: 91.30 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.776 min
Delta R.T.: 0.000 min
Response: 13398021
Conc: 91.10 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.516 min
Delta R.T.: 0.000 min
Response: 11781393
Conc: 92.52 ug/ml



#21 n-Octatriacontane (C38)


R.T.: 21.272 min
Delta R.T.: 0.000 min
Response: 10906555
Conc: 95.31 ug/ml

Instrument :

FID_E

ClientSampleId :

100 PPM ALIPHATIC HC STD1



#22 n-Tetracontane (C40)

R.T.: 22.223 min
Delta R.T.: 0.000 min
Response: 10108109
Conc: 95.43 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056575.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 10:20
 Sample : 100 PPM ALIPHATIC HC STD1
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.337	3.300	3.397	BB	1587559	15510371	81.38%	4.543%
2	4.591	4.534	4.619	BBA	1581962	15797359	82.88%	4.627%
3	6.308	6.252	6.419	BB	1757870	17418486	91.39%	5.102%
4	6.765	6.700	6.809	BB	1584765	16070610	84.32%	4.707%
5	7.409	7.362	7.530	PV	1636096	16955970	88.96%	4.966%
6	8.601	8.535	8.655	BB	1480313	16108586	84.52%	4.718%
7	10.215	10.165	10.275	BB	1502827	16849850	88.40%	4.935%
8	11.666	11.610	11.724	BB	1469692	17249418	90.50%	5.052%
9	11.992	11.924	12.057	BB	1639468	19059851	100.00%	5.583%
10	12.979	12.909	13.037	BB	1409836	16787284	88.08%	4.917%
11	13.427	13.335	13.494	BB	1223044	14749289	77.38%	4.320%
12	13.593	13.520	13.649	BB	1361908	16598194	87.08%	4.862%
13	14.181	14.105	14.240	BB	1330972	16594646	87.07%	4.861%
14	15.288	15.217	15.344	BB	1288586	16611261	87.15%	4.865%
15	16.313	16.255	16.369	BB	1263736	16325668	85.65%	4.782%
16	17.266	17.200	17.299	BBA	1155432	15859780	83.21%	4.645%
17	18.155	18.089	18.210	BB	1127298	15787066	82.83%	4.624%
18	18.990	18.920	19.042	BB	1047781	14882432	78.08%	4.359%
19	19.776	19.720	19.799	BBA	935373	13398021	70.29%	3.924%
20	20.516	20.450	20.549	BBA	822409	11781393	61.81%	3.451%
21	21.272	21.210	21.434	BB	606103	10906555	57.22%	3.195%
22	22.223	22.129	22.317	BB	398455	10108109	53.03%	2.961%
Sum of corrected areas:						341410195		

Aliphatic EPH 103025.M Thu Oct 30 12:55:02 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056576.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 10:50
 Operator : YP\AJ
 Sample : 50 PPM ALIPHATIC HC STD2
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
 Quant Time: Oct 30 12:05:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 11:59:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.985	10390390	51.615 ug/ml
Spiked Amount 50.000		Recovery =	103.23%
12) S 1-chlorooctadecane (S...	13.422	8069798	51.782 ug/ml
Spiked Amount 50.000		Recovery =	103.56%
Target Compounds			
1) T n-Nonane (C9)	3.334	8283699	51.108 ug/ml
2) T n-Decane (C10)	4.587	8458707	51.101 ug/ml
3) T A~Naphthalene (C11.7)	6.303	9255687	51.178 ug/ml
4) T n-Dodecane (C12)	6.761	8630420	51.255 ug/ml
5) T A~2-methylnaphthalene...	7.404	9018328	51.408 ug/ml
6) T n-Tetradecane (C14)	8.597	8702299	51.502 ug/ml
7) T n-Hexadecane (C16)	10.210	9165065	51.685 ug/ml
8) T n-Octadecane (C18)	11.660	9466424	51.842 ug/ml
10) T n-Eicosane (C20)	12.974	9224749	51.777 ug/ml
11) T n-Heneicosane (C21)	13.588	9116016	51.761 ug/ml
13) T n-Docosane (C22)	14.176	9188407	52.021 ug/ml
14) T n-Tetracosane (C24)	15.283	9376021	52.583 ug/ml
15) T n-Hexacosane (C26)	16.307	9244371	52.644 ug/ml
16) T n-Octacosane (C28)	17.261	9040985	52.620 ug/ml
17) T n-Tricontane (C30)	18.150	9116641	52.435 ug/ml
18) T n-Dotriacontane (C32)	18.985	8697901	52.190 ug/ml
19) T n-Tetratriacontane (C34)	19.770	7726773	51.665 ug/ml
20) T n-Hexatriacontane (C36)	20.510	6593672	51.172 ug/ml
21) T n-Octatriacontane (C38)	21.266	5889118	50.965 ug/ml
22) T n-Tetracontane (C40)	22.214	5525608	51.424 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056576.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 10:50
Operator : YP\AJ
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
Quant Time: Oct 30 12:05:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 11:59:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.334 min
Delta R.T.: 0.000 min
Response: 8283699
Conc: 51.11 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#2 n-Decane (C10)

R.T.: 4.587 min
Delta R.T.: 0.000 min
Response: 8458707
Conc: 51.10 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.303 min
Delta R.T.: 0.000 min
Response: 9255687
Conc: 51.18 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.761 min
Delta R.T.: 0.000 min
Response: 8630420
Conc: 51.25 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 9018328
Conc: 51.41 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#6 n-Tetradecane (C14)

R.T.: 8.597 min
Delta R.T.: 0.000 min
Response: 8702299
Conc: 51.50 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.210 min
Delta R.T.: 0.000 min
Response: 9165065
Conc: 51.68 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.660 min
Delta R.T.: 0.000 min
Response: 9466424
Conc: 51.84 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.985 min
Delta R.T.: 0.000 min
Response: 10390390
Conc: 51.61 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#10 n-Eicosane (C20)

R.T.: 12.974 min
Delta R.T.: 0.000 min
Response: 9224749
Conc: 51.78 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.588 min
Delta R.T.: 0.000 min
Response: 9116016
Conc: 51.76 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.422 min
Delta R.T.: 0.000 min
Response: 8069798
Conc: 51.78 ug/ml



#13 n-Docosane (C22)


R.T.: 14.176 min
Delta R.T.: 0.000 min
Response: 9188407
Conc: 52.02 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#14 n-Tetracosane (C24)

R.T.: 15.283 min
Delta R.T.: 0.000 min
Response: 9376021
Conc: 52.58 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.307 min
Delta R.T.: 0.000 min
Response: 9244371
Conc: 52.64 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.261 min
Delta R.T.: 0.000 min
Response: 9040985
Conc: 52.62 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.150 min
Delta R.T.: 0.000 min
Response: 9116641
Conc: 52.43 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#18 n-Dotriacontane (C32)

R.T.: 18.985 min
Delta R.T.: 0.000 min
Response: 8697901
Conc: 52.19 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.770 min
Delta R.T.: 0.000 min
Response: 7726773
Conc: 51.67 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.510 min
Delta R.T.: 0.000 min
Response: 6593672
Conc: 51.17 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.266 min
Delta R.T.: 0.000 min
Response: 5889118
Conc: 50.97 ug/ml

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2



#22 n-Tetracontane (C40)

R.T.: 22.214 min
Delta R.T.: 0.000 min
Response: 5525608
Conc: 51.42 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056576.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 10:50
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.334	3.295	3.390	BB	859468	8283699	79.72%	4.402%
2	4.587	4.554	4.619	BBA	868249	8458707	81.41%	4.495%
3	6.303	6.240	6.415	BB	974202	9255687	89.08%	4.919%
4	6.761	6.697	6.814	BB	869786	8630420	83.06%	4.586%
5	7.404	7.364	7.522	BB	890333	9018328	86.79%	4.792%
6	8.597	8.535	8.647	BB	849680	8702299	83.75%	4.624%
7	10.210	10.152	10.272	BB	848274	9165065	88.21%	4.870%
8	11.660	11.549	11.722	BB	822893	9466424	91.11%	5.030%
9	11.985	11.924	12.049	BB	981415	10390390	100.00%	5.521%
10	12.974	12.912	13.025	BB	797927	9224749	88.78%	4.902%
11	13.422	13.335	13.479	BB	689355	8069798	77.67%	4.288%
12	13.588	13.520	13.642	BB	780350	9116016	87.74%	4.844%
13	14.176	14.142	14.232	VB	743785	9188407	88.43%	4.883%
14	15.283	15.194	15.345	BB	757600	9376021	90.24%	4.982%
15	16.307	16.235	16.362	BB	727588	9244371	88.97%	4.912%
16	17.261	17.200	17.299	BBA	690402	9040985	87.01%	4.804%
17	18.150	18.082	18.205	BB	679486	9116641	87.74%	4.845%
18	18.985	18.920	19.037	BB	654654	8697901	83.71%	4.622%
19	19.770	19.720	19.799	BBA	541984	7726773	74.36%	4.106%
20	20.510	20.450	20.549	BBA	457176	6593672	63.46%	3.504%
21	21.266	21.210	21.332	BB	342971	5889118	56.68%	3.129%
22	22.214	22.129	22.307	BB	230200	5525608	53.18%	2.936%
Sum of corrected areas:						188181079		

Aliphatic EPH 103025.M Thu Oct 30 12:57:04 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056577.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 11:20
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
 Quant Time: Oct 30 12:00:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 11:59:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.980	4110239	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	13.418	3172652	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.333	3309356	20.000 ug/ml
2) T n-Decane (C10)	4.585	3388814	20.000 ug/ml
3) T A~Naphthalene (C11.7)	6.302	3665296	20.000 ug/ml
4) T n-Dodecane (C12)	6.758	3436704	20.000 ug/ml
5) T A~2-methylnaphthalene...	7.402	3527076	20.000 ug/ml
6) T n-Tetradecane (C14)	8.593	3435589	20.000 ug/ml
7) T n-Hexadecane (C16)	10.207	3603610	20.000 ug/ml
8) T n-Octadecane (C18)	11.656	3719660	20.000 ug/ml
10) T n-Eicosane (C20)	12.970	3642327	20.000 ug/ml
11) T n-Heneicosane (C21)	13.584	3600979	20.000 ug/ml
13) T n-Docosane (C22)	14.172	3603427	20.000 ug/ml
14) T n-Tetracosane (C24)	15.278	3625778	20.000 ug/ml
15) T n-Hexacosane (C26)	16.302	3573189	20.000 ug/ml
16) T n-Octacosane (C28)	17.255	3520565	20.000 ug/ml
17) T n-Tricontane (C30)	18.146	3627924	20.000 ug/ml
18) T n-Dotriacontane (C32)	18.981	3543930	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.766	3202986	20.000 ug/ml
20) T n-Hexatriacontane (C36)	20.507	2737391	20.000 ug/ml
21) T n-Octatriacontane (C38)	21.262	2396120	20.000 ug/ml
22) T n-Tetracontane (C40)	22.211	2215261	20.000 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056577.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 11:20
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
Quant Time: Oct 30 12:00:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 11:59:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 3309356
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#2 n-Decane (C10)

R.T.: 4.585 min
Delta R.T.: 0.000 min
Response: 3388814
Conc: 20.00 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.302 min
Delta R.T.: 0.000 min
Response: 3665296
Conc: 20.00 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.758 min
Delta R.T.: 0.000 min
Response: 3436704
Conc: 20.00 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.402 min
Delta R.T.: 0.000 min
Response: 3527076
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#6 n-Tetradecane (C14)

R.T.: 8.593 min
Delta R.T.: 0.000 min
Response: 3435589
Conc: 20.00 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.207 min
Delta R.T.: 0.000 min
Response: 3603610
Conc: 20.00 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.656 min
Delta R.T.: 0.000 min
Response: 3719660
Conc: 20.00 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.980 min
Delta R.T.: 0.000 min
Response: 4110239
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#10 n-Eicosane (C20)

R.T.: 12.970 min
Delta R.T.: 0.000 min
Response: 3642327
Conc: 20.00 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.584 min
Delta R.T.: 0.000 min
Response: 3600979
Conc: 20.00 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.418 min
Delta R.T.: 0.000 min
Response: 3172652
Conc: 20.00 ug/ml



#13 n-Docosane (C22)


R.T.: 14.172 min
Delta R.T.: 0.000 min
Response: 3603427
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#14 n-Tetracosane (C24)

R.T.: 15.278 min
Delta R.T.: 0.000 min
Response: 3625778
Conc: 20.00 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.302 min
Delta R.T.: 0.000 min
Response: 3573189
Conc: 20.00 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.255 min
Delta R.T.: 0.000 min
Response: 3520565
Conc: 20.00 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.146 min
Delta R.T.: 0.000 min
Response: 3627924
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#18 n-Dotriacontane (C32)

R.T.: 18.981 min
Delta R.T.: 0.000 min
Response: 3543930
Conc: 20.00 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.766 min
Delta R.T.: 0.000 min
Response: 3202986
Conc: 20.00 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.507 min
Delta R.T.: 0.000 min
Response: 2737391
Conc: 20.00 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.262 min
Delta R.T.: 0.000 min
Response: 2396120
Conc: 20.00 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD3



#22 n-Tetracontane (C40)

R.T.: 22.211 min
Delta R.T.: 0.000 min
Response: 2215261
Conc: 20.00 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056577.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 11:20
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.333	3.302	3.387	BB	346822	3309356	80.51%	4.433%
2	4.585	4.547	4.619	BBA	352192	3388814	82.45%	4.539%
3	6.302	6.244	6.407	BB	364861	3665296	89.17%	4.909%
4	6.758	6.702	6.802	BB	351549	3436704	83.61%	4.603%
5	7.402	7.354	7.499	BB	342824	3527076	85.81%	4.724%
6	8.593	8.535	8.644	BB	338958	3435589	83.59%	4.602%
7	10.207	10.152	10.264	BB	338730	3603610	87.67%	4.827%
8	11.656	11.589	11.714	BB	336890	3719660	90.50%	4.982%
9	11.980	11.915	12.049	BB	384534	4110239	100.00%	5.505%
10	12.970	12.909	13.022	BB	322608	3642327	88.62%	4.879%
11	13.418	13.362	13.474	BB	264565	3172652	77.19%	4.250%
12	13.584	13.515	13.639	BB	315131	3600979	87.61%	4.823%
13	14.172	14.115	14.222	BB	303235	3603427	87.67%	4.827%
14	15.278	15.215	15.339	BB	298227	3625778	88.21%	4.856%
15	16.302	16.242	16.354	BB	285534	3573189	86.93%	4.786%
16	17.255	17.200	17.299	BBA	264764	3520565	85.65%	4.716%
17	18.146	18.082	18.195	BB	266055	3627924	88.27%	4.859%
18	18.981	18.920	19.037	BB	246347	3543930	86.22%	4.747%
19	19.766	19.720	19.799	BBA	237128	3202986	77.93%	4.290%
20	20.507	20.450	20.549	BBA	189105	2737391	66.60%	3.667%
21	21.262	21.210	21.314	BB	138965	2396120	58.30%	3.209%
22	22.211	22.129	22.297	BB	92560	2215261	53.90%	2.967%
Sum of corrected areas:							74658870	

Aliphatic EPH 103025.M Thu Oct 30 12:57:56 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056578.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 11:51
 Operator : YP\AJ
 Sample : 10 PPM ALIPHATIC HC STD4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
 Quant Time: Oct 30 12:21:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 12:21:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.979	2175105	10.592 ug/ml
Spiked Amount 50.000		Recovery =	21.18%
12) S 1-chlorooctadecane (S...	13.417	1673178	10.542 ug/ml
Spiked Amount 50.000		Recovery =	21.08%
Target Compounds			
1) T n-Nonane (C9)	3.332	1739861	10.541 ug/ml
2) T n-Decane (C10)	4.584	1778707	10.549 ug/ml
3) T A~Naphthalene (C11.7)	6.302	1900301	10.376 ug/ml
4) T n-Dodecane (C12)	6.757	1805400	10.532 ug/ml
5) T A~2-methylnaphthalene...	7.401	1809756	10.235 ug/ml
6) T n-Tetradecane (C14)	8.592	1806960	10.512 ug/ml
7) T n-Hexadecane (C16)	10.206	1900814	10.530 ug/ml
8) T n-Octadecane (C18)	11.654	1972225	10.589 ug/ml
10) T n-Eicosane (C20)	12.969	1949024	10.689 ug/ml
11) T n-Heneicosane (C21)	13.583	1922827	10.673 ug/ml
13) T n-Docosane (C22)	14.171	1965821	10.824 ug/ml
14) T n-Tetracosane (C24)	15.278	2071768	11.167 ug/ml
15) T n-Hexacosane (C26)	16.302	2057752	11.236 ug/ml
16) T n-Octacosane (C28)	17.255	2055084	11.402 ug/ml
17) T n-Tricontane (C30)	18.145	2121276	11.564 ug/ml
18) T n-Dotriacontane (C32)	18.980	2084284	11.769 ug/ml
19) T n-Tetratriacontane (C34)	19.764	1830892	11.592 ug/ml
20) T n-Hexatriacontane (C36)	20.507	1529216	11.338 ug/ml
21) T n-Octatriacontane (C38)	21.262	1330818	11.096 ug/ml
22) T n-Tetracontane (C40)	22.208	1245352	11.147 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056578.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 11:51
Operator : YP\AJ
Sample : 10 PPM ALIPHATIC HC STD4
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
Quant Time: Oct 30 12:21:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 12:21:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.332 min
Delta R.T.: 0.000 min
Response: 1739861
Conc: 10.54 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#2 n-Decane (C10)

R.T.: 4.584 min
Delta R.T.: 0.000 min
Response: 1778707
Conc: 10.55 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.302 min
Delta R.T.: 0.000 min
Response: 1900301
Conc: 10.38 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.757 min
Delta R.T.: 0.000 min
Response: 1805400
Conc: 10.53 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.401 min
Delta R.T.: 0.000 min
Response: 1809756
Conc: 10.24 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#6 n-Tetradecane (C14)

R.T.: 8.592 min
Delta R.T.: 0.000 min
Response: 1806960
Conc: 10.51 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.206 min
Delta R.T.: 0.000 min
Response: 1900814
Conc: 10.53 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.654 min
Delta R.T.: 0.000 min
Response: 1972225
Conc: 10.59 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.979 min
Delta R.T.: 0.000 min
Response: 2175105
Conc: 10.59 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#10 n-Eicosane (C20)

R.T.: 12.969 min
Delta R.T.: 0.000 min
Response: 1949024
Conc: 10.69 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.583 min
Delta R.T.: 0.000 min
Response: 1922827
Conc: 10.67 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.417 min
Delta R.T.: 0.000 min
Response: 1673178
Conc: 10.54 ug/ml



#13 n-Docosane (C22)


R.T.: 14.171 min
Delta R.T.: 0.000 min
Response: 1965821
Conc: 10.82 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#14 n-Tetracosane (C24)

R.T.: 15.278 min
Delta R.T.: 0.000 min
Response: 2071768
Conc: 11.17 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.302 min
Delta R.T.: 0.000 min
Response: 2057752
Conc: 11.24 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.255 min
Delta R.T.: 0.000 min
Response: 2055084
Conc: 11.40 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.145 min
Delta R.T.: 0.000 min
Response: 2121276
Conc: 11.56 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#18 n-Dotriacontane (C32)

R.T.: 18.980 min
Delta R.T.: 0.000 min
Response: 2084284
Conc: 11.77 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.764 min
Delta R.T.: 0.000 min
Response: 1830892
Conc: 11.59 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.507 min
Delta R.T.: 0.000 min
Response: 1529216
Conc: 11.34 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.262 min
Delta R.T.: 0.000 min
Response: 1330818
Conc: 11.10 ug/ml

Instrument :
FID_E
ClientSampleId :
10 PPM ALIPHATIC HC STD4



#22 n-Tetracontane (C40)

R.T.: 22.208 min
Delta R.T.: 0.000 min
Response: 1245352
Conc: 11.15 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056578.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 11:51
 Sample : 10 PPM ALIPHATIC HC STD4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.332	3.304	3.372	BB	183333	1739861	79.99%	4.272%
2	4.584	4.547	4.620	BBA	182225	1778707	81.78%	4.367%
3	6.302	6.264	6.397	BB	182183	1900301	87.37%	4.666%
4	6.757	6.727	6.795	BB	185256	1805400	83.00%	4.433%
5	7.401	7.340	7.500	BB	165998	1809756	83.20%	4.444%
6	8.592	8.537	8.637	BB	181787	1806960	83.07%	4.437%
7	10.206	10.152	10.269	BB	174738	1900814	87.39%	4.667%
8	11.654	11.589	11.714	BB	176404	1972225	90.67%	4.843%
9	11.979	11.927	12.045	BB	199071	2175105	100.00%	5.341%
10	12.969	12.912	13.024	BB	166384	1949024	89.61%	4.786%
11	13.417	13.362	13.467	BB	147215	1673178	76.92%	4.108%
12	13.583	13.524	13.637	BB	163350	1922827	88.40%	4.721%
13	14.171	14.144	14.222	VB	165444	1965821	90.38%	4.827%
14	15.278	15.192	15.335	BB	161436	2071768	95.25%	5.087%
15	16.302	16.242	16.355	BB	158775	2057752	94.60%	5.053%
16	17.255	17.200	17.300	BBA	155015	2055084	94.48%	5.046%
17	18.145	18.077	18.195	BB	156516	2121276	97.53%	5.209%
18	18.980	18.920	19.030	BB	156591	2084284	95.82%	5.118%
19	19.764	19.720	19.800	BBA	130721	1830892	84.17%	4.496%
20	20.507	20.450	20.550	BBA	104860	1529216	70.31%	3.755%
21	21.262	21.210	21.319	BB	76821	1330818	61.18%	3.268%
22	22.208	22.129	22.294	BB	50974	1245352	57.25%	3.058%
Sum of corrected areas:							40726422	

Aliphatic EPH 103025.M Thu Oct 30 12:58:36 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056579.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 12:21
 Operator : YP\AJ
 Sample : 5 PPM ALIPHATIC HC STD5
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
 Quant Time: Oct 30 12:48:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 12:47:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.979	1115676	5.340 ug/ml
Spiked Amount 50.000		Recovery =	10.68%
12) S 1-chlorooctadecane (S...	13.417	853728	5.299 ug/ml
Spiked Amount 50.000		Recovery =	10.60%
Target Compounds			
1) T n-Nonane (C9)	3.333	897930	5.346 ug/ml
2) T n-Decane (C10)	4.585	917558	5.347 ug/ml
3) T A~Naphthalene (C11.7)	6.303	951651	5.156 ug/ml
4) T n-Dodecane (C12)	6.757	919808	5.288 ug/ml
5) T A~2-methylnaphthalene...	7.403	894288	5.046 ug/ml
6) T n-Tetradecane (C14)	8.592	925756	5.304 ug/ml
7) T n-Hexadecane (C16)	10.206	963615	5.267 ug/ml
8) T n-Octadecane (C18)	11.654	1006643	5.319 ug/ml
10) T n-Eicosane (C20)	12.969	1012155	5.431 ug/ml
11) T n-Heneicosane (C21)	13.582	993871	5.405 ug/ml
13) T n-Docosane (C22)	14.171	1047683	5.597 ug/ml
14) T n-Tetracosane (C24)	15.277	1186109	6.056 ug/ml
15) T n-Hexacosane (C26)	16.302	1195552	6.152 ug/ml
16) T n-Octacosane (C28)	17.255	1236772	6.386 ug/ml
17) T n-Tricontane (C30)	18.145	1343777	6.702 ug/ml
18) T n-Dotriacontane (C32)	18.981	1374204	6.988 ug/ml
19) T n-Tetratriacontane (C34)	19.765	1156477	6.700 ug/ml
20) T n-Hexatriacontane (C36)	20.507	919020	6.353 ug/ml
21) T n-Octatriacontane (C38)	21.262	747858	5.942 ug/ml
22) T n-Tetracontane (C40)	22.209	655889	5.673 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056579.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 12:21
Operator : YP\AJ
Sample : 5 PPM ALIPHATIC HC STD5
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
Quant Time: Oct 30 12:48:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 12:47:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.333 min
Delta R.T.: 0.000 min
Response: 897930
Conc: 5.35 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#2 n-Decane (C10)

R.T.: 4.585 min
Delta R.T.: 0.000 min
Response: 917558
Conc: 5.35 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.303 min
Delta R.T.: 0.000 min
Response: 951651
Conc: 5.16 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.757 min
Delta R.T.: 0.000 min
Response: 919808
Conc: 5.29 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.403 min
Delta R.T.: 0.000 min
Response: 894288
Conc: 5.05 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#6 n-Tetradecane (C14)

R.T.: 8.592 min
Delta R.T.: 0.000 min
Response: 925756
Conc: 5.30 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.206 min
Delta R.T.: 0.000 min
Response: 963615
Conc: 5.27 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.654 min
Delta R.T.: 0.000 min
Response: 1006643
Conc: 5.32 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.979 min
Delta R.T.: 0.000 min
Response: 1115676
Conc: 5.34 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#10 n-Eicosane (C20)

R.T.: 12.969 min
Delta R.T.: 0.000 min
Response: 1012155
Conc: 5.43 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.582 min
Delta R.T.: 0.000 min
Response: 993871
Conc: 5.40 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.417 min
Delta R.T.: 0.000 min
Response: 853728
Conc: 5.30 ug/ml



#13 n-Docosane (C22)


R.T.: 14.171 min
Delta R.T.: 0.000 min
Response: 1047683
Conc: 5.60 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#14 n-Tetracosane (C24)

R.T.: 15.277 min
Delta R.T.: 0.000 min
Response: 1186109
Conc: 6.06 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.302 min
Delta R.T.: 0.000 min
Response: 1195552
Conc: 6.15 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.255 min
Delta R.T.: 0.000 min
Response: 1236772
Conc: 6.39 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.145 min
Delta R.T.: 0.000 min
Response: 1343777
Conc: 6.70 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#18 n-Dotriacontane (C32)

R.T.: 18.981 min
Delta R.T.: 0.000 min
Response: 1374204
Conc: 6.99 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.765 min
Delta R.T.: 0.000 min
Response: 1156477
Conc: 6.70 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.507 min
Delta R.T.: 0.000 min
Response: 919020
Conc: 6.35 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.262 min
Delta R.T.: 0.000 min
Response: 747858
Conc: 5.94 ug/ml

Instrument :
FID_E
ClientSampleId :
5 PPM ALIPHATIC HC STD5



#22 n-Tetracontane (C40)

R.T.: 22.209 min
Delta R.T.: 0.000 min
Response: 655889
Conc: 5.67 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056579.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 12:21
 Sample : 5 PPM ALIPHATIC HC STD5
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.333	3.303	3.377	BB	92471	897930	65.34%	4.024%
2	4.585	4.545	4.619	BBA	95384	917558	66.77%	4.112%
3	6.303	6.268	6.377	BB	88968	951651	69.25%	4.264%
4	6.757	6.700	6.800	BB	95428	919808	66.93%	4.122%
5	7.403	7.347	7.487	BB	80203	894288	65.08%	4.007%
6	8.592	8.540	8.642	BB	88772	925756	67.37%	4.148%
7	10.206	10.138	10.265	BB	89800	963615	70.12%	4.318%
8	11.654	11.588	11.710	BB	89973	1006643	73.25%	4.511%
9	11.979	11.920	12.035	BB	104613	1115676	81.19%	4.999%
10	12.969	12.905	13.018	BB	87940	1012155	73.65%	4.536%
11	13.417	13.363	13.465	BB	72050	853728	62.13%	3.826%
12	13.582	13.517	13.635	BB	85261	993871	72.32%	4.454%
13	14.171	14.145	14.218	VB	86814	1047683	76.24%	4.695%
14	15.277	15.187	15.330	BB	94557	1186109	86.31%	5.315%
15	16.302	16.243	16.357	BB	94717	1195552	87.00%	5.357%
16	17.255	17.200	17.299	BBA	95523	1236772	90.00%	5.542%
17	18.145	18.077	18.192	BB	99085	1343777	97.79%	6.022%
18	18.981	18.920	19.032	BB	95776	1374204	100.00%	6.158%
19	19.765	19.720	19.799	BBA	84124	1156477	84.16%	5.182%
20	20.507	20.450	20.549	BBA	66647	919020	66.88%	4.118%
21	21.262	21.210	21.328	BB	41776	747858	54.42%	3.351%
22	22.209	22.123	22.287	BB	26317	655889	47.73%	2.939%
Sum of corrected areas:							22316018	

Aliphatic EPH 103025.M Thu Oct 30 12:59:22 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056580.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 12:51
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
 Quant Time: Oct 30 13:21:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 30 12:50:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.979	3698663	17.704 ug/ml
Spiked Amount 50.000		Recovery =	35.41%
12) S 1-chlorooctadecane (S...	13.417	2853138	17.708 ug/ml
Spiked Amount 50.000		Recovery =	35.42%
Target Compounds			
1) T n-Nonane (C9)	3.331	2974720	17.711 ug/ml
2) T n-Decane (C10)	4.584	3032863	17.675 ug/ml
3) T A~Naphthalene (C11.7)	6.301	3268504	17.707 ug/ml
4) T n-Dodecane (C12)	6.757	3080927	17.714 ug/ml
5) T A~2-methylnaphthalene...	7.401	3138399	17.709 ug/ml
6) T n-Tetradecane (C14)	8.592	3087511	17.688 ug/ml
7) T n-Hexadecane (C16)	10.206	3246411	17.744 ug/ml
8) T n-Octadecane (C18)	11.655	3352800	17.714 ug/ml
10) T n-Eicosane (C20)	12.969	3290342	17.656 ug/ml
11) T n-Heneicosane (C21)	13.583	3256292	17.709 ug/ml
13) T n-Docosane (C22)	14.171	3252345	17.374 ug/ml
14) T n-Tetracosane (C24)	15.279	3265557	16.673 ug/ml
15) T n-Hexacosane (C26)	16.303	3203978	16.487 ug/ml
16) T n-Octacosane (C28)	17.256	3101810	16.017 ug/ml
17) T n-Tricontane (C30)	18.145	3129479	15.609 ug/ml
18) T n-Dotriacontane (C32)	18.980	2966227	15.084 ug/ml
19) T n-Tetratriacontane (C34)	19.765	2678828	15.520 ug/ml
20) T n-Hexatriacontane (C36)	20.508	2321116	16.046 ug/ml
21) T n-Octatriacontane (C38)	21.262	2080984	16.534 ug/ml
22) T n-Tetracontane (C40)	22.209	1952947	16.892 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
Data File : FE056580.D
Signal(s) : FID1B.ch
Acq On : 30 Oct 2025 12:51
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD ICV
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
Quant Time: Oct 30 13:21:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
Quant Title : GC Extractables
QLast Update : Thu Oct 30 12:50:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um





#1 n-Nonane (C9)


R.T.: 3.331 min
Delta R.T.: -0.002 min
Response: 2974720
Conc: 17.71 ug/ml

Instrument :

FID_E


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#2 n-Decane (C10)

R.T.: 4.584 min
Delta R.T.: 0.000 min
Response: 3032863
Conc: 17.67 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.301 min
Delta R.T.: -0.003 min
Response: 3268504
Conc: 17.71 ug/ml



#4 n-Dodecane (C12)

R.T.: 6.757 min
Delta R.T.: 0.000 min
Response: 3080927
Conc: 17.71 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.401 min
Delta R.T.: -0.002 min
Response: 3138399
Conc: 17.71 ug/ml

Instrument :

FID_E


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#6 n-Tetradecane (C14)

R.T.: 8.592 min
Delta R.T.: 0.000 min
Response: 3087511
Conc: 17.69 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.206 min
Delta R.T.: 0.000 min
Response: 3246411
Conc: 17.74 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.655 min
Delta R.T.: 0.000 min
Response: 3352800
Conc: 17.71 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.979 min
Delta R.T.: 0.000 min
Response: 3698663
Conc: 17.70 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#10 n-Eicosane (C20)

R.T.: 12.969 min
Delta R.T.: 0.000 min
Response: 3290342
Conc: 17.66 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.583 min
Delta R.T.: 0.001 min
Response: 3256292
Conc: 17.71 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.417 min
Delta R.T.: 0.000 min
Response: 2853138
Conc: 17.71 ug/ml



#13 n-Docosane (C22)


R.T.: 14.171 min
Delta R.T.: 0.000 min
Response: 3252345
Conc: 17.37 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#14 n-Tetracosane (C24)

R.T.: 15.279 min
Delta R.T.: 0.001 min
Response: 3265557
Conc: 16.67 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.303 min
Delta R.T.: 0.000 min
Response: 3203978
Conc: 16.49 ug/ml



#16 n-Octacosane (C28)

R.T.: 17.256 min
Delta R.T.: 0.001 min
Response: 3101810
Conc: 16.02 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.145 min
Delta R.T.: 0.000 min
Response: 3129479
Conc: 15.61 ug/ml

Instrument :

FID_E


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#18 n-Dotriacontane (C32)

R.T.: 18.980 min
Delta R.T.: 0.000 min
Response: 2966227
Conc: 15.08 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.765 min
Delta R.T.: 0.000 min
Response: 2678828
Conc: 15.52 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.508 min
Delta R.T.: 0.000 min
Response: 2321116
Conc: 16.05 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.262 min
Delta R.T.: 0.000 min
Response: 2080984
Conc: 16.53 ug/ml

Instrument :
FID_E
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV



#22 n-Tetracontane (C40)

R.T.: 22.209 min
Delta R.T.: 0.000 min
Response: 1952947
Conc: 16.89 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103025AL\
 Data File : FE056580.D
 Signal(s) : FID1B.ch
 Acq On : 30 Oct 2025 12:51
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 103025.M
 Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.331	3.300	3.384	BB	310481	2974720	80.43%	4.491%
2	4.584	4.549	4.620	BBA	315244	3032863	82.00%	4.579%
3	6.301	6.269	6.407	BB	321401	3268504	88.37%	4.935%
4	6.757	6.702	6.799	BB	318158	3080927	83.30%	4.652%
5	7.401	7.345	7.512	BB	293420	3138399	84.85%	4.738%
6	8.592	8.532	8.645	BB	305141	3087511	83.48%	4.662%
7	10.206	10.152	10.279	BB	304154	3246411	87.77%	4.901%
8	11.655	11.589	11.717	BB	305409	3352800	90.65%	5.062%
9	11.979	11.917	12.040	BB	334249	3698663	100.00%	5.584%
10	12.969	12.910	13.025	BB	280820	3290342	88.96%	4.968%
11	13.417	13.362	13.475	BB	241413	2853138	77.14%	4.308%
12	13.583	13.515	13.637	BB	286535	3256292	88.04%	4.916%
13	14.171	14.112	14.229	BB	266967	3252345	87.93%	4.910%
14	15.279	15.215	15.335	BB	270242	3265557	88.29%	4.930%
15	16.303	16.242	16.349	BB	253168	3203978	86.63%	4.837%
16	17.256	17.200	17.300	BBA	243564	3101810	83.86%	4.683%
17	18.145	18.080	18.204	BB	236990	3129479	84.61%	4.725%
18	18.980	18.920	19.027	BB	221501	2966227	80.20%	4.478%
19	19.765	19.720	19.800	BBA	187610	2678828	72.43%	4.044%
20	20.508	20.450	20.550	BBA	162660	2321116	62.76%	3.504%
21	21.262	21.210	21.315	BB	117144	2080984	56.26%	3.142%
22	22.209	22.129	22.302	BB	79731	1952947	52.80%	2.949%
Sum of corrected areas:							66233841	

Aliphatic EPH 103025.M Thu Oct 30 13:22:51 2025