

Continuing Calibration Report for SequenceID : FG111225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016878.D

Aliphatic C9-C12	8051141.000	60.000	3.261	6.902	134185.683	122003.227	-9.985
Aliphatic C12-C16	5513270.000	40.000	6.903	10.360	137831.750	122851.240	-12.194
Aliphatic C16-C21	8697375.000	60.000	10.361	13.749	144956.250	132396.253	-9.487
Aliphatic C21-C28	11593691.000	80.000	13.750	17.437	144921.138	139034.659	-4.234
Aliphatic C28-C40	13716515.000	120.000	17.438	22.463	114304.292	113794.545	-0.448
Aliphatic EPH	47571992.000	360.000	3.261	22.463	132144.422	124878.157	-5.819