

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ROMA02
SDG NO.: Q3604

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO114871.D</u>		CF 750 = <u>PO114872.D</u>				
CF 500 = <u>PO114873.D</u>		CF 250 = <u>PO114874.D</u>		CF 050 = <u>PO114875.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	235183609	242405119	252300612	252330724	247880040	246020021	3
Aroclor-1016-2	(2)	380671765	391174169	403435360	408228592	401844700	397070917	3
Aroclor-1016-3	(3)	214892716	220811199	230510330	231508016	217282560	223000964	3
Aroclor-1016-4	(4)	180735079	184273587	191842228	192742164	194782360	188875084	3
Aroclor-1016-5	(5)	166479276	173610481	179587174	180772736	190536000	178197133	5
Aroclor-1260-1	(1)	291080181	301659089	311914366	329562032	293951540	305633442	5
Aroclor-1260-2	(2)	393658847	402595176	414352186	418536840	350242960	395877202	7
Aroclor-1260-3	(3)	331096941	340266576	347857958	348147780	291720640	331817979	7
Aroclor-1260-4	(4)	303823322	310642864	321126472	321507392	268000240	305020058	7
Aroclor-1260-5	(5)	714393776	724690244	737226162	719415648	670367440	713218654	3
Decachlorobiphenyl		4451734610	4516548787	4612561720	4524869360	3995527600	4420248415	6
Tetrachloro-m-xylene		7020403280	7162783533	7307727940	7156075200	6543618600	7038121711	4
Aroclor-1242-1	(1)	180395803	184480117	189796832	192180524	184350480	186240751	2
Aroclor-1242-2	(2)	290607708	295647931	305830174	302003732	299791360	298776181	2
Aroclor-1242-3	(3)	164449629	167684461	173929682	174410736	176369820	171368866	3
Aroclor-1242-4	(4)	136844790	140633719	146402846	139660276	152376540	143183634	4
Aroclor-1242-5	(5)	138362947	141839337	145002188	138385972	141196180	140957325	2
Decachlorobiphenyl		4729850460	4804993427	4858375920	4746069120	4160138600	4659885505	6
Tetrachloro-m-xylene		7348004600	7445204373	7532777260	7262019280	6694579200	7256516943	5
Aroclor-1248-1	(1)	130126160	135244476	137909000	141742196	154206600	139845686	6
Aroclor-1248-2	(2)	182893435	187680161	194064692	196392780	211353600	194476934	5
Aroclor-1248-3	(3)	203368877	206306241	213983954	212927108	217587500	210834736	3
Aroclor-1248-4	(4)	240181487	246210361	255848644	255957264	264773260	252594203	4
Aroclor-1248-5	(5)	225312398	230738423	239469078	240115800	243213880	235769916	3
Decachlorobiphenyl		4590720290	4647354867	4744288880	4661202760	4211361200	4570985599	5
Tetrachloro-m-xylene		7316541360	7442934853	7528054520	7402608560	7218400400	7381707939	2
Aroclor-1254-1	(1)	240591852	247416493	258765478	257075152	263299240	253429643	3
Aroclor-1254-2	(2)	346021204	353529297	371589652	368897016	412184820	370444398	7
Aroclor-1254-3	(3)	372893964	380446979	397009938	384293176	389678940	384864599	2
Aroclor-1254-4	(4)	315539209	321424513	336840998	326904952	342871860	328716306	3
Aroclor-1254-5	(5)	295373361	300690525	312155076	300474784	300573260	301853401	2
Decachlorobiphenyl		4791769470	4896862187	5061968640	4831244600	4661831200	4848735219	3
Tetrachloro-m-xylene		7410585440	7526809333	7802184660	7481754040	7408041400	7525874975	2
Aroclor-1268-1	(1)	876279188	885717404	893041022	887529864	813508780	871215252	4

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	748334854	756337347	765525536	759915712	702804360	746583562	3
Aroclor-1268-3	(3)	663299706	672085827	679866912	671661004	617886840	660960058	4
Aroclor-1268-4	(4)	274155738	277531193	282310020	282161716	286808980	280593529	2
Aroclor-1268-5	(5)	1966581606	1979440064	1985729368	1956798376	1807416340	1939193151	4
Decachlorobiphenyl		8288385350	8396434587	8504946880	8419357920	7722111200	8266247187	4
Tetrachloro-m-xylene		7475792880	7636683133	7655291620	7597134240	7121440400	7497268455	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ROMA02
SDG NO.: Q3604

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: CF 1000 = <u>PO114871.D</u> CF 750 = <u>PO114872.D</u> CF 500 = <u>PO114873.D</u> CF 250 = <u>PO114874.D</u> CF 050 = <u>PO114875.D</u>							
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	178795127	185535456	193106802	200932132	192907960	4
Aroclor-1016-2	(2)	260787154	267435697	277593168	286495324	262010880	4
Aroclor-1016-3	(3)	140310175	144533207	150201526	154962192	149858780	4
Aroclor-1016-4	(4)	108607533	112021955	117721682	122827936	125909940	6
Aroclor-1016-5	(5)	142235561	147493437	153611756	157990124	162236540	5
Aroclor-1260-1	(1)	272865489	283696256	302549666	306994580	314431380	5
Aroclor-1260-2	(2)	410849028	420744668	440173148	441650472	436738740	3
Aroclor-1260-3	(3)	328919579	344471297	362206406	366201912	353957800	4
Aroclor-1260-4	(4)	295043629	302500340	313130718	316509244	287959420	4
Aroclor-1260-5	(5)	819932115	830411719	851082708	846532168	760181660	4
Decachlorobiphenyl		3768213910	3862452533	4051635320	3997632280	3839185200	3
Tetrachloro-m-xylene		4444199150	4532189360	4610968540	4573862320	4172678600	4
Aroclor-1242-1	(1)	138347633	144438088	149108388	148877560	150336280	3
Aroclor-1242-2	(2)	199908386	204961789	213926082	211880332	205062860	3
Aroclor-1242-3	(3)	108216818	112229521	116418372	113485296	109537600	3
Aroclor-1242-4	(4)	101919263	105301932	110004430	109236748	109193360	3
Aroclor-1242-5	(5)	135980897	140103160	145653410	150301700	142549620	4
Decachlorobiphenyl		4093300380	4301738307	4403735460	4338775160	4380898600	3
Tetrachloro-m-xylene		4800115840	4683347560	4961464680	4527345840	4437059000	4
Aroclor-1248-1	(1)	101027399	105625799	111262386	113093188	115539800	5
Aroclor-1248-2	(2)	136540550	142594189	148530518	152692968	164144180	7
Aroclor-1248-3	(3)	143887712	149887199	155474250	158349236	169440400	6
Aroclor-1248-4	(4)	171266913	177666692	184171754	187513908	195642380	5
Aroclor-1248-5	(5)	174649655	180220384	187038042	189642400	190092660	3
Decachlorobiphenyl		4076042330	4166088587	4290124020	4403270520	4144788800	3
Tetrachloro-m-xylene		4758509080	4829545453	4892334280	4811209560	4727034000	1
Aroclor-1254-1	(1)	275439068	284238351	300243386	299923620	341055960	8
Aroclor-1254-2	(2)	245333366	253279232	267984098	268318040	313122940	9
Aroclor-1254-3	(3)	398238109	410614839	430545692	426530308	445294740	4
Aroclor-1254-4	(4)	284785979	294895399	305027804	301955520	300306220	3
Aroclor-1254-5	(5)	385708423	392326712	409498118	398089984	402006780	2
Decachlorobiphenyl		4273373360	4367226933	4604811080	4533103920	4272134600	3
Tetrachloro-m-xylene		4791085320	4887004773	5044010700	4876428000	4843036000	2
Aroclor-1268-1	(1)	948830555	965398547	981732904	993582148	991082540	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	784312077	796694211	811227528	817006596	768929160	795633914	2
Aroclor-1268-3	(3)	672635507	681509284	695723330	696225136	632763280	675771307	4
Aroclor-1268-4	(4)	270862616	275402580	280086838	282372280	267609400	275266743	2
Aroclor-1268-5	(5)	1768544208	1786014605	1818457210	1831952004	1717683400	1784530285	2
Decachlorobiphenyl		7349980390	7437822467	7706201340	7861559120	8321979800	7735508623	5
Tetrachloro-m-xylene		4791521570	4902547960	4915920660	4883273320	4724156000	4843483902	2