

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Continuing Calib Date: 11/12/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 15:55

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.33	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Continuing Calib Date: 11/12/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 15:55

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.18	5.18	5.08	5.28	0.00
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00