

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Continuing Calib Date: 11/12/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 10:23

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|-------|------------|
| Aroclor-1242-1 (1)   | 5.79       | 5.80      | 5.70                   | 5.90  | 0.01       |
| Aroclor-1242-2 (2)   | 5.81       | 5.82      | 5.72                   | 5.92  | 0.01       |
| Aroclor-1242-3 (3)   | 5.88       | 5.88      | 5.78                   | 5.98  | 0.00       |
| Aroclor-1242-4 (4)   | 5.97       | 5.98      | 5.88                   | 6.08  | 0.01       |
| Aroclor-1242-5 (5)   | 6.70       | 6.71      | 6.61                   | 6.81  | 0.01       |
| Tetrachloro-m-xylene | 4.64       | 4.65      | 4.55                   | 4.75  | 0.01       |
| Decachlorobiphenyl   | 10.41      | 10.42     | 10.32                  | 10.52 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Continuing Calib Date: 11/12/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 10:23

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.87       | 4.88      | 4.78      | 4.98 | 0.01       |
| Aroclor-1242-2 (2)   | 4.93       | 4.94      | 4.84      | 5.04 | 0.01       |
| Aroclor-1242-3 (3)   | 5.05       | 5.06      | 4.96      | 5.16 | 0.01       |
| Aroclor-1242-4 (4)   | 5.13       | 5.14      | 5.04      | 5.24 | 0.01       |
| Aroclor-1242-5 (5)   | 5.65       | 5.66      | 5.56      | 5.76 | 0.01       |
| Tetrachloro-m-xylene | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Decachlorobiphenyl   | 8.77       | 8.79      | 8.69      | 8.89 | 0.02       |