

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115031.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	502.390	500.000	0.5
Aroclor-1016-2	5.507	5.408	5.608	500.300	500.000	0.1
Aroclor-1016-3	5.569	5.470	5.670	514.140	500.000	2.8
Aroclor-1016-4	5.665	5.566	5.766	503.350	500.000	0.7
Aroclor-1016-5	5.957	5.858	6.058	492.960	500.000	-1.4
Aroclor-1260-1	7.070	6.971	7.171	480.960	500.000	-3.8
Aroclor-1260-2	7.324	7.225	7.425	516.090	500.000	3.2
Aroclor-1260-3	7.682	7.582	7.782	515.890	500.000	3.2
Aroclor-1260-4	7.903	7.803	8.003	499.810	500.000	0.0
Aroclor-1260-5	8.210	8.110	8.310	497.030	500.000	-0.6
Decachlorobiphenyl	9.932	9.830	10.030	49.750	50.000	-0.5
Tetrachloro-m-xylene	4.346	4.248	4.448	51.820	50.000	3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115031.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	494.480	500.000	-1.1
Aroclor-1016-2	4.745	4.647	4.847	491.490	500.000	-1.7
Aroclor-1016-3	4.919	4.821	5.021	501.010	500.000	0.2
Aroclor-1016-4	4.962	4.864	5.064	502.510	500.000	0.5
Aroclor-1016-5	5.174	5.076	5.276	500.560	500.000	0.1
Aroclor-1260-1	6.205	6.106	6.306	472.430	500.000	-5.5
Aroclor-1260-2	6.392	6.294	6.494	465.770	500.000	-6.8
Aroclor-1260-3	6.545	6.446	6.646	471.710	500.000	-5.7
Aroclor-1260-4	7.015	6.917	7.117	480.570	500.000	-3.9
Aroclor-1260-5	7.257	7.158	7.358	476.520	500.000	-4.7
Decachlorobiphenyl	8.639	8.541	8.741	51.660	50.000	3.3
Tetrachloro-m-xylene	3.648	3.548	3.748	49.940	50.000	-0.1