

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076469.D **Time Analyzed:** 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.794	5.695	5.895	519.400	500.000	3.9
Aroclor-1016-2	5.815	5.717	5.917	535.200	500.000	7.0
Aroclor-1016-3	5.878	5.780	5.980	530.110	500.000	6.0
Aroclor-1016-4	5.975	5.877	6.077	542.530	500.000	8.5
Aroclor-1016-5	6.267	6.169	6.369	539.540	500.000	7.9
Aroclor-1260-1	7.385	7.287	7.487	553.370	500.000	10.7
Aroclor-1260-2	7.638	7.540	7.740	544.070	500.000	8.8
Aroclor-1260-3	7.997	7.899	8.099	560.250	500.000	12.1
Aroclor-1260-4	8.225	8.127	8.327	565.840	500.000	13.2
Aroclor-1260-5	8.550	8.452	8.652	526.810	500.000	5.4
Decachlorobiphenyl	10.413	10.319	10.519	48.750	50.000	-2.5
Tetrachloro-m-xylene	4.641	4.542	4.742	52.940	50.000	5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076469.D **Time Analyzed:** 15:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.872	4.778	4.978	512.340	500.000	2.5
Aroclor-1016-2	4.930	4.836	5.036	518.280	500.000	3.7
Aroclor-1016-3	5.049	4.954	5.154	501.000	500.000	0.2
Aroclor-1016-4	5.090	4.996	5.196	537.150	500.000	7.4
Aroclor-1016-5	5.305	5.210	5.410	523.160	500.000	4.6
Aroclor-1260-1	6.331	6.238	6.438	534.460	500.000	6.9
Aroclor-1260-2	6.519	6.425	6.625	511.160	500.000	2.2
Aroclor-1260-3	6.672	6.578	6.778	535.490	500.000	7.1
Aroclor-1260-4	7.140	7.047	7.247	567.330	500.000	13.5
Aroclor-1260-5	7.382	7.287	7.487	507.090	500.000	1.4
Decachlorobiphenyl	8.774	8.682	8.882	42.170	50.000	-15.7
Tetrachloro-m-xylene	3.778	3.681	3.881	51.520	50.000	3.0