284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance
Contract: ROMA02
Lab Code: ACE
SDG NO.: Q3604
Instrument ID: ECD_S
Calibration Date(s): 11/03/2025 11/03/2025
Calibration Times: 09:40 11:17

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID: RT 200 = PS032226.D RT 500 = PS032227.D
RT 750 = PS032228.D RT 1000 = PS032229.D RT 1500 = PS032230.D

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM	RT WINDOW TO
2,4,5-TP(Silvex)	9.28	9.28	9.28	9.28	9.28	9.28	9.18	9.38
2,4-D	8.43	8.42	8.41	8.41	8.41	8.42	8.32	8.52
2,4-DCAA	7.29	7.29	7.28	7.28	7.28	7.28	7.18	7.38

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_S
Contract: ROMA02
SDG NO.: Q3604
Calibration Date(s): 11/03/2025 11/03/2025
Calibration Times: 09:40 11:17

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID: RT 200 = PS032226.D RT 500 = PS032227.D
RT 750 = PS032228.D RT 1000 = PS032229.D RT 1500 = PS032230.D

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM	RT WINDOW TO
2,4,5-TP(Silvex)	9.75	9.75	9.74	9.74	9.74	9.75	9.65	9.85
2,4-D	8.89	8.87	8.86	8.86	8.86	8.87	8.77	8.97
2,4-DCAA	7.63	7.62	7.62	7.62	7.61	7.62	7.52	7.72