

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Instrument ID: ECD_S

Calibration Date(s): 11/03/2025 11/03/2025

Calibration Times: 09:40 11:17

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS032226.D</u>	CF 500 =	<u>PS032227.D</u>		
CF 750 =		<u>PS032228.D</u>	CF 1000 =	<u>PS032229.D</u>	CF 1500 =	<u>PS032230.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	29002100000	27213000000	26386400000	24343100000	23113100000	26011600000	9
2,4-D	5213180000	4699390000	4506050000	4152780000	3987670000	4511810000	11
2,4-DCAA	5119490000	4707160000	4538440000	4190170000	4101110000	4531270000	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Instrument ID: ECD_S

Calibration Date(s): 11/03/2025 11/03/2025

Calibration Times: 09:40 11:17

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS032226.D</u>	CF 500 =	<u>PS032227.D</u>		
	CF 750 =	<u>PS032228.D</u>	CF 1000 =	<u>PS032229.D</u>	CF 1500 =	<u>PS032230.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	30276100000	28543900000	27314100000	24890000000	23245400000	26853900000	10
2,4-D	5684310000	5535780000	5360230000	4911890000	4736080000	5245660000	8
2,4-DCAA	3895870000	3715630000	3686230000	3472610000	3457140000	3645500000	5