

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032383.D **Time Analyzed:** 20:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.276	9.180	9.380	760.790	712.500	6.8
2,4-D	8.410	8.314	8.514	748.450	705.000	6.2
2,4-DCAA	7.282	7.183	7.383	790.220	750.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032383.D **Time Analyzed:** 20:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.745	9.644	9.844	754.130	712.500	5.8
2,4-D	8.863	8.764	8.964	752.680	705.000	6.8
2,4-DCAA	7.615	7.517	7.717	822.760	750.000	9.7