

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ROMA02
Lab Code: ACE
SDG NO.: Q3604
GC Column: RTX-CLP
ID: 0.32 (mm)
 Initi. Calib. Date(s): 11/03/2025 11/03/2025

Client Sample No.: CCAL02
Date Analyzed: 11/13/2025
Lab Sample No.: HSTDCCC750
Data File : PS032392.D
Time Analyzed: 01:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.276	9.180	9.380	776.890	712.500	9.0
2,4-D	8.411	8.314	8.514	757.130	705.000	7.4
2,4-DCAA	7.282	7.183	7.383	808.580	750.000	7.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/03/2025 11/03/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/13/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032392.D **Time Analyzed:** 01:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.745	9.644	9.844	758.490	712.500	6.5
2,4-D	8.866	8.764	8.964	750.930	705.000	6.5
2,4-DCAA	7.617	7.517	7.717	801.960	750.000	6.9