

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3604</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Roman E&G Corp</u>	Datafile :	<u>PS032385.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170516BS (Column 1)	2,4-D	5	5.40	ug/L	108				70 (83)	130 (130)	
	2,4,5-TP(Silvex)	5	5.50	ug/L	110				70 (78)	130 (127)	
PB170516BS (Column 2)	2,4-D	5	5.40	ug/L	108				70 (83)	130 (130)	
	2,4,5-TP(Silvex)	5	5.50	ug/L	110				70 (78)	130 (127)	