

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170498BS

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Lab Sample ID: PB170498BS

Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	0.43	15.1
	2	6.74	6.69	6.79	0.50	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	0.47	9.4
	2	3.72	3.67	3.77	0.52	
Heptachlor	1	4.92	4.87	4.97	0.55	4.9
	2	4.07	4.02	4.12	0.52	
Heptachlor epoxide	1	5.68	5.63	5.73	0.48	6.3
	2	4.86	4.81	4.91	0.51	
Endrin	1	6.57	6.52	6.62	0.41	12.8
	2	5.78	5.73	5.83	0.46	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

S-1MS

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Lab Sample ID: Q3604-01MS

Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.00	11.3
	2	6.74	6.69	6.79	5.60	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	5.00	2
	2	3.72	3.67	3.77	5.10	
Heptachlor	1	4.92	4.87	4.97	5.80	10.9
	2	4.07	4.02	4.12	5.20	
Heptachlor epoxide	1	5.68	5.63	5.73	5.00	2
	2	4.86	4.81	4.91	5.10	
Endrin	1	6.57	6.52	6.62	4.90	5.9
	2	5.78	5.73	5.83	5.20	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

S-1MSD

Lab Name: Alliance

Contract: ROMA02

Lab Code: ACE

SDG NO.: Q3604

Lab Sample ID: Q3604-01MSD

Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.00	11.3
	2	6.74	6.69	6.79	5.60	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	4.90	4
	2	3.72	3.67	3.77	5.10	
Heptachlor	1	4.92	4.87	4.97	5.80	10.9
	2	4.07	4.02	4.12	5.20	
Heptachlor epoxide	1	5.68	5.63	5.73	5.00	2
	2	4.86	4.81	4.91	5.10	
Endrin	1	6.57	6.52	6.62	4.90	5.9
	2	5.78	5.73	5.83	5.20	