

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091130.D **Time Analyzed:** 21:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.964	9.164	43.910	50.000	-12.2
Endrin	6.567	6.468	6.668	44.320	50.000	-11.4
gamma-BHC (Lindane)	4.324	4.225	4.425	47.940	50.000	-4.1
Heptachlor	4.922	4.823	5.023	55.760	50.000	11.5
Heptachlor epoxide	5.683	5.584	5.784	47.360	50.000	-5.3
Methoxychlor	7.489	7.388	7.588	43.050	50.000	-13.9
Tetrachloro-m-xylene	3.544	3.445	3.645	47.140	50.000	-5.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091130.D **Time Analyzed:** 21:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.962	8.162	53.920	50.000	7.8
Endrin	5.778	5.681	5.881	50.650	50.000	1.3
gamma-BHC (Lindane)	3.722	3.623	3.823	52.040	50.000	4.1
Heptachlor	4.073	3.975	4.175	52.980	50.000	6.0
Heptachlor epoxide	4.862	4.764	4.964	50.700	50.000	1.4
Methoxychlor	6.743	6.645	6.845	49.480	50.000	-1.0
Tetrachloro-m-xylene	2.874	2.775	2.975	51.990	50.000	4.0