

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ROMA02
Lab Code: ACE
SDG NO.: Q3604
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL02
Date Analyzed: 11/12/2025
Lab Sample No.: PSTDCCC050
Data File : PD091139.D
Time Analyzed: 23:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.964	9.164	45.650	50.000	-8.7
Endrin	6.567	6.468	6.668	45.080	50.000	-9.8
gamma-BHC (Lindane)	4.323	4.225	4.425	48.230	50.000	-3.5
Heptachlor	4.921	4.823	5.023	55.720	50.000	11.4
Heptachlor epoxide	5.683	5.584	5.784	47.920	50.000	-4.2
Methoxychlor	7.488	7.388	7.588	44.090	50.000	-11.8
Tetrachloro-m-xylene	3.543	3.445	3.645	48.400	50.000	-3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3604
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091139.D **Time Analyzed:** 23:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.962	8.162	56.660	50.000	13.3
Endrin	5.778	5.681	5.881	50.570	50.000	1.1
gamma-BHC (Lindane)	3.721	3.623	3.823	51.310	50.000	2.6
Heptachlor	4.073	3.975	4.175	51.770	50.000	3.5
Heptachlor epoxide	4.862	4.764	4.964	50.220	50.000	0.4
Methoxychlor	6.742	6.645	6.845	50.100	50.000	0.2
Tetrachloro-m-xylene	2.873	2.775	2.975	51.500	50.000	3.0