

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090646.D	17 Oct 2025 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD090647.D	17 Oct 2025 10:53		AR\AJ	Ok
3	PEM	PEM	PD090648.D	17 Oct 2025 11:07		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090649.D	17 Oct 2025 11:20		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD090650.D	17 Oct 2025 11:34		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090651.D	17 Oct 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090652.D	17 Oct 2025 12:01		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090653.D	17 Oct 2025 12:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090654.D	17 Oct 2025 12:28		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090656.D	17 Oct 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090657.D	17 Oct 2025 13:09		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090658.D	17 Oct 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090659.D	17 Oct 2025 13:36		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090660.D	17 Oct 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090661.D	17 Oct 2025 14:03		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090662.D	17 Oct 2025 14:17		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090663.D	17 Oct 2025 14:30		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090664.D	17 Oct 2025 16:18		AR\AJ	Ok
20	PSTDICV050	ICVPD101725	PD090665.D	17 Oct 2025 16:46		AR\AJ	Ok
21	PCHLORICV500	ICVPD101725CHLOR	PD090666.D	17 Oct 2025 17:50		AR\AJ	Ok
22	PTOXICV500	ICVPD101725TOX	PD090667.D	17 Oct 2025 18:42		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD090668.D	17 Oct 2025 19:50		AR\AJ	Ok
24	PEM	PEM	PD090669.D	17 Oct 2025 20:04		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090670.D	17 Oct 2025 20:18		AR\AJ	Ok
26	PB170142BL	PB170142BL	PD090671.D	17 Oct 2025 20:31		AR\AJ	Ok
27	PB170142BS	PB170142BS	PD090672.D	17 Oct 2025 20:45		AR\AJ	Ok
28	Q3366-01	TR-06-101625	PD090673.D	17 Oct 2025 20:59		AR\AJ	Ok,M
29	Q3366-03	TR-1-101625	PD090674.D	17 Oct 2025 21:12		AR\AJ	Ok,M
30	Q3367-01	BU-03-101625	PD090675.D	17 Oct 2025 21:26		AR\AJ	Ok,M
31	Q3369-01	MH-22	PD090676.D	17 Oct 2025 21:40		AR\AJ	Ok,M
32	Q3369-05	MH-21	PD090677.D	17 Oct 2025 21:53		AR\AJ	Ok,M
33	Q3363-01	PR132-SO1-085011-20	PD090678.D	17 Oct 2025 22:48		AR\AJ	Ok,M
34	Q3363-02	PR132-SO1-039085-20	PD090679.D	17 Oct 2025 23:02		AR\AJ	Ok,M
35	Q3363-03	PR132-SO3-028010-20	PD090680.D	17 Oct 2025 23:15		AR\AJ	Ok,M
36	Q3363-04	PR132-SO3-010013-20	PD090681.D	17 Oct 2025 23:29		AR\AJ	Ok,M
37	Q3363-04MS	PR132-SO3-010013-20	PD090682.D	17 Oct 2025 23:43		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3363-04MSD	PR132-SO3-010013-20	PD090683.D	17 Oct 2025 23:56		AR\AJ	Ok,M
39	Q3363-05	COMP01-20251015	PD090684.D	18 Oct 2025 00:10		AR\AJ	Ok,M
40	Q3343-04MS	SOIL-ROLLOFF-WC-11	PD090685.D	18 Oct 2025 00:24		AR\AJ	Ok,M
41	Q3343-04MSD	SOIL-ROLLOFF-WC-11	PD090686.D	18 Oct 2025 00:37		AR\AJ	Ok
42	I.BLK	I.BLK	PD090687.D	18 Oct 2025 00:51		AR\AJ	Ok
43	PSTDCCC050	PSTDCCC050	PD090688.D	18 Oct 2025 02:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111225

Review By	rahul	Review On	11/12/2025 9:01:58 AM
Supervise By	yogesh	Supervise On	11/13/2025 1:18:12 PM
SubDirectory	PD111225	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091080.D	12 Nov 2025 08:48		AR\AJ	Ok
2	I.BLK	I.BLK	PD091081.D	12 Nov 2025 09:01		AR\AJ	Ok
3	PEM	PEM	PD091082.D	12 Nov 2025 09:15		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD091083.D	12 Nov 2025 09:28		AR\AJ	Ok
5	PB170476BL	PB170476BL	PD091084.D	12 Nov 2025 09:44		AR\AJ	Ok
6	PB170476BS	PB170476BS	PD091085.D	12 Nov 2025 09:57		AR\AJ	Ok,M
7	PB170485BL	PB170485BL	PD091086.D	12 Nov 2025 10:11		AR\AJ	Ok
8	PB170485BS	PB170485BS	PD091087.D	12 Nov 2025 10:25		AR\AJ	Ok,M
9	PB1701485BSD	PB1701485BSD	PD091088.D	12 Nov 2025 10:38		AR\AJ	Ok,M
10	Q3593-04	A-LAW-25-0167-0175-C	PD091089.D	12 Nov 2025 10:52		AR\AJ	Ok
11	Q3593-06	B-LAW-25-0167-0175-C	PD091090.D	12 Nov 2025 11:06		AR\AJ	Ok
12	Q3593-08	C-LAW-25-0156-0166-C	PD091091.D	12 Nov 2025 11:19	DCB high 1st column	AR\AJ	Ok
13	Q3593-10	D-LAW-25-0156-0166-C	PD091092.D	12 Nov 2025 11:33	Need 10X	AR\AJ	Not Ok
14	PB170485BSD	PB170485BSD	PD091093.D	12 Nov 2025 11:47		AR\AJ	Ok,M
15	Q3569-01	MW-2	PD091094.D	12 Nov 2025 12:00		AR\AJ	Ok,NS
16	Q3569-02	MW-12	PD091095.D	12 Nov 2025 12:14		AR\AJ	Ok
17	Q3584-01	SW-1	PD091096.D	12 Nov 2025 12:28		AR\AJ	Ok,NS
18	Q3593-10DL	D-LAW-25-0156-0166-C	PD091097.D	12 Nov 2025 12:42		AR\AJ	Not Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111225

Review By	rahul	Review On	11/12/2025 9:01:58 AM
Supervise By	yogesh	Supervise On	11/13/2025 1:18:12 PM
SubDirectory	PD111225	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3584-02	SW-2	PD091098.D	12 Nov 2025 12:55		AR\AJ	Ok,NS
20	Q3593-10	D-LAW-25-0156-0166-Q	PD091099.D	12 Nov 2025 13:09	Need 10X	AR\AJ	Dilution
21	Q3584-03	SW-3	PD091100.D	12 Nov 2025 13:22		AR\AJ	Ok,NS
22	Q3593-10DL	D-LAW-25-0156-0166-Q	PD091101.D	12 Nov 2025 13:36		AR\AJ	Ok,M
23	Q3584-04	EB-2	PD091102.D	12 Nov 2025 13:50		AR\AJ	Ok
24	Q3584-05	FB-2	PD091103.D	12 Nov 2025 14:03		AR\AJ	Ok
25	I.BLK	I.BLK	PD091104.D	12 Nov 2025 14:17		AR\AJ	Ok
26	PEM	PEM	PD091105.D	12 Nov 2025 14:31		AR\AJ	Ok,NS
27	PSTDCCC050	PSTDCCC050	PD091106.D	12 Nov 2025 14:44		AR\AJ	Ok
28	PB170506BL	PB170506BL	PD091107.D	12 Nov 2025 14:58		AR\AJ	Ok
29	PB170506BS	PB170506BS	PD091108.D	12 Nov 2025 15:12	surrogate fail and recovery fail & Heptachlor	AR\AJ	Not Ok
30	Q3597-01	TP-1	PD091109.D	12 Nov 2025 15:26		AR\AJ	Ok,NS
31	Q3597-01MS	TP-1MS	PD091110.D	12 Nov 2025 15:39		AR\AJ	Ok,NS
32	Q3597-01MSD	TP-1MSD	PD091111.D	12 Nov 2025 15:53		AR\AJ	Ok,NS
33	Q3597-04	TP-2	PD091112.D	12 Nov 2025 16:07		AR\AJ	Ok,NS
34	Q3597-07	TP-3	PD091113.D	12 Nov 2025 16:20		AR\AJ	Ok,NS
35	Q3597-10	TP-4	PD091114.D	12 Nov 2025 16:34		AR\AJ	Ok,NS
36	Q3598-01	EO-CONCRETE	PD091115.D	12 Nov 2025 16:47		AR\AJ	Ok,NS
37	Q3600-01	CENTRAL-ROW-CONC	PD091116.D	12 Nov 2025 17:01		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111225

Review By	rahul	Review On	11/12/2025 9:01:58 AM
Supervise By	yogesh	Supervise On	11/13/2025 1:18:12 PM
SubDirectory	PD111225	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	I.BLK	I.BLK	PD091117.D	12 Nov 2025 17:15		AR\AJ	Ok
39	PSTDCCC050	PSTDCCC050	PD091118.D	12 Nov 2025 17:28		AR\AJ	Ok
40	Q3601-01	TP-1	PD091119.D	12 Nov 2025 18:09		AR\AJ	Ok,NS
41	Q3601-05	TP-2	PD091120.D	12 Nov 2025 18:23		AR\AJ	Ok,NS
42	Q3601-09	TP-3	PD091121.D	12 Nov 2025 18:37		AR\AJ	Ok,NS
43	Q3601-13	TP-4	PD091122.D	12 Nov 2025 18:50		AR\AJ	Ok,NS
44	Q3601-17	TP-5	PD091123.D	12 Nov 2025 19:04		AR\AJ	Ok,NS
45	Q3584-07	SEEP-1	PD091124.D	12 Nov 2025 19:18		AR\AJ	Ok,NS
46	Q3599-01	LIVINGSTON-SOIL	PD091125.D	12 Nov 2025 19:31		AR\AJ	Ok,NS
47	Q3590-01	MOO-25-0314-0316-CC	PD091126.D	12 Nov 2025 19:45		AR\AJ	Ok
48	Q3590-03	MOO-25-0317-0318-CC	PD091127.D	12 Nov 2025 19:59		AR\AJ	Ok,NS
49	Q3590-05	MOO-25-0319	PD091128.D	12 Nov 2025 20:12		AR\AJ	Ok,NS
50	I.BLK	I.BLK	PD091129.D	12 Nov 2025 20:53		AR\AJ	Ok
51	PSTDCCC050	PSTDCCC050	PD091130.D	12 Nov 2025 21:07		AR\AJ	Ok
52	PB170498BL	PB170498BL	PD091131.D	12 Nov 2025 21:48		AR\AJ	Ok
53	PB170498BS	PB170498BS	PD091132.D	12 Nov 2025 22:02		AR\AJ	Ok
54	PB170493TB	PB170493TB	PD091133.D	12 Nov 2025 22:15		AR\AJ	Ok
55	Q3604-01	S-1	PD091134.D	12 Nov 2025 22:29		AR\AJ	Ok,NS
56	Q3604-01MS	S-1MS	PD091135.D	12 Nov 2025 22:43		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111225

Review By	rahul	Review On	11/12/2025 9:01:58 AM
Supervise By	yogesh	Supervise On	11/13/2025 1:18:12 PM
SubDirectory	PD111225	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	Q3604-01MSD	S-1MSD	PD091136.D	12 Nov 2025 22:56		AR\AJ	Ok
58	Q3604-02	S-2	PD091137.D	12 Nov 2025 23:10		AR\AJ	Ok,NS
59	I.BLK	I.BLK	PD091138.D	12 Nov 2025 23:24		AR\AJ	Ok
60	PSTDCCC050	PSTDCCC050	PD091139.D	12 Nov 2025 23:37		AR\AJ	Ok

M : Manual Integration