

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120125\
 Data File : BP026197.D
 Acq On : 01 Dec 2025 14:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 01 14:29:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	-0.01
2	1,4-Dioxane	0.505	0.497	1.6	82	0.00
3	Pyridine	1.461	1.372	6.1	76	0.00
4	n-Nitrosodimethylamine	0.546	0.531	2.7	79	0.00
5 S	2-Fluorophenol	1.246	1.183	5.1	76	0.00
6	Aniline	2.094	1.944	7.2	74	0.00
7 S	Phenol-d6	1.586	1.496	5.7	74	-0.01
8	2-Chlorophenol	1.418	1.381	2.6	78	0.00
9	Benzaldehyde	0.831	0.844	-1.6	74	0.00
10 C	Phenol	1.709	1.612	5.7	74	0.00
11	bis(2-Chloroethyl)ether	1.338	1.312	1.9	78	0.00
12	1,3-Dichlorobenzene	1.528	1.530	-0.1	81	-0.01
13 C	1,4-Dichlorobenzene	1.540	1.537	0.2	80	0.00
14	1,2-Dichlorobenzene	1.479	1.472	0.5	80	0.00
15	Benzyl Alcohol	1.162	1.087	6.5	74	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.879	0.0	79	-0.02
17	2-Methylphenol	1.163	1.110	4.6	75	-0.02
18	Hexachloroethane	0.560	0.573	-2.3	82	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.953	3.2	75	-0.01
20	3+4-Methylphenols	1.612	1.495	7.3	74	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	77	-0.02
22	Acetophenone	0.508	0.500	1.6	76	-0.01
23 S	Nitrobenzene-d5	0.352	0.352	0.0	76	-0.01
24	Nitrobenzene	0.356	0.356	0.0	76	-0.01
25	Isophorone	0.698	0.683	2.1	74	-0.02
26 C	2-Nitrophenol	0.180	0.175	2.8	71	-0.02
27	2,4-Dimethylphenol	0.325	0.289	11.1	68	-0.02
28	bis(2-Chloroethoxy)methane	0.437	0.426	2.5	75	-0.02
29 C	2,4-Dichlorophenol	0.325	0.323	0.6	75	-0.02
30	1,2,4-Trichlorobenzene	0.331	0.337	-1.8	79	-0.02
31	Naphthalene	1.081	1.080	0.1	78	-0.02
32	Benzoic acid	0.274	0.216	21.2	60	-0.02
33	4-Chloroaniline	0.460	0.432	6.1	70	-0.02
34 C	Hexachlorobutadiene	0.215	0.227	-5.6	81	-0.03
35	Caprolactam	0.116	0.105	9.5	69	-0.02
36 C	4-Chloro-3-methylphenol	0.345	0.334	3.2	73	-0.02
37	2-Methylnaphthalene	0.766	0.763	0.4	76	-0.02
38	1-Methylnaphthalene	0.749	0.751	-0.3	76	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	73	-0.04
40	1,2,4,5-Tetrachlorobenzene	0.605	0.635	-5.0	78	-0.02
41 P	Hexachlorocyclopentadiene	0.396	0.358	9.6	66	-0.02
42 S	2,4,6-Tribromophenol	0.259	0.261	-0.8	72	-0.02
43 C	2,4,6-Trichlorophenol	0.415	0.422	-1.7	74	-0.02
44	2,4,5-Trichlorophenol	0.463	0.466	-0.6	73	-0.02
45 S	2-Fluorobiphenyl	1.365	1.415	-3.7	78	-0.03
46	1,1'-Biphenyl	1.506	1.511	-0.3	74	-0.02
47	2-Chloronaphthalene	1.184	1.183	0.1	74	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120125\
 Data File : BP026197.D
 Acq On : 01 Dec 2025 14:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 01 14:29:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.316	4.0	68	-0.02
49	Acenaphthylene	1.953	1.984	-1.6	75	-0.02
50	Dimethylphthalate	1.494	1.496	-0.1	73	-0.02
51	2,6-Dinitrotoluene	0.296	0.297	-0.3	69	-0.03
52 C	Acenaphthene	1.145	1.146	-0.1	74	-0.03
53	3-Nitroaniline	0.349	0.327	6.3	66	-0.03
54 P	2,4-Dinitrophenol	0.179	0.160	10.6	63	-0.03
55	Dibenzofuran	1.773	1.788	-0.8	74	-0.03
56 P	4-Nitrophenol	0.315	0.265	15.9	62	-0.03
57	2,4-Dinitrotoluene	0.442	0.444	-0.5	70	-0.04
58	Fluorene	1.401	1.429	-2.0	76	-0.03
59	2,3,4,6-Tetrachlorophenol	0.393	0.381	3.1	70	-0.04
60	Diethylphthalate	1.505	1.488	1.1	72	-0.03
61	4-Chlorophenyl-phenylether	0.697	0.715	-2.6	76	-0.03
62	4-Nitroaniline	0.362	0.328	9.4	66	-0.03
63	Azobenzene	1.269	1.269	0.0	72	-0.03
64 I	Phenanthrene-d10	1.000	1.000	0.0	72	-0.03
65	4,6-Dinitro-2-methylphenol	0.127	0.120	5.5	65	-0.03
66 c	n-Nitrosodiphenylamine	0.619	0.618	0.2	71	-0.03
67	4-Bromophenyl-phenylether	0.221	0.229	-3.6	72	-0.04
68	Hexachlorobenzene	0.257	0.266	-3.5	72	-0.03
69	Atrazine	0.233	0.231	0.9	70	-0.04
70 C	Pentachlorophenol	0.180	0.179	0.6	70	-0.04
71	Phenanthrene	1.127	1.149	-2.0	74	-0.03
72	Anthracene	1.149	1.145	0.3	70	-0.04
73	Carbazole	1.091	1.074	1.6	70	-0.03
74	Di-n-butylphthalate	1.294	1.325	-2.4	70	-0.04
75 C	Fluoranthene	1.369	1.373	-0.3	72	-0.03
76 I	Chrysene-d12	1.000	1.000	0.0	76	-0.03
77	Benzidine	0.635	0.579	8.8	71	-0.04
78	Pyrene	1.311	1.288	1.8	74	-0.03
79 S	Terphenyl-d14	0.981	0.994	-1.3	76	-0.04
80	Butylbenzylphthalate	0.578	0.557	3.6	70	-0.03
81	Benzo(a)anthracene	1.374	1.338	2.6	74	-0.03
82	3,3'-Dichlorobenzidine	0.500	0.463	7.4	70	-0.05
83	Chrysene	1.295	1.277	1.4	76	-0.03
84	Bis(2-ethylhexyl)phthalate	0.874	0.870	0.5	71	-0.04
85 c	Di-n-octyl phthalate	1.529	1.473	3.7	71	-0.05
86 I	Perylene-d12	1.000	1.000	0.0	77	-0.08
87	Indeno(1,2,3-cd)pyrene	1.500	1.526	-1.7	77	-0.14
88	Benzo(b)fluoranthene	1.218	1.215	0.2	75	-0.06
89	Benzo(k)fluoranthene	1.217	1.260	-3.5	79	-0.07
90 C	Benzo(a)pyrene	1.154	1.169	-1.3	77	-0.08
91	Dibenzo(a,h)anthracene	1.228	1.250	-1.8	77	-0.12
92	Benzo(g,h,i)perylene	1.220	1.252	-2.6	78	-0.15

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120125\
Data File : BP026197.D
Acq On : 01 Dec 2025 14:09
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 01 14:29:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0