

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3604

Client ID : SSTDCCC040

Date Analyzed: 11/26/2025

Lab File ID: BF144358.D

Time Analyzed: 13:30

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	61998	6.863	227764	8.15	130458	9.90
UPPER LIMIT	123996	7.363	455528	8.645	260916	10.404
LOWER LIMIT	30999	6.363	113882	7.645	65229	9.404
EPA SAMPLE NO.						
01 PB170735BL	62889	6.86	250119	8.14	143140	9.90
02 PB170735BS	62512	6.86	243072	8.15	133431	9.90
03 S-1	56195	6.86	208707	8.15	118617	9.90
04 S-2	55371	6.86	213302	8.15	114867	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.