

Surrogate Summary

SDG No.: **Q3604**

Client: **Roman E&G Corp**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3604-04	S-1	1,2-Dichloroethane-d4	50	68.0	136	*	70 (63)	130 (155)
		Dibromofluoromethane	50	53.3	107		70 (70)	130 (134)
		Toluene-d8	50	46.3	93		70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.8	108		70 (52)	130 (138)
Q3604-05	S-2	1,2-Dichloroethane-d4	50	70.4	141	*	70 (63)	130 (155)
		Dibromofluoromethane	50	53.4	107		70 (70)	130 (134)
		Toluene-d8	50	46.3	93		70 (74)	130 (123)
		4-Bromofluorobenzene	50	56.4	113		70 (52)	130 (138)
VW1126SBL01	VW1126SBL01	1,2-Dichloroethane-d4	50	61.0	122		70 (63)	130 (155)
		Dibromofluoromethane	50	54.3	109		70 (70)	130 (134)
		Toluene-d8	50	49.6	99		70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.2	94		70 (52)	130 (138)
VW1126SBS01	VW1126SBS01	1,2-Dichloroethane-d4	50	57.6	115		70 (63)	130 (155)
		Dibromofluoromethane	50	52.4	105		70 (70)	130 (134)
		Toluene-d8	50	51.9	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.4	103		70 (52)	130 (138)
VW1126SBSD0	VW1126SBSD01	1,2-Dichloroethane-d4	50	58.8	118		70 (63)	130 (155)
		Dibromofluoromethane	50	54.0	108		70 (70)	130 (134)
		Toluene-d8	50	52.0	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	53.7	107		70 (52)	130 (138)

