

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROMA02
Lab Code: ACE SDG NO.: Q3604
Lab File ID: VW032536.D Date Analyzed: 11/26/2025
Instrument ID: MSVOA_W Time Analyzed: 12:05
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	353885	7.96	632151	8.85	586351	11.63
UPPER LIMIT	707770	8.459	1264300	9.349	1172700	12.129
LOWER LIMIT	176943	7.459	316076	8.349	293176	11.129
EPA SAMPLE NO.						
S-1	202457	7.96	487508	8.85	546383	11.63
S-2	191034	7.96	461380	8.85	522966	11.63
VW1126SBL01	322545	7.96	615508	8.85	556062	11.63
VW1126SBS01	337562	7.96	629179	8.85	578751	11.63
VW1126SBSD01	335704	7.96	620682	8.86	568775	11.63

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.