

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROMA02
 Lab Code: ACE SDG NO.: Q3604
 Lab File ID: VW032536.D Date Analyzed: 11/26/2025
 Instrument ID: MSVOA_W Time Analyzed: 12:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	285431	13.555				
UPPER LIMIT	570862	14.055				
LOWER LIMIT	142716	13.055				
EPA SAMPLE NO.						
S-1	253441	13.56				
S-2	256269	13.56				
VW1126SBL01	253580	13.56				
VW1126SBS01	285723	13.56				
VW1126SBSD01	279860	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.