

Preparation Blank Summary

Client: ATG - AKRON LAB

SDG No.: Q3605

Project: PO 25101605

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137853BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/11/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/11/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/11/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/11/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/11/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/11/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/11/2025