

Duplicate Sample Summary

Client: ATG - AKRON LAB	SDG No.: Q3605
Project: PO 25101605	Sample ID: Q3603-01
Client ID: SHORT-HILLS-DWMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Nitrate	mg/L	+/-20	2.50	OR	2.50	OR	1	0		11/11/2025
Chloride	mg/L	+/-20	77.1		77.1		1	0		11/11/2025
Sulfate	mg/L	+/-20	36.0		35.5		1	1		11/11/2025
Orthophosphate as P	mg/L	+/-20	5.20		5.10		1	2		11/11/2025
Bromide	mg/L	+/-20	10.3		10.0		1	3		11/11/2025
Nitrite	mg/L	+/-20	3.10		3.00		1	3		11/11/2025
Fluoride	mg/L	+/-20	2.20		2.10		1	5		11/11/2025