

Matrix Spike Summary

Client:	ALS Environmental	SDG No.:	Q3606
Project:	ALS Middletown	Sample ID:	Q3614-01
Client ID:	COMP-1MS	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	57.5		3.20	J	58.4	1	93		11/17/2025

Matrix Spike Summary

Client:	ALS Environmental	SDG No.:	Q3606
Project:	ALS Middletown	Sample ID:	Q3614-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	82.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/Kg	75-125	56.1		3.20	J	59.6	1	89		11/17/2025